



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91060

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/24/2017	17:10
2	CAL2	1	Water			7.00	10/24/2017	17:11
3	CAL3	1	Water			10.03	10/24/2017	17:12
4	ICV	1	Water			7.00	10/24/2017	17:15
5	CCV1	1	Water			2.01	10/24/2017	17:16
6	I6093-10	1	Solid	20.02	20	6.46	10/24/2017	17:19
7	I6093-11	1	Solid	20.03	20	7.63	10/24/2017	17:20
8	I6093-12	1	Solid	20.04	20	10.44	10/24/2017	17:21
9	I6093-13	1	Solid	20.02	20	8.81	10/24/2017	17:22
10	I6101-01	1	Solid	20.03	20	7.56	10/24/2017	17:24
11	I6101-02	1	Solid	20.03	20	7.25	10/24/2017	17:26
12	I6101-03	1	Solid	20.01	20	7.68	10/24/2017	17:28
13	I6101-04	1	Solid	20.04	20	8.61	10/24/2017	17:30
14	I6101-05	1	Solid	20.02	20	7.84	10/24/2017	17:31
15	I6109-03	1	Solid	20.02	20	8.04	10/24/2017	17:33
16	CCV2	1	Water			2.01	10/24/2017	17:34
17	I6109-03DUP	1	Solid	20.03	20	8.05	10/24/2017	17:35
18	I6112-04	1	Solid	20.02	20	8.43	10/24/2017	17:36
19	I6112-08	1	Solid	20.04	20	8.14	10/24/2017	17:37
20	I6112-12	1	Solid	20.03	20	8.13	10/24/2017	17:38

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I6112-16	1	Solid	20.03	20	7.90	10/24/2017	17:39
22	I6112-20	1	Solid	20.01	20	8.12	10/24/2017	17:41
23	I6112-24	1	Solid	20.02	20	7.62	10/24/2017	17:42
24	I6113-03	1	Solid	20.03	20	8.06	10/24/2017	17:43
25	I6117-03	1	Solid	20.05	20	8.74	10/24/2017	17:44
26	I6117-03DUP	1	Solid	20.03	20	8.75	10/24/2017	17:45
27	CCV3	1	Water			12.01	10/24/2017	17:46

WORKLIST(Hardcopy Internal Chain)

1B91060

WorkList Name : CORROSIONITY I6112

WorkList ID : 104612

Date : 10/24/2017 4:11:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2017	Solid	I6093-10	Corrosivity	Cool 4 deg C	TECT01	D22	TEC-B6-B7-COMP-0-4	10/23/2017	9045C
11/04/2017	Solid	I6093-11	Corrosivity	Cool 4 deg C	TECT01	D22	TEC-B6-B7-COMP-4-12	10/23/2017	9045C
11/04/2017	Solid	I6093-12	Corrosivity	Cool 4 deg C	TECT01	D22	TEC-B8-B9-COMP-1-4	10/23/2017	9045C
11/04/2017	Solid	I6093-13	Corrosivity	Cool 4 deg C	TECT01	D22	TEC-B8-B9-COMP-4-12	10/23/2017	9045C
10/31/2017	Solid	I6101-01	Corrosivity	Cool 4 deg C	USAR03	D62	CB-1017-S-TCLP-48S(3+4)	10/23/2017	9045C
10/31/2017	Solid	I6101-02	Corrosivity	Cool 4 deg C	USAR03	D62	CB-1017-S-TCLP-48S(1+2)	10/23/2017	9045C
10/31/2017	Solid	I6101-03	Corrosivity	Cool 4 deg C	USAR03	D62	CB-1017-S-TCLP-43S3	10/23/2017	9045C
10/31/2017	Solid	I6101-04	Corrosivity	Cool 4 deg C	USAR03	D62	CB-1017-S-TCLP-43S2	10/23/2017	9045C
10/31/2017	Solid	I6101-05	Corrosivity	Cool 4 deg C	USAR03	D62	CB-1017-S-TCLP-43S1	10/23/2017	9045C
10/27/2017	Solid	I6109-03	Corrosivity	Cool 4 deg C	SKAN06	D52	S13-SP1-C	10/24/2017	9045C
10/27/2017	Solid	I6112-04	Corrosivity	Cool 4 deg C	PSEG01	D52	1-MARION26KV-(2-3)	10/24/2017	9045C
10/27/2017	Solid	I6112-08	Corrosivity	Cool 4 deg C	PSEG01	D52	2-MARION26KV-(2-3)	10/24/2017	9045C
10/27/2017	Solid	I6112-12	Corrosivity	Cool 4 deg C	PSEG01	D52	3-MARION26KV-(2-3)	10/24/2017	9045C
10/27/2017	Solid	I6112-16	Corrosivity	Cool 4 deg C	PSEG01	D52	4-MARION26KV-(2-3)	10/24/2017	9045C
10/27/2017	Solid	I6112-20	Corrosivity	Cool 4 deg C	PSEG01	D52	5-MARION26KV-(2-3)	10/24/2017	9045C
10/27/2017	Solid	I6112-24	Corrosivity	Cool 4 deg C	PSEG01	D52	6-MARION26KV-(2-3)	10/24/2017	9045C
10/27/2017	Solid	I6113-03	Corrosivity	Cool 4 deg C	PSEG05	D52	SU-01-102417	10/24/2017	9045C
10/27/2017	Solid	I6117-03	Corrosivity	Cool 4 deg C	PSEG05	D52	JC-01-102417	10/24/2017	9045C

Date/Time 10/24/17 5:00 PM
 Received by: TP
 Relinquished by: CP

Date/Time 10-24-17 7:00 PM
 Received by: CP
 Relinquished by: JP