

245.1


LB91072

INSTRUMENT ID: CY2

Linear

 μ Abs.:

32237



A= 0.0000e+000

B= 3.1140e-004

C= 2.3036e-002

Rho= 0.9999229

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	-8	0.000	-8				
0.2	0.200	0.230	0.030	666	0.0 %	666				
2.5	2.500	2.455	-0.045	7810	0.0 %	7810				
5.0	5.000	5.006	0.006	16002	0.0 %	16002				
7.5	7.500	7.426	-0.074	23775	0.0 %	23775				
10.0	10.000	10.062	0.062	32237	0.0 %	32237				

LB91072 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Oct 2017 10:23:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-8	-	0.0000245.1		PPB	25 Oct 2017 10:27:29	S	Std
0.2 - 1	S02	666	-	0.2000245.1		PPB	25 Oct 2017 10:29:38	S	Std
2.5 - 1	S2.5	7810	-	2.5000245.1		PPB	25 Oct 2017 10:31:46	S	Std
5.0 - 1	S5	16002	-	5.0000245.1		PPB	25 Oct 2017 10:33:54	S	Std
7.5 - 1	S7.5	23775	-	7.5000245.1		PPB	25 Oct 2017 10:40:20	S	Std
10.0 - 1	S10	32237	-	10.0000245.1		PPB	25 Oct 2017 10:42:27	S	Std
ICV28 - 1	ICV28	12154	3.8077	-245.1		PPB	25 Oct 2017 10:45:10	U	SMPL
ICB28 - 1	ICB28	-114	-0.0125	-245.1		PPB	25 Oct 2017 10:47:17	U	SMPL
CCV01 - 1	CCV01	15677	4.9048	-245.1		PPB	25 Oct 2017 10:49:24	U	SMPL
CCB01 - 1	CCB01	-128	-0.0168	-245.1		PPB	25 Oct 2017 10:51:30	U	SMPL
CRA - 1	CRA	649	0.2251	-245.1		PPB	25 Oct 2017 10:53:38	U	SMPL
HighStd - 1	HighStd	31976	9.9802	-245.1		PPB	25 Oct 2017 10:55:44	U	SMPL
ChkStd - 1	ChkStd	22462	7.0176	-245.1		PPB	25 Oct 2017 10:57:51	U	SMPL
PB103617BL - 1	PBW	-247	-0.0539	-245.1		PPB	25 Oct 2017 11:00:01	U	SMPL
PB103617BS - 1	LCSW	12957	4.0578	-245.1		PPB	25 Oct 2017 11:02:09	U	SMPL
I6024-25 - 1	TWP-02	263	0.1049	-245.1		PPB	25 Oct 2017 11:04:17	U	SMPL
I6024-25DUP - 1	TWP-02DUP	176	0.0778	-245.1		PPB	25 Oct 2017 11:06:26	U	SMPL
I6024-25MS - 1	TWP-02MS	14063	4.4022	-245.1		PPB	25 Oct 2017 11:08:34	U	SMPL
I6024-25MSD - 1	TWP-02MSD	14360	4.4947	-245.1		PPB	25 Oct 2017 11:10:43	U	SMPL
I6024-26 - 1	TWP-13	-128	-0.0168	-245.1		PPB	25 Oct 2017 11:12:49	U	SMPL
CCV02 - 1	CCV02	15658	4.8989	-245.1		PPB	25 Oct 2017 11:14:56	U	SMPL
CCB02 - 1	CCB02	-136	-0.0193	-245.1		PPB	25 Oct 2017 11:17:03	U	SMPL
I6057-01 - 1	101917-TIDO-F-D-B	30	0.0324	-245.1		PPB	25 Oct 2017 11:19:10	U	SMPL
I6057-02 - 1	101917-TIDO-F-D-S	-3	0.0221	-245.1		PPB	25 Oct 2017 11:21:17	U	SMPL
I6057-03 - 1	101917-TIDO-F-U-B	9	0.0258	-245.1		PPB	25 Oct 2017 11:23:24	U	SMPL
I6057-04 - 1	101917-TIDO-F-U-S	4	0.0243	-245.1		PPB	25 Oct 2017 11:25:31	U	SMPL
I6057-05 - 1	101917-SIDI-F-D-B	-13	0.0190	-245.1		PPB	25 Oct 2017 11:27:38	U	SMPL
I6057-06 - 1	101917-SIDI-F-D-M	6	0.0249	-245.1		PPB	25 Oct 2017 11:29:46	U	SMPL
I6057-07 - 1	101917-SIDI-F-D-S	-3	0.0221	-245.1		PPB	25 Oct 2017 11:31:54	U	SMPL
I6057-08 - 1	101917-SIDI-F-U-B	5	0.0246	-245.1		PPB	25 Oct 2017 11:34:02	U	SMPL
I6057-09 - 1	101917-SIDI-F-U-M	-27	0.0146	-245.1		PPB	25 Oct 2017 11:36:11	U	SMPL
I6057-10 - 1	101917-SIDI-F-U-S	-29	0.0140	-245.1		PPB	25 Oct 2017 11:38:18	U	SMPL
CCV03 - 1	CCV03	15542	4.8628	-245.1		PPB	25 Oct 2017 11:40:25	U	SMPL
CCB03 - 1	CCB03	-142	-0.0212	-245.1		PPB	25 Oct 2017 11:42:32	U	SMPL
I6059-01 - 1	101817-TIDO-F-D-B	8	0.0255	-245.1		PPB	25 Oct 2017 11:44:40	U	SMPL
I6059-02 - 1	101817-TIDO-F-D-S	15	0.0277	-245.1		PPB	25 Oct 2017 11:46:46	U	SMPL
I6059-03 - 1	101817-TIDO-F-U-B	5	0.0246	-245.1		PPB	25 Oct 2017 11:48:53	U	SMPL
I6059-04 - 1	101817-TIDO-F-U-S	-10	0.0199	-245.1		PPB	25 Oct 2017 11:51:01	U	SMPL
I6100-01 - 1	102317-TIDO-F-D-B	24	0.0305	-245.1		PPB	25 Oct 2017 11:53:08	U	SMPL
I6100-02 - 1	102317-TIDO-F-D-S	-10	0.0199	-245.1		PPB	25 Oct 2017 11:55:16	U	SMPL
I6100-03 - 1	102317-TIDO-F-U-B	-15	0.0184	-245.1		PPB	25 Oct 2017 11:57:23	U	SMPL
I6100-04 - 1	102317-TIDO-F-U-S	-8	0.0205	-245.1		PPB	25 Oct 2017 11:59:32	U	SMPL
PB103619BL - 1	PBW	-20	0.0168	-245.1		PPB	25 Oct 2017 12:01:40	U	SMPL
PB103619BS - 1	LCSW	13803	4.3212	-245.1		PPB	25 Oct 2017 12:03:47	U	SMPL
CCV04 - 1	CCV04	15433	4.8288	-245.1		PPB	25 Oct 2017 12:05:54	U	SMPL
CCB04 - 1	CCB04	-161	-0.0271	-245.1		PPB	25 Oct 2017 12:08:01	U	SMPL
I6083-01 - 1	102017-TIDO-F-D-B	-53	0.0065	-245.1		PPB	25 Oct 2017 12:10:10	U	SMPL
I6083-02 - 1	102017-TIDO-F-D-S	15	0.0277	-245.1		PPB	25 Oct 2017 12:12:17	U	SMPL
I6083-04 - 1	102017-TIDO-F-U-S	18	0.0286	-245.1		PPB	25 Oct 2017 12:16:31	U	SMPL
I6083-05 - 1	102017-TIFR-F-D-B	23	0.0302	-245.1		PPB	25 Oct 2017 12:18:39	U	SMPL
I6083-06 - 1	102017-TIFR-F-D-M	5	0.0246	-245.1		PPB	25 Oct 2017 12:20:46	U	SMPL
I6083-07 - 1	102017-TIFR-F-D-S	6	0.0249	-245.1		PPB	25 Oct 2017 12:22:55	U	SMPL
I6083-08 - 1	102017-TIFR-F-U-B	8	0.0255	-245.1		PPB	25 Oct 2017 12:25:03	U	SMPL
I6083-09 - 1	102017-TIFR-F-U-M	-18	0.0174	-245.1		PPB	25 Oct 2017 12:27:12	U	SMPL
I6083-10 - 1	102017-TIFR-F-U-S	-4	0.0218	-245.1		PPB	25 Oct 2017 12:29:19	U	SMPL
CCV05 - 1	CCV05	15944	4.9879	-245.1		PPB	25 Oct 2017 12:31:26	U	SMPL
CCB05 - 1	CCB05	-137	-0.0196	-245.1		PPB	25 Oct 2017 12:33:33	U	SMPL
I6083-11 - 1	102017-SIDI-F-D-B	17	0.0283	-245.1		PPB	25 Oct 2017 12:35:41	U	SMPL
I6083-12 - 1	102017-SIDI-F-D-M	7	0.0252	-245.1		PPB	25 Oct 2017 12:37:48	U	SMPL
I6083-13 - 1	102017-SIDI-F-D-S	32	0.0330	-245.1		PPB	25 Oct 2017 12:39:55	U	SMPL
I6083-14 - 1	102017-SIDI-F-U-B	-29	0.0140	-245.1		PPB	25 Oct 2017 12:42:03	U	SMPL
I6083-15 - 1	102017-SIDI-F-U-M	13	0.0271	-245.1		PPB	25 Oct 2017 12:44:10	U	SMPL
I6083-16 - 1	102017-SIDI-F-U-S	-17	0.0177	-245.1		PPB	25 Oct 2017 12:46:18	U	SMPL
I6084-01 - 1	102117-TIDO-F-D-B	16	0.0280	-245.1		PPB	25 Oct 2017 12:48:26	U	SMPL
I6084-02 - 1	102117-TIDO-F-D-S	17	0.0283	-245.1		PPB	25 Oct 2017 12:50:34	U	SMPL
I6084-03 - 1	102117-TIDO-F-U-B	28	0.0318	-245.1		PPB	25 Oct 2017 12:52:43	U	SMPL
I6084-04 - 1	102117-TIDO-F-U-S	29	0.0321	-245.1		PPB	25 Oct 2017 12:54:50	U	SMPL
CCV06 - 1	CCV06	16316	5.1038	-245.1		PPB	25 Oct 2017 12:56:58	U	SMPL
CCB06 - 1	CCB06	-123	-0.0153	-245.1		PPB	25 Oct 2017 12:59:05	U	SMPL
I6084-04DUP - 1	102117-TIDO-F-U-SDUP	-5	0.0215	-245.1		PPB	25 Oct 2017 13:01:14	U	SMPL
I6084-04MS - 1	102117-TIDO-F-U-SMS	15532	4.8596	-245.1		PPB	25 Oct 2017 13:03:21	U	SMPL
I6084-04MSD - 1	102117-TIDO-F-U-SMSD	14847	4.6463	-245.1		PPB	25 Oct 2017 13:05:29	U	SMPL
I6083-03 - 1	1020147-TIDO-F-U-B	-20	0.0168	-245.1		PPB	25 Oct 2017 13:07:36	U	SMPL
I6024-25LX5 - 1		3	0.0240	-245.1		PPB	25 Oct 2017 13:09:44	U	SMPL
I6024-25A - 1		14688	4.5968	-245.1		PPB	25 Oct 2017 13:11:52	U	SMPL
I6084-04LX5 - 1		-124	-0.0156	-245.1		PPB	25 Oct 2017 13:14:00	U	SMPL
I6084-04A - 1		15659	4.8992	-245.1		PPB	25 Oct 2017 13:16:09	U	SMPL
CCV07 - 1	CCV07	15390	4.8154	-245.1		PPB	25 Oct 2017 13:18:18	U	SMPL
CCB07 - 1	CCB07	-120	-0.0143	-245.1		PPB	25 Oct 2017 13:20:25	U	SMPL