

7470A


LB91077

INSTRUMENT ID: CV2

Linear

 μ Abs.:

35193



A= 0.0000e+000

B= 2.8761e-004

C= 4.1241e-003

Rho= 0.9993741

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	-14	0.000	-14				
0.2	0.200	0.212	0.012	723	0.0 %	723				
2.5	2.500	2.503	0.003	8690	0.0 %	8690				
5.0	5.000	5.124	0.124	17803	0.0 %	17803				
7.5	7.500	7.234	-0.266	25138	0.0 %	25138				
10.0	10.000	10.126	0.126	35193	0.0 %	35193				

LB91077 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Oct 2017 14:11:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-14	-	0.00007470A	PPB		25 Oct 2017 14:15:38	S	Std
0.2 - 1	50.2	723	-	0.20007470A	PPB		25 Oct 2017 14:17:47	S	Std
2.5 - 1	52.5	8690	-	2.50007470A	PPB		25 Oct 2017 14:19:55	S	Std
5.0 - 1	55	17803	-	5.00007470A	PPB		25 Oct 2017 14:22:03	S	Std
7.5 - 1	57.5	25138	-	7.50007470A	PPB		25 Oct 2017 14:24:11	S	Std
10.0 - 1	510	35193	-	10.00007470A	PPB		25 Oct 2017 14:26:19	S	Std
ICV29 - 1	ICV29	13220	3.8063	-7470A	PPB		25 Oct 2017 14:28:54	U	SMPL
ICB29 - 1	ICB29	-142	-0.0367	-7470A	PPB		25 Oct 2017 14:31:00	U	SMPL
CCV08 - 1	CCV08	17936	5.1627	-7470A	PPB		25 Oct 2017 14:33:08	U	SMPL
CCB08 - 1	CCB08	-189	-0.0502	-7470A	PPB		25 Oct 2017 14:35:14	U	SMPL
CRA - 1	CRA	726	0.2129	-7470A	PPB		25 Oct 2017 14:37:21	U	SMPL
HighStd - 1	HighStd	36124	10.3937	-7470A	PPB		25 Oct 2017 14:39:28	U	SMPL
ChkStd - 1	ChkStd	24346	7.0062	-7470A	PPB		25 Oct 2017 14:41:35	U	SMPL
PB103620BL - 1	PBW	-240	-0.0649	-7470A	PPB		25 Oct 2017 14:43:44	U	SMPL
PB103620BS - 1	LCSW	12534	3.6090	-7470A	PPB		25 Oct 2017 14:45:53	U	SMPL
I6076-03 - 1	FB-2017-10-19	-141	-0.0364	-7470A	PPB		25 Oct 2017 14:48:00	U	SMPL
I6090-01 - 1	FADW29217L	361	0.1080	-7470A	PPB		25 Oct 2017 14:54:41	U	SMPL
I6090-01DUP - 1	FADW29217LDUP	370	0.1105	-7470A	PPB		25 Oct 2017 14:56:49	U	SMPL
I6090-01MS - 1	FADW29217LMS	12787	3.6818	-7470A	PPB		25 Oct 2017 14:58:58	U	SMPL
I6090-01MSD - 1	FADW29217LMSD	12592	3.6257	-7470A	PPB		25 Oct 2017 15:01:04	U	SMPL
CCV09 - 1	CCV09	17965	5.1710	-7470A	PPB		25 Oct 2017 15:03:12	U	SMPL
CCB09 - 1	CCB09	-175	-0.0462	-7470A	PPB		25 Oct 2017 15:05:18	U	SMPL
I6111-01 - 1	MARION-26KV-3-WATER	201310	57.9026	-7470A	PPB		25 Oct 2017 15:07:26	U	SMPL
I6111-03 - 1	MARION-26KV-3-WATER	-89	-0.0215	-7470A	PPB		25 Oct 2017 15:12:21	U	SMPL
I6118-01 - 1	FRAC-TANK-102417	-30	-0.0045	-7470A	PPB		25 Oct 2017 15:14:28	U	SMPL
PB103628BL - 1	PBW	-22	-0.0022	-7470A	PPB		25 Oct 2017 15:16:35	U	SMPL
PB103628BS - 1	LCSW	13528	3.8949	-7470A	PPB		25 Oct 2017 15:18:42	U	SMPL
PB103628TB - 1	PB103574TB	-123	-0.0313	-7470A	PPB		25 Oct 2017 15:20:50	U	SMPL
I6080-15 - 1	SB-7-FILL(4-7)	-24	-0.0028	-7470A	PPB		25 Oct 2017 15:22:58	U	SMPL
I6080-16 - 1	SB-7-NATIVE(8-11)	5	0.0056	-7470A	PPB		25 Oct 2017 15:25:07	U	SMPL
I6081-18 - 1	SB-16-FILL(1-3)	12	0.0076	-7470A	PPB		25 Oct 2017 15:27:15	U	SMPL
I6081-19 - 1	SB-16-NATIVE(4-6)	16	0.0087	-7470A	PPB		25 Oct 2017 15:29:22	U	SMPL
CCV10 - 1	CCV10	18127	5.2176	-7470A	PPB		25 Oct 2017 15:31:28	U	SMPL
CCB10 - 1	CCB10	-145	-0.0376	-7470A	PPB		25 Oct 2017 15:33:35	U	SMPL
I6081-20 - 1	SB-9-NATIVE(10-11)	5	0.0056	-7470A	PPB		25 Oct 2017 15:35:43	U	SMPL
I6085-07 - 1	ROLLOFF-28-20	0	0.0041	-7470A	PPB		25 Oct 2017 15:37:50	U	SMPL
I6090-02 - 1	FADW29217L	-1	0.0038	-7470A	PPB		25 Oct 2017 15:39:56	U	SMPL
I6091-03 - 1	HD-01-102317	-6	0.0024	-7470A	PPB		25 Oct 2017 15:42:03	U	SMPL
I6091-03DUP - 1	HD-01-102317DUP	-17	-0.0008	-7470A	PPB		25 Oct 2017 15:44:11	U	SMPL
I6091-03MS - 1	HD-01-102317MS	13643	3.9280	-7470A	PPB		25 Oct 2017 15:46:19	U	SMPL
I6091-03MSD - 1	HD-01-102317MSD	13836	3.9835	-7470A	PPB		25 Oct 2017 15:48:27	U	SMPL
CCV11 - 1	CCV11	17822	5.1299	-7470A	PPB		25 Oct 2017 15:50:35	U	SMPL
CCB11 - 1	CCB11	-198	-0.0528	-7470A	PPB		25 Oct 2017 15:52:44	U	SMPL
I6090-01LX5 - 1		28	0.0122	-7470A	PPB		25 Oct 2017 15:54:52	U	SMPL
I6090-01A - 1		13154	3.7873	-7470A	PPB		25 Oct 2017 15:56:59	U	SMPL
I6091-03LX5 - 1		-129	-0.0330	-7470A	PPB		25 Oct 2017 15:59:06	U	SMPL
I6091-03A - 1		13722	3.9507	-7470A	PPB		25 Oct 2017 16:01:14	U	SMPL
I6111-01DLX10 - 1		20183	5.8089	-7470A	PPB		25 Oct 2017 16:03:21	U	SMPL
CCV12 - 1	CCV12	18032	5.1903	-7470A	PPB		25 Oct 2017 16:05:28	U	SMPL
CCB12 - 1	CCB12	-165	-0.0433	-7470A	PPB		25 Oct 2017 16:07:36	U	SMPL