



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91086

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	10/25/2017	09:15
2	I6024-25		1	>100.0	>212.0	10/25/2017	09:50
3	I6024-25DUP		1	>100.0	>212.0	10/25/2017	10:12
4	I6024-26		1	>100.0	>212.0	10/25/2017	10:35
5	I6034-10		1	>100.0	>212.0	10/25/2017	11:05
6	I6090-02		1	>100.0	>212.0	10/25/2017	11:40
7	I6111-02		1	>100.0	>212.0	10/25/2017	12:05
8	I6118-02		1	>100.0	>212.0	10/25/2017	12:35

Result = (Celsius * 1.8) + 32

LB91086

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP102517

WorkList ID : 104677

Date : 10/25/2017 08:30AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/26/2017	Water	I6024-25	Flash Point	Cool 4 deg C	LIRO01	M41	TWP-02	10/19/2017	1010A
10/26/2017	Water	I6024-26	Flash Point	Cool 4 deg C	LIRO01	M41	TWP-13	10/19/2017	1010A
10/26/2017	Water	I6034-10	Flash Point	Cool 4 deg C	PSEG01	M52	BAKER-TANK	10/19/2017	1010A
10/26/2017	Water	I6090-02	Flash Point	Cool 4 deg C	PSEG01	D52	FADW29217L	10/23/2017	1010A
10/27/2017	Water	I6111-02	Flash Point	Cool 4 deg C	PSEG01	D52	MARION-26KV-3-WATER	10/24/2017	1010A
10/27/2017	Water	I6118-02	Flash Point		PSEG01		FRAC-TANK-102417	10/24/2017	1010A

Date/Time 10/25/17 8:35AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/25/17 12:45PM
 Received by: CP
 Relinquished by: KS