


7471A

LB91089
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42345



A= 0.0000e+000

B= 2.3727e-004

C= 2.9290e-002

Rho= 0.9996280

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.038	0.038	38	0.000	38				
0.2	0.200	0.237	0.037	874	0.0 %	874				
2.5	2.500	2.550	0.050	10625	0.0 %	10625				
5.0	5.000	4.778	-0.222	20015	0.0 %	20015				
7.5	7.500	7.520	0.020	31571	0.0 %	31571	0			
10.0	10.000	10.076	0.076	42345	0.0 %	42345				

LB91089 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 25 Oct 2017 16:57:11

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	38	-	0.00007471A	PPB	25 Oct 2017 16:59:26	S	Std
0.2 - 1	50.2	874	-	0.20007471A	PPB	25 Oct 2017 17:01:44	S	Std
2.5 - 1	52.5	10625	-	2.50007471A	PPB	25 Oct 2017 17:04:02	S	Std
5.0 - 1	55	20015	-	5.00007471A	PPB	25 Oct 2017 17:06:19	S	Std
7.5 - 1	57.5	31571	-	7.50007471A	PPB	25 Oct 2017 17:08:37	S	Std
10.0 - 1	50	42345	-	10.00007471A	PPB	25 Oct 2017 17:10:54	S	Std
ICV30 - 1	ICV30	16171	3.8662	-7471A	PPB	25 Oct 2017 17:13:41	U	SMPL
ICB30 - 1	ICB30	-188	-0.0153	-7471A	PPB	25 Oct 2017 17:15:57	U	SMPL
CCV13 - 1	CCV13	19867	4.7431	-7471A	PPB	25 Oct 2017 17:18:13	U	SMPL
CCB13 - 1	CCB13	-172	-0.0115	-7471A	PPB	25 Oct 2017 17:20:28	U	SMPL
CRA - 1	CRA	850	0.2310	-7471A	PPB	25 Oct 2017 17:22:45	U	SMPL
HighStd - 1	HighStd	41995	9.9934	-7471A	PPB	25 Oct 2017 17:25:01	U	SMPL
ChkStd - 1	ChkStd	27944	6.6595	-7471A	PPB	25 Oct 2017 17:27:16	U	SMPL
PB103621BL - 1	PBS	-256	-0.0315	-7471A	PPB	25 Oct 2017 17:29:35	U	SMPL
PB103621BS - 1	LCSS	16370	3.9134	-7471A	PPB	25 Oct 2017 17:34:24	U	SMPL
I6076-01 - 1	SB-4R(5.5-6.0)	4357	1.0631	-7471A	PPB	25 Oct 2017 17:36:41	U	SMPL
I6076-02 - 1	SB-1R(5.5-6.0)	3249	0.8002	-7471A	PPB	25 Oct 2017 17:38:58	U	SMPL
I6076-04 - 1	SB-2R(5.5-6.0)	4626	1.1269	-7471A	PPB	25 Oct 2017 17:41:16	U	SMPL
I6076-05 - 1	SB-2R(5.5-6.0)-DUP	775	0.2132	-7471A	PPB	25 Oct 2017 17:43:34	U	SMPL
I6085-01 - 1	ROLLOFF-28-20	1197	0.3133	-7471A	PPB	25 Oct 2017 17:45:50	U	SMPL
CCV14 - 1	CCV14	20096	4.7974	-7471A	PPB	25 Oct 2017 17:48:06	U	SMPL
CCB14 - 1	CCB14	-171	-0.0113	-7471A	PPB	25 Oct 2017 17:50:22	U	SMPL
I6091-01 - 1	HD-01-102317	3908	0.9565	-7471A	PPB	25 Oct 2017 17:52:39	U	SMPL
I6091-01DUP - 1	HD-01-102317DUP	3606	0.8849	-7471A	PPB	25 Oct 2017 17:54:55	U	SMPL
I6091-01MS - 1	HD-01-102317MS	18130	4.3310	-7471A	PPB	25 Oct 2017 17:57:11	U	SMPL
I6091-01MSD - 1	HD-01-102317MSD	17990	4.2978	-7471A	PPB	25 Oct 2017 17:59:28	U	SMPL
I6093-01 - 1	TEC-B6-B7-COMP-0-4	4924	1.1976	-7471A	PPB	25 Oct 2017 18:01:44	U	SMPL
I6093-04 - 1	TEC-B6-B7-COMP-4-12	1134	0.2984	-7471A	PPB	25 Oct 2017 18:04:01	U	SMPL
I6093-07 - 1	TEC-B8-B9-COMP-1-4	2403	0.5994	-7471A	PPB	25 Oct 2017 18:06:19	U	SMPL
I6093-08 - 1	TEC-B8-B9-COMP-4-12	447	0.1353	-7471A	PPB	25 Oct 2017 18:08:36	U	SMPL
PB103622BL - 1	PBS	-29	0.0224	-7471A	PPB	25 Oct 2017 18:10:54	U	SMPL
PB103622BS - 1	LCSS	13438	3.2177	-7471A	PPB	25 Oct 2017 18:13:10	U	SMPL
CCV15 - 1	CCV15	19901	4.7512	-7471A	PPB	25 Oct 2017 18:15:27	U	SMPL
CCB15 - 1	CCB15	-155	-0.0075	-7471A	PPB	25 Oct 2017 18:17:43	U	SMPL
I6099-01 - 1	ERM-P-1	1417	0.3655	-7471A	PPB	25 Oct 2017 18:20:00	U	SMPL
I6099-02 - 1	ERM-P-2	1875	0.4742	-7471A	PPB	25 Oct 2017 18:22:17	U	SMPL
I6099-03 - 1	ERM-FS-1	3959	0.9686	-7471A	PPB	25 Oct 2017 18:24:33	U	SMPL
I6099-04 - 1	ERM-FS-2	2487	0.6194	-7471A	PPB	25 Oct 2017 18:26:49	U	SMPL
I6099-05 - 1	ERM-FS-2-FD	2199	0.5510	-7471A	PPB	25 Oct 2017 18:29:06	U	SMPL
I6109-01 - 1	S13-SP1-C	64900	15.4280	-7471A	PPB	25 Oct 2017 18:31:23	U	SMPL
I6112-01 - 1	1-MARION26KV-(2-3)	13961	3.3418	-7471A	PPB	25 Oct 2017 18:33:40	U	SMPL
I6112-05 - 1	2-MARION26KV-(2-3)	615	0.1752	-7471A	PPB	25 Oct 2017 18:36:00	U	SMPL
I6112-09 - 1	3-MARION26KV-(2-3)	17928	4.2830	-7471A	PPB	25 Oct 2017 18:38:18	U	SMPL
I6112-13 - 1	4-MARION26KV-(2-3)	14632	3.5010	-7471A	PPB	25 Oct 2017 18:40:34	U	SMPL
CCV16 - 1	CCV16	20051	4.7868	-7471A	PPB	25 Oct 2017 18:42:51	U	SMPL
CCB16 - 1	CCB16	-172	-0.0115	-7471A	PPB	25 Oct 2017 18:45:07	U	SMPL
I6112-17 - 1	5-MARION26KV-(2-3)	14887	3.5615	-7471A	PPB	25 Oct 2017 18:47:24	U	SMPL
I6112-21 - 1	6-MARION26KV-(2-3)	25372	6.0493	-7471A	PPB	25 Oct 2017 18:49:41	U	SMPL
I6113-01 - 1	SU-01-102417	758	0.2091	-7471A	PPB	25 Oct 2017 18:51:57	U	SMPL
I6117-01 - 1	JC-01-102417	6849	1.6543	-7471A	PPB	25 Oct 2017 18:54:15	U	SMPL
I6117-01DUP - 1	JC-01-102417DUP	6307	1.5257	-7471A	PPB	25 Oct 2017 18:56:32	U	SMPL
I6117-01MS - 1	JC-01-102417MS	23498	5.6046	-7471A	PPB	25 Oct 2017 18:58:49	U	SMPL
I6117-01MSD - 1	JC-01-102417MSD	23689	5.6499	-7471A	PPB	25 Oct 2017 19:01:06	U	SMPL
CCV17 - 1	CCV17	19909	4.7531	-7471A	PPB	25 Oct 2017 19:03:25	U	SMPL
CCB17 - 1	CCB17	-140	-0.0039	-7471A	PPB	25 Oct 2017 19:05:44	U	SMPL
I6091-01LX5 - 1		730	0.2025	-7471A	PPB	25 Oct 2017 19:08:00	U	SMPL
I6091-01A - 1		22010	5.2516	-7471A	PPB	25 Oct 2017 19:10:17	U	SMPL
6117-01LX5 - 1		1148	0.3017	-7471A	PPB	25 Oct 2017 19:12:33	U	SMPL
6117-01A - 1		23189	5.5313	-7471A	PPB	25 Oct 2017 19:14:49	U	SMPL
6109-01DLX5 - 1		12066	2.8922	-7471A	PPB	25 Oct 2017 19:17:05	U	SMPL
CCV18 - 1	CCV18	20174	4.8159	-7471A	PPB	25 Oct 2017 19:19:22	U	SMPL
CCB18 - 1	CCB18	-156	-0.0077	-7471A	PPB	25 Oct 2017 19:21:39	U	SMPL