



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91096

Temp.: 20.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	10/25/2017	16:10
2	CAL2	1	Water			7.01	10/25/2017	16:11
3	CAL3	1	Water			10.03	10/25/2017	16:13
4	ICV	1	Water			7.00	10/25/2017	16:14
5	CCV1	1	Water			2.01	10/25/2017	16:15
6	I6143-01	1	Water			6.39	10/25/2017	16:16
7	I6143-02	1	Water			6.33	10/25/2017	16:17
8	I6143-06	1	Water			6.50	10/25/2017	16:19
9	I6143-06DUP	1	Water			6.51	10/25/2017	16:21
10	CCV2	1	Water			12.01	10/25/2017	16:23

1391096

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6143

WorkList ID : 104683

Date : 10/25/2017 3:54:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2017	Water	I6143-01 -12	pH	Cool 4 deg C	CHAZ02	E62	RLF-GW-7S	10/24/2017	SM 4500-PH B
10/24/2017	Water	I6143-02 -12	pH	Cool 4 deg C	CHAZ02	E62	RLF-GW-9S	10/24/2017	SM 4500-PH B
10/24/2017	Water	I6143-06 -12	pH	Cool 4 deg C	CHAZ02	E62	RLF-FD-01	10/24/2017	SM 4500-PH B

Date/Time 10/25/17 1:00 PM
Received by: JD
Relinquished by: CR

Date/Time 10/25/17 5:00 PM
Received by: CR
Relinquished by: JD