



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91100

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/25/2017	17:15
2	CAL2	1	Water			7.00	10/25/2017	17:16
3	CAL3	1	Water			10.04	10/25/2017	17:19
4	ICV	1	Water			6.99	10/25/2017	17:21
5	CCV1	1	Water			2.01	10/25/2017	17:22
6	I6138-04	1	Solid	20.02	20	5.10	10/25/2017	17:25
7	I6138-08	1	Solid	20.04	20	10.03	10/25/2017	17:26
8	I6141-03	1	Solid	20.01	20	7.90	10/25/2017	17:30
9	I6141-06	1	Solid	20.01	20	8.15	10/25/2017	17:31
10	I6146-03	1	Solid	20.02	20	6.64	10/25/2017	17:32
11	I6146-06	1	Solid	20.05	20	6.48	10/25/2017	17:34
12	I6146-09	1	Solid	20.03	20	7.36	10/25/2017	17:35
13	I6146-09DUP	1	Solid	20.05	20	7.37	10/25/2017	17:36
14	CCV2	1	Water			12.01	10/25/2017	17:37

2891100

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6141

WorkList ID : 104682

Date : 10/25/2017 3:52:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2017	Solid	I6138-04	Corrosivity	Cool 4 deg C	PSEG01	D52	220-3A-B-STONES	10/25/2017	9045C
10/29/2017	Solid	I6138-08	Corrosivity	Cool 4 deg C	PSEG01	D52	220-3A-B-CONCRETE	10/25/2017	9045C
10/29/2017	Solid	I6141-03	Corrosivity	Cool 4 deg C	PSEG01	D52	SP-1-A-COMP	10/25/2017	9045C
10/29/2017	Solid	I6141-06	Corrosivity	Cool 4 deg C	PSEG01	D52	SP-1-B-COMP	10/25/2017	9045C
10/29/2017	Solid	I6146-03	Corrosivity	Cool 4 deg C	PSEG05	D52	AU-06-102517-A	10/25/2017	9045C
10/29/2017	Solid	I6146-06	Corrosivity	Cool 4 deg C	PSEG05	D52	AU-06-102517-B	10/25/2017	9045C
10/29/2017	Solid	I6146-09	Corrosivity	Cool 4 deg C	PSEG05	D52	AU-06-102517-C	10/25/2017	9045C

Date/Time 10/23/17 4:00PM

Received by: JP

Relinquished by: CB

Date/Time 10/25/17 6:35AM

Received by: CB

Relinquished by: JP