

7470A


LB 91117

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34324



A= 0.0000e+000

B= 2.9330e-004

C= -3.6059e-002

Rho= 0.9999517

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	27	0.000	27				
0.2	0.200	0.215	0.015	855	0.0 %	855				
2.5	2.500	2.525	0.025	8733	0.0 %	8733				
5.0	5.000	5.025	0.025	17256	0.0 %	17256				
7.5	7.500	7.432	-0.068	25462	0.0 %	25462				
10.0	10.000	10.031	0.031	34324	0.0 %	34324				

LB91117 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Oct 2017 11:10:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	27	-	0.0000	7470A	PPB	26 Oct 2017 11:12:29	S	Std
0.2 - 1	S0.2	855	-	0.2000	7470A	PPB	26 Oct 2017 11:14:37	S	Std
2.5 - 1	S2.5	8733	-	2.5000	7470A	PPB	26 Oct 2017 11:16:46	S	Std
5.0 - 1	S5	17256	-	5.0000	7470A	PPB	26 Oct 2017 11:18:54	S	Std
7.5 - 1	S7.5	25462	-	7.5000	7470A	PPB	26 Oct 2017 11:21:02	S	Std
10.0 - 1	S10	34324	-	10.0000	7470A	PPB	26 Oct 2017 11:23:09	S	Std
ICV31 - 1	ICV31	13594	3.9510	-	7470A	PPB	26 Oct 2017 11:25:43	U	SMPL
ICB31 - 1	ICB31	-147	-0.0792	-	7470A	PPB	26 Oct 2017 11:27:50	U	SMPL
CCV19 - 1	CCV19	17636	5.1366	-	7470A	PPB	26 Oct 2017 11:29:56	U	SMPL
CCB19 - 1	CCB19	-132	-0.0748	-	7470A	PPB	26 Oct 2017 11:32:03	U	SMPL
CRA - 1	CRA	832	0.2080	-	7470A	PPB	26 Oct 2017 11:34:10	U	SMPL
HighStd - 1	HighStd	34824	10.1778	-	7470A	PPB	26 Oct 2017 11:36:17	U	SMPL
ChkStd - 1	ChkStd	23522	6.8629	-	7470A	PPB	26 Oct 2017 11:38:24	U	SMPL
PB103658BL - 1	PBW	-20	-0.0419	-	7470A	PPB	26 Oct 2017 11:42:58	U	SMPL
PB103658BS - 1	LCBW	13561	3.9414	-	7470A	PPB	26 Oct 2017 11:45:05	U	SMPL
PB103658TB - 1	PB103612TB	-120	-0.0713	-	7470A	PPB	26 Oct 2017 11:47:13	U	SMPL
I6101-01 - 1	CB-1017-S-TCLP-48S(3+4)	54	-0.0202	-	7470A	PPB	26 Oct 2017 11:49:21	U	SMPL
I6101-02 - 1	CB-1017-S-TCLP-48S(1+2)	224	0.0296	-	7470A	PPB	26 Oct 2017 11:51:29	U	SMPL
I6101-03 - 1	CB-1017-S-TCLP-43S3	20	-0.0302	-	7470A	PPB	26 Oct 2017 11:53:37	U	SMPL
I6101-04 - 1	CB-1017-S-TCLP-43S2	37	-0.0252	-	7470A	PPB	26 Oct 2017 11:55:44	U	SMPL
CCV20 - 1	CCV20	17369	5.0582	-	7470A	PPB	26 Oct 2017 11:57:50	U	SMPL
CCB20 - 1	CCB20	-153	-0.0809	-	7470A	PPB	26 Oct 2017 11:59:57	U	SMPL
I6101-05 - 1	CB-1017-S-TCLP-43S1	25	-0.0287	-	7470A	PPB	26 Oct 2017 12:02:04	U	SMPL
I6111-02 - 1	MARION-26KV-3-WATER	32	-0.0267	-	7470A	PPB	26 Oct 2017 12:04:11	U	SMPL
I6112-04 - 1	1-MARION26KV-(2-3)	11	-0.0328	-	7470A	PPB	26 Oct 2017 12:06:18	U	SMPL
I6112-08 - 1	2-MARION26KV-(2-3)	6	-0.0343	-	7470A	PPB	26 Oct 2017 12:08:25	U	SMPL
I6112-12 - 1	3-MARION26KV-(2-3)	6	-0.0343	-	7470A	PPB	26 Oct 2017 12:10:32	U	SMPL
I6112-16 - 1	4-MARION26KV-(2-3)	21	-0.0299	-	7470A	PPB	26 Oct 2017 12:12:40	U	SMPL
I6112-20 - 1	5-MARION26KV-(2-3)	26	-0.0284	-	7470A	PPB	26 Oct 2017 12:14:48	U	SMPL
I6112-24 - 1	6-MARION26KV-(2-3)	22	-0.0296	-	7470A	PPB	26 Oct 2017 12:16:56	U	SMPL
I6113-03 - 1	SU-01-102417	21	-0.0299	-	7470A	PPB	26 Oct 2017 12:19:05	U	SMPL
I6117-03 - 1	JC-01-102417	23	-0.0293	-	7470A	PPB	26 Oct 2017 12:21:12	U	SMPL
CCV21 - 1	CCV21	17316	5.0427	-	7470A	PPB	26 Oct 2017 12:23:19	U	SMPL
CCB21 - 1	CCB21	-144	-0.0783	-	7470A	PPB	26 Oct 2017 12:25:26	U	SMPL
I6117-03DUP - 1	JC-01-102417DUP	12	-0.0325	-	7470A	PPB	26 Oct 2017 12:27:34	U	SMPL
I6117-03MS - 1	JC-01-102417MS	12598	3.6589	-	7470A	PPB	26 Oct 2017 12:29:41	U	SMPL
I6117-03MSD - 1	JC-01-102417MSD	12599	3.6592	-	7470A	PPB	26 Oct 2017 12:34:15	U	SMPL
I6118-02 - 1	FRAC-TANK-102417	-115	-0.0698	-	7470A	PPB	26 Oct 2017 12:36:21	U	SMPL
I6117-03LX5 - 1		-6	-0.0378	-	7470A	PPB	26 Oct 2017 12:38:29	U	SMPL
I6117-03A - 1		13749	3.9965	-	7470A	PPB	26 Oct 2017 12:40:37	U	SMPL
CCV22 - 1	CCV22	17512	5.1002	-	7470A	PPB	26 Oct 2017 12:42:45	U	SMPL
CCB22 - 1	CCB22	-149	-0.0798	-	7470A	PPB	26 Oct 2017 12:44:54	U	SMPL