


ISM02.4

LB 91124
INSTRUMENT ID: CV2

Linear

 μ Abs.:

45480



A= 0.0000e+000

B= 2.2050e-004

C= -1.4176e-002

Rho= 0.9999966

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	0.007	0.007	98	0.000	98					1
0.2	0.200	0.205	0.005	992	0.0 %	992					2
2.5	2.500	2.492	-0.008	11365	0.0 %	11365					0
5.0	5.000	4.995	-0.005	22718	0.0 %	22718					0
7.5	7.500	7.487	-0.013	34016	0.0 %	34016					0
10.0	10.000	10.014	0.014	45480	0.0 %	45480					0

NB
10/26/12

LB91124
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	98	-		0 ISM02.4	PPB	10/26/2017 14:36	Std
	0.2 S0.2	992	-		0.2 ISM02.4	PPB	10/26/2017 14:38	Std
	2.5 S2.5	11365	-		2.5 ISM02.4	PPB	10/26/2017 14:40	Std
	5 S5	22718	-		5 ISM02.4	PPB	10/26/2017 14:43	Std
	7.5 S7.5	34016	-		7.5 ISM02.4	PPB	10/26/2017 14:45	Std
	10 S10	45480	-		10 ISM02.4	PPB	10/26/2017 14:47	Std
ICV	ICV	17853	3.9225 -		ISM02.4	PPB	10/26/2017 14:50	SMPL
ICB	ICB	-122	-0.0411 -		ISM02.4	PPB	10/26/2017 14:52	SMPL
CCV023	CCV023	22172	4.8749 -		ISM02.4	PPB	10/26/2017 14:55	SMPL
CCB023	CCB023	-134	-0.0437 -		ISM02.4	PPB	10/26/2017 14:57	SMPL
PB103662BL	PBS004	321	0.0566 -		ISM02.4	PPB	10/26/2017 14:59	SMPL
I5941-01	MBDXP8	20220	4.4444 -		ISM02.4	PPB	10/26/2017 15:01	SMPL
I5941-02	MBDXQ4	7875	1.7223 -		ISM02.4	PPB	10/26/2017 15:04	SMPL
I5941-03	MBDXN3	542	0.1053 -		ISM02.4	PPB	10/26/2017 15:06	SMPL
I5941-04	MBDXN9	684	0.1366 -		ISM02.4	PPB	10/26/2017 15:08	SMPL
I5941-05	MBDXP2	557	0.1086 -		ISM02.4	PPB	10/26/2017 15:10	SMPL
I5941-07	MBDXP2D	595	0.117 -		ISM02.4	PPB	10/26/2017 15:13	SMPL
I5941-08	MBDXP2S	12313	2.7009 -		ISM02.4	PPB	10/26/2017 15:15	SMPL
CCV024	CCV024	22481	4.943 -		ISM02.4	PPB	10/26/2017 15:51	SMPL
CCB024	CCB024	-192	-0.0565 -		ISM02.4	PPB	10/26/2017 15:53	SMPL