

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91128

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59121
calibration std. phosphate 0.5 ppm	WP59120
calibration std. phosphate 0.3 ppm	WP59119
calibration std. 0 ppm	WP59116
phosphate ICV std.	WP59123
phosphate CCV std.	WP59122
calibration std. phosphate 0.1 ppm	WP59118
calibration std. phosphate 0.05 ppm	WP59117
Combined reagent	WP59125
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0008Slope: 0.6506Regression: 0.99981

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/26/2017	13:30
2	CAL2	0.05	1	50	50	0.033	10/26/2017	13:30
3	CAL3	0.10	1	50	50	0.065	10/26/2017	13:31
4	CAL4	0.30	1	50	50	0.186	10/26/2017	13:31
5	CAL5	0.50	1	50	50	0.331	10/26/2017	13:32
6	CAL6	1.00	1	50	50	0.649	10/26/2017	13:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91128

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	10/26/2017	13:33
2	ICB		1	50	50	0.000	0.001	10/26/2017	13:33
3	CCV1	0.50	1	50	50	0.330	0.508	10/26/2017	13:34
4	CCB1		1	50	50	0.000	0.001	10/26/2017	13:34
5	PB103698BL		1	50	50	0.000	0.001	10/26/2017	13:35
6	PB103698BS	0.50	1	50	50	0.327	0.504	10/26/2017	13:35
7	I6139-01		1	50	50	0.018	0.029	10/26/2017	13:36
8	I6139-02		1	50	50	0.039	0.061	10/26/2017	13:36
9	I6140-01		1	50	50	0.146	0.226	10/26/2017	13:37
10	I6143-01		1	50	50	0.021	0.034	10/26/2017	13:37
11	I6143-02		1	50	50	0.038	0.060	10/26/2017	13:38
12	I6143-06		1	50	50	0.040	0.063	10/26/2017	13:38
13	I6143-06DUP		1	50	50	0.038	0.060	10/26/2017	13:39
14	I6143-06MS	0.50	1	50	50	0.341	0.525	10/26/2017	13:39
15	CCV2	0.50	1	50	50	0.330	0.508	10/26/2017	13:40
16	CCB2		1	50	50	0.000	0.001	10/26/2017	13:40
17	I6143-06MSD	0.50	1	50	50	0.345	0.532	10/26/2017	13:41
18	I6160-01		1	50	50	0.046	0.072	10/26/2017	13:41
19	I6160-02		1	50	50	0.041	0.064	10/26/2017	13:42
20	I6160-03		1	50	50	0.053	0.083	10/26/2017	13:42
21	I6160-04		1	50	50	0.045	0.070	10/26/2017	13:43
22	I6160-05		1	50	50	0.009	0.015	10/26/2017	13:43
23	I6160-06		1	50	50	0.008	0.014	10/26/2017	13:44
24	CCV3	0.50	1	50	50	0.328	0.505	10/26/2017	13:44
25	CCB3		1	50	50	0.000	0.001	10/26/2017	13:45