



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: AMANDEEP

Parameter: Color

Supervisor Review By: apatel

Run Number: LB91131

Reagent/Standard	Lot/Log #
30 color units	WP59067
25 color units	WP59066
20 color units	WP59065
15 color units	WP59064
5 color units	WP59063
70 color units	WP59072
60 color units	WP59071
50 color units	WP59070
40 color units	WP59069
35 color units	WP59068
0 color units	WP59061
10 color units	WP59062

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB91131BL	LB91131BL	1	5.00	N/A	0	10/26/2017	14:02
2	I6160-01	RLF-SW-1	1	7.84	N/A	10	10/26/2017	14:10
3	I6160-01DUP	RLF-SW-1DUP	1	7.84	N/A	10	10/26/2017	14:15
4	I6160-02	RLF-SW-3	1	7.85	N/A	10	10/26/2017	14:19
5	I6160-03	RLF-SW-4	1	7.70	N/A	15	10/26/2017	14:24
6	I6160-04	RLF-SW-5	1	7.42	N/A	15	10/26/2017	14:29
7	I6160-05	RLF-SW-6	1	7.04	N/A	20	10/26/2017	14:35
8	I6160-06	RLF-FD-02	1	7.06	N/A	20	10/26/2017	14:40

LB91131

WORKLIST(Hardcopy Internal Chain)

WorkList Name : color102617

WorkList ID : 104737

Date : 10/26/2017 1:09:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/27/2017	Water	I6160-01	Color	Cool 4 deg C	CHAZ02	F12	RLF-SW-1	10/25/2017	SM2120 B
10/27/2017	Water	I6160-02	Color	Cool 4 deg C	CHAZ02	F12	RLF-SW-3	10/25/2017	SM2120 B
10/27/2017	Water	I6160-03	Color	Cool 4 deg C	CHAZ02	F12	RLF-SW-4	10/25/2017	SM2120 B
10/27/2017	Water	I6160-04	Color	Cool 4 deg C	CHAZ02	F12	RLF-SW-5	10/25/2017	SM2120 B
10/27/2017	Water	I6160-05	Color	Cool 4 deg C	CHAZ02	F12	RLF-SW-6	10/25/2017	SM2120 B
10/27/2017	Water	I6160-06	Color	Cool 4 deg C	CHAZ02	F12	RLF-FD-02	10/25/2017	SM2120 B

Date/Time

Received by:

Relinquished by:

10/26/17 13:40

AUC



Date/Time

Received by:

Relinquished by:

10/26/17 15:10

