



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91134

Temp.: 20.2 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/26/2017	16:15
2	CAL2	1	Water			7.00	10/26/2017	16:16
3	CAL3	1	Water			10.03	10/26/2017	16:18
4	ICV	1	Water			7.00	10/26/2017	16:20
5	CCV1	1	Water			2.02	10/26/2017	16:23
6	I6118-02	1	Water			7.56	10/26/2017	16:25
7	I6153-01	1	Water			6.24	10/26/2017	16:26
8	I6166-05	1	Water			6.89	10/26/2017	16:28
9	I6166-05DUP	1	Water			6.90	10/26/2017	16:29
10	CCV2	1	Water			12.01	10/26/2017	16:30

1391134

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6160 WorkList ID : 104739

Date : 10/26/2017 2:46:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2017	Water	I6118-02-A	Corrosivity		PSEG01		FRAC-TANK-102417	10/24/2017	9040C
10/25/2017	Water	I6153-01-C	Corrosivity	Cool 4 deg C	WOOD09	F12	WC-102517	10/25/2017	9040C
10/26/2017	Water	I6166-05 - C	Corrosivity	Cool 4 deg C	PSEG01	D52	EXCAVATION-WATER	10/26/2017	9040C

Date/Time 10/26/17 3:00 PM
Received by: TH
Relinquished by: SD

Date/Time 10/26/17 4:30 PM
Received by: SD
Relinquished by: JP