


7471A

LB 91135
INSTRUMENT ID: CV2

Linear

 μ Abs.:

44542



A= 0.0000e+000

B= 2.2966e-004

C= 1.1947e-002

Rho= 0.9987073

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	54	0.000	54				
0.2	0.200	0.237	0.037	981	0.0 %	981				
2.5	2.500	2.565	0.065	11117	0.0 %	11117				
5.0	5.000	5.014	0.014	21780	0.0 %	21780				
7.5	7.500	7.118	-0.382	30939	0.0 %	30939	0			
10.0	10.000	10.242	0.242	44542	0.0 %	44542				

LB91135 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 26 Oct 2017 17:24:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.00007471A	PPB		26 Oct 2017 17:25:58	S	Std
0.2 - 1	50.2	981	-	0.20007471A	PPB		26 Oct 2017 17:28:16	S	Std
2.5 - 1	52.5	11117	-	2.50007471A	PPB		26 Oct 2017 17:30:34	S	Std
5.0 - 1	55	21780	-	5.00007471A	PPB		26 Oct 2017 17:32:51	S	Std
7.5 - 1	57.5	30939	-	7.50007471A	PPB		26 Oct 2017 17:35:09	S	Std
10.0 - 1	510	44542	-	10.00007471A	PPB		26 Oct 2017 17:41:38	S	Std
ICV32 - 1	ICV32	17175	3.9564	-7471A	PPB		26 Oct 2017 17:44:29	U	SMPL
ICB32 - 1	ICB32	-113	-0.0140	-7471A	PPB		26 Oct 2017 17:46:45	U	SMPL
CCV25 - 1	CCV25	22242	5.1202	-7471A	PPB		26 Oct 2017 17:49:01	U	SMPL
CCB25 - 1	CCB25	-133	-0.0186	-7471A	PPB		26 Oct 2017 17:51:16	U	SMPL
CRA - 1	CRA	982	0.2375	-7471A	PPB		26 Oct 2017 17:53:33	U	SMPL
HighStd - 1	HighStd	45404	10.4396	-7471A	PPB		26 Oct 2017 17:55:49	U	SMPL
ChkStd - 1	ChkStd	29590	6.8077	-7471A	PPB		26 Oct 2017 17:58:05	U	SMPL
PB103664BL - 1	PBS	-131	-0.0181	-7471A	PPB		26 Oct 2017 18:00:24	U	SMPL
PB103664BS - 1	LCSS	15098	3.4794	-7471A	PPB		26 Oct 2017 18:02:42	U	SMPL
I6138-01 - 1	220-3A-B-STONES	1554	0.3688	-7471A	PPB		26 Oct 2017 18:04:59	U	SMPL
I6138-05 - 1	220-3A-B-CONCRETE	452	0.1158	-7471A	PPB		26 Oct 2017 18:07:16	U	SMPL
I6141-01 - 1	SP-1-A-COMP	2523	0.5914	-7471A	PPB		26 Oct 2017 18:09:34	U	SMPL
I6141-04 - 1	SP-1-B-COMP	2141	0.5037	-7471A	PPB		26 Oct 2017 18:11:52	U	SMPL
I6146-01 - 1	AU-06-102517-A	1553	0.3686	-7471A	PPB		26 Oct 2017 18:14:08	U	SMPL
CCV26 - 1	CCV26	22081	5.0832	-7471A	PPB		26 Oct 2017 18:16:24	U	SMPL
CCB26 - 1	CCB26	-128	-0.0175	-7471A	PPB		26 Oct 2017 18:18:40	U	SMPL
I6146-04 - 1	AU-06-102517-B	1892	0.4465	-7471A	PPB		26 Oct 2017 18:20:57	U	SMPL
I6146-04DUP - 1	AU-06-102517-BDUP	1578	0.3744	-7471A	PPB		26 Oct 2017 18:23:13	U	SMPL
I6146-04MS - 1	AU-06-102517-BMS	17165	3.9541	-7471A	PPB		26 Oct 2017 18:25:29	U	SMPL
I6146-04MSD - 1	AU-06-102517-BMSD	16914	3.8965	-7471A	PPB		26 Oct 2017 18:27:45	U	SMPL
I6146-07 - 1	AU-06-102517-C	938	0.2274	-7471A	PPB		26 Oct 2017 18:30:03	U	SMPL
I6146-04LX5 - 1		371	0.0972	-7471A	PPB		26 Oct 2017 18:32:20	U	SMPL
I6146-04A - 1		18632	4.2911	-7471A	PPB		26 Oct 2017 18:34:37	U	SMPL
CCV27 - 1	CCV27	21826	5.0246	-7471A	PPB		26 Oct 2017 18:36:55	U	SMPL
CCB27 - 1	CCB27	-139	-0.0200	-7471A	PPB		26 Oct 2017 18:39:12	U	SMPL