



### Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91138

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/26/2017	17:20
2	CAL2	1	Water			7.01	10/26/2017	17:21
3	CAL3	1	Water			10.04	10/26/2017	17:22
4	ICV	1	Water			6.99	10/26/2017	17:25
5	CCV1	1	Water			2.01	10/26/2017	17:26
6	I6153-03	1	Solid	20.02	20	9.34	10/26/2017	17:29
7	I6166-03	1	Solid	20.05	20	8.51	10/26/2017	17:30
8	I6169-02	1	Solid	20.02	20	8.78	10/26/2017	17:31
9	I6171-02	1	Solid	20.03	20	7.81	10/26/2017	17:33
10	I6171-04	1	Solid	20.01	20	10.15	10/26/2017	17:35
11	I6171-06	1	Solid	20.03	20	8.82	10/26/2017	17:36
12	I6171-08	1	Solid	20.03	20	8.38	10/26/2017	17:37
13	I6171-10	1	Solid	20.05	20	8.20	10/26/2017	17:38
14	I6171-10DUP	1	Solid	20.03	20	8.21	10/26/2017	17:39
15	CCV2	1	Water			12.02	10/26/2017	17:40

2091138

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6166

WorkList ID : 104742

Date : 10/26/2017 3:19:38 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2017	Solid	I6153-03	Corrosivity	Cool 4 deg C	WOOD09	F12	WC-102517	10/25/2017	9045C
10/31/2017	Solid	I6166-03	Corrosivity	Cool 4 deg C	PSEG01	D52	RO-72-11966	10/26/2017	9045C
10/31/2017	Solid	I6169-02	Corrosivity	Cool 4 deg C	PSEG05	D52	SU-04-102617	10/26/2017	9045C
10/31/2017	Solid	I6171-02	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-01-102617-A	10/26/2017	9045C
10/31/2017	Solid	I6171-04	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-01-102617-B	10/26/2017	9045C
10/31/2017	Solid	I6171-06	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-01-102617-C	10/26/2017	9045C
10/31/2017	Solid	I6171-08	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-01-102617-D	10/26/2017	9045C
10/31/2017	Solid	I6171-10	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-01-102617-E	10/26/2017	9045C

Date/Time 10/26/17 3:20PM  
 Received by: JH  
 Relinquished by: [Signature]

Date/Time 10/26/17 3:19PM  
 Received by: [Signature]  
 Relinquished by: JH