

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: DHVANI

Date: 10/27/2017

DATE & TIME IN: 10/27/2017 11:00

OVEN TEMP: 103 °C

Run Number: LB91146

DATE & TIME OUT: 10/27/2017 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB91146BL	LB91146BL	1.4003	1.4003	100	1.4003	1.4003	1.4003	0	0
2	LB91146BS	LB91146BS	1.4018	1.4018	100	1.4573	1.4573	1.4573	0.0555	555
3	I6137-01	COMP	1.4179	1.4179	200	1.4405	1.4405	1.4405	0.0226	113
4	I6139-01	STORMDRAIN-1	1.4245	1.4245	900	1.4338	1.4338	1.4338	0.0093	10.3
5	I6139-02	STORMDRAIN-2	1.423	1.423	900	1.4359	1.4359	1.4359	0.0129	14.3
6	I6140-01	DSN001	1.4415	1.4415	200	1.4535	1.4535	1.4535	0.012	60
7	I6160-01	RLF-SW-1	1.4357	1.4357	400	1.4494	1.4494	1.4494	0.0137	34.3
8	I6160-02	RLF-SW-3	1.4303	1.4303	300	1.441	1.441	1.441	0.0107	35.7
9	I6160-03	RLF-SW-4	1.4229	1.4229	500	1.4242	1.4242	1.4242	0.0013	2.6
10	I6160-04	RLF-SW-5	1.4258	1.4258	500	1.4269	1.4269	1.4269	0.0011	2.2
11	I6160-05	RLF-SW-6	1.4174	1.4174	400	1.4228	1.4228	1.4228	0.0054	13.5
12	I6160-06	RLF-FD-02	1.3883	1.3883	400	1.3949	1.3949	1.3949	0.0066	16.5
13	I6166-04	EXCAVATION-WATER	1.4267	1.4267	100	1.4465	1.4465	1.4465	0.0198	198
14	I6166-04DUP	EXCAVATION-WATERDUP	1.3959	1.3959	100	1.4164	1.4164	1.4164	0.0205	205
15	I6168-01	SPDES-1	1.4201	1.4201	900	1.4202	1.4202	1.4202	0.0001	0.1

