


7470A

LB91151
INSTRUMENT ID: CX2

Linear

 μ Abs.:

34848



A= 0.0000e+000

B= 2.8913e-004

C= 1.1984e-002

Rho= 0.9998500

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	48	0.000	48				
0.2	0.200	0.213	0.013	694	0.0 %	694				
2.5	2.500	2.541	0.041	8746	0.0 %	8746				
5.0	5.000	4.901	-0.099	16911	0.0 %	16911				
7.5	7.500	7.432	-0.068	25663	0.0 %	25663				
10.0	10.000	10.088	0.088	34848	0.0 %	34848				

LB91151 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Oct 2017 12:10:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	48	-	0.0000	7470A	PPB	27 Oct 2017 12:14:04	S	Std
0.2 - 1	S0.2	694	-	0.2000	7470A	PPB	27 Oct 2017 12:16:13	S	Std
2.5 - 1	S2.5	8746	-	2.5000	7470A	PPB	27 Oct 2017 12:18:21	S	Std
5.0 - 1	S5	16911	-	5.0000	7470A	PPB	27 Oct 2017 12:20:29	S	Std
7.5 - 1	S7.5	25663	-	7.5000	7470A	PPB	27 Oct 2017 12:22:37	S	Std
10.0 - 1	S10	34848	-	10.0000	7470A	PPB	27 Oct 2017 12:24:45	S	Std
ICV33 - 1	ICV33	13243	3.8409	-	7470A	PPB	27 Oct 2017 12:27:37	U	SMPL
ICB33 - 1	ICB33	-120	-0.0227	-	7470A	PPB	27 Oct 2017 12:29:43	U	SMPL
CCV31 - 1	CCV31	16953	4.9136	-	7470A	PPB	27 Oct 2017 12:31:50	U	SMPL
CCB31 - 1	CCB31	-131	-0.0259	-	7470A	PPB	27 Oct 2017 12:33:56	U	SMPL
CRA - 1	CRA	703	0.2152	-	7470A	PPB	27 Oct 2017 12:36:03	U	SMPL
HighStd - 1	HighStd	34640	10.0274	-	7470A	PPB	27 Oct 2017 12:38:10	U	SMPL
ChkStd - 1	ChkStd	24383	7.0618	-	7470A	PPB	27 Oct 2017 12:40:17	U	SMPL
PB103702BL - 1	PBW	-206	-0.0476	-	7470A	PPB	27 Oct 2017 12:42:26	U	SMPL
PB103702BS - 1	LCSW	13524	3.9221	-	7470A	PPB	27 Oct 2017 12:44:35	U	SMPL
PB103702TB - 1	PB103660TB	-90	-0.0140	-	7470A	PPB	27 Oct 2017 12:46:43	U	SMPL
I6109-03 - 1	S13-SP1-C	2	0.0126	-	7470A	PPB	27 Oct 2017 12:48:51	U	SMPL
I6109-03DUP - 1	S13-SP1-CDUP	20	0.0178	-	7470A	PPB	27 Oct 2017 12:51:00	U	SMPL
I6109-03MS - 1	S13-SP1-CMS	13187	3.8247	-	7470A	PPB	27 Oct 2017 12:53:08	U	SMPL
I6109-03MSD - 1	S13-SP1-CMSD	13270	3.8487	-	7470A	PPB	27 Oct 2017 12:55:15	U	SMPL
CCV32 - 1	CCV32	16989	4.9240	-	7470A	PPB	27 Oct 2017 12:57:23	U	SMPL
CCB32 - 1	CCB32	-163	-0.0351	-	7470A	PPB	27 Oct 2017 12:59:30	U	SMPL
I6138-04 - 1	220-3A-B-STONES	25	0.0192	-	7470A	PPB	27 Oct 2017 13:01:38	U	SMPL
I6138-08 - 1	220-3A-B-CONCRETE	30	0.0207	-	7470A	PPB	27 Oct 2017 13:03:45	U	SMPL
I6141-03 - 1	SP-1-A-COMP	44	0.0247	-	7470A	PPB	27 Oct 2017 13:05:52	U	SMPL
I6141-06 - 1	SP-1-B-COMP	33	0.0215	-	7470A	PPB	27 Oct 2017 13:07:59	U	SMPL
I6146-03 - 1	AU-06-102517-A	30	0.0207	-	7470A	PPB	27 Oct 2017 13:10:06	U	SMPL
I6146-06 - 1	AU-06-102517-B	40	0.0235	-	7470A	PPB	27 Oct 2017 13:12:14	U	SMPL
I6146-09 - 1	AU-06-102517-C	25	0.0192	-	7470A	PPB	27 Oct 2017 13:14:22	U	SMPL
I6109-03LX5 - 1		36	0.0224	-	7470A	PPB	27 Oct 2017 13:16:30	U	SMPL
I6109-03A - 1		14364	4.1650	-	7470A	PPB	27 Oct 2017 13:18:39	U	SMPL
CCV33 - 1	CCV33	17065	4.9459	-	7470A	PPB	27 Oct 2017 13:20:46	U	SMPL
CCB33 - 1	CCB33	-158	-0.0337	-	7470A	PPB	27 Oct 2017 13:22:53	U	SMPL