


ISM02.4

LB 9/163
INSTRUMENT ID: CV1

 μ Abs.:

41635



Linear

A= 0.0000e+000

B= 2.4150e-004

C= -2.4918e-002

Rho= 0.9999642

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.005	0.005	123	0.000	123					-
0.2	0.200	0.189	-0.011	887	0.0 %	887					-6
2.5	2.500	2.511	0.011	10500	0.0 %	10500					0
5.0	5.000	5.028	0.028	20921	0.0 %	20921					1
7.5	7.500	7.437	-0.063	30899	0.0 %	30899					-1
10.0	10.000	10.030	0.030	41635	0.0 %	41635					0

NB
10/27/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	123	-		0 ISM02.4	PPB	10/27/2017 16:15	Std
	0.2 S0.2	887	-		0.2 ISM02.4	PPB	10/27/2017 16:17	Std
	2.5 S2.5	10500	-		2.5 ISM02.4	PPB	10/27/2017 16:20	Std
	5 S5	20921	-		5 ISM02.4	PPB	10/27/2017 16:22	Std
	7.5 S7.5	30899	-		7.5 ISM02.4	PPB	10/27/2017 16:24	Std
	10 S10	41635	-		10 ISM02.4	PPB	10/27/2017 16:26	Std
ICV	ICV	16667	4.0002 -		ISM02.4	PPB	10/27/2017 16:29	SMPL
ICB	ICB	-105	-0.0503 -		ISM02.4	PPB	10/27/2017 16:31	SMPL
CCV034	CCV034	20916	5.0264 -		ISM02.4	PPB	10/27/2017 16:34	SMPL
CCB034	CCB034	-110	-0.0515 -		ISM02.4	PPB	10/27/2017 16:36	SMPL
PB103691BL	PBW003	16	-0.0211 -		ISM02.4	PPB	10/27/2017 16:38	SMPL
I5941-06	MBDY24	37	-0.016 -		ISM02.4	PPB	10/27/2017 16:41	SMPL
PB103692BL	PBW003	113	0.0024 -		ISM02.4	PPB	10/27/2017 16:43	SMPL
I6032-04	TWP-B-11	9549	2.2812 -		ISM02.4	PPB	10/27/2017 16:45	SMPL
I6032-05	TWP-B-20	1068	0.233 -		ISM02.4	PPB	10/27/2017 16:47	SMPL
I6032-06	TWP-B-31	1471	0.3303 -		ISM02.4	PPB	10/27/2017 16:50	SMPL
I6120-01	TWP-B-28	457	0.0854 -		ISM02.4	PPB	10/27/2017 16:52	SMPL
I6120-03	TWP-B-34	2	-0.0244 -		ISM02.4	PPB	10/27/2017 16:54	SMPL
I6120-05	EQB-5-102417	-47	-0.0363 -		ISM02.4	PPB	10/27/2017 16:56	SMPL
CCV035	CCV035	19997	4.8044 -		ISM02.4	PPB	10/27/2017 17:15	SMPL
CCB035	CCB035	-65	-0.0406 -		ISM02.4	PPB	10/27/2017 17:17	SMPL