


7471A

LB91173

INSTRUMENT ID: LV1

 μ Abs.:

42121



Linear

A= 0.0000e+000

B= 2.3647e-004

C= 7.7092e-003

Rho= 0.9995903

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	25	0.000	25				
0.2	0.200	0.219	0.019	892	0.0 %	892				
2.5	2.500	2.550	0.050	10752	0.0 %	10752				
5.0	5.000	4.799	-0.201	20262	0.0 %	20262				
7.5	7.500	7.650	0.150	32318	0.0 %	32318	0			
10.0	10.000	9.968	-0.032	42121	0.0 %	42121				

LB91173 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Oct 2017 18:30:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.00007471A	PPB		27 Oct 2017 18:31:40	S	Std
0.2 - 1	50.2	892	-	0.20007471A	PPB		27 Oct 2017 18:33:59	S	Std
2.5 - 1	52.5	10752	-	2.50007471A	PPB		27 Oct 2017 18:36:16	S	Std
5.0 - 1	55	20262	-	5.00007471A	PPB		27 Oct 2017 18:38:34	S	Std
7.5 - 1	57.5	32318	-	7.50007471A	PPB		27 Oct 2017 18:40:51	S	Std
10.0 - 1	510	42121	-	10.00007471A	PPB		27 Oct 2017 18:46:45	S	Std
ICV34 - 1	ICV34	16636	3.9417	-7471A	PPB		27 Oct 2017 18:49:21	U	SMPL
ICB34 - 1	ICB34	-174	-0.0334	-7471A	PPB		27 Oct 2017 18:51:37	U	SMPL
CCV36 - 1	CCV36	19681	4.6618	-7471A	PPB		27 Oct 2017 18:53:53	U	SMPL
CCB36 - 1	CCB36	-171	-0.0327	-7471A	PPB		27 Oct 2017 18:56:09	U	SMPL
CRA - 1	CRA	871	0.2137	-7471A	PPB		27 Oct 2017 18:58:26	U	SMPL
HighStd - 1	HighStd	41748	9.8800	-7471A	PPB		27 Oct 2017 19:00:42	U	SMPL
ChkStd - 1	ChkStd	28223	6.6817	-7471A	PPB		27 Oct 2017 19:02:58	U	SMPL
PB103713BL - 1	PBS	-226	-0.0457	-7471A	PPB		27 Oct 2017 19:05:16	U	SMPL
PB103713BS - 1	LCSS	16225	3.8445	-7471A	PPB		27 Oct 2017 19:10:18	U	SMPL
I6119-03 - 1	EPE-120-3(0-5)	5001	1.1903	-7471A	PPB		27 Oct 2017 19:12:34	U	SMPL
I6119-03DUP - 1	EPE-120-3(0-5)DUP	4326	1.0307	-7471A	PPB		27 Oct 2017 19:14:52	U	SMPL
I6119-03MS - 1	EPE-120-3(0-5)MS	22905	5.4241	-7471A	PPB		27 Oct 2017 19:17:10	U	SMPL
I6119-03MSD - 1	EPE-120-3(0-5)MSD	22111	5.2364	-7471A	PPB		27 Oct 2017 19:22:02	U	SMPL
I6119-04 - 1	EPE-120-3(5-10)	453	0.1148	-7471A	PPB		27 Oct 2017 19:24:18	U	SMPL
CCV37 - 1	CCV37	20915	4.9536	-7471A	PPB		27 Oct 2017 19:26:34	U	SMPL
CCB37 - 1	CCB37	-163	-0.0308	-7471A	PPB		27 Oct 2017 19:28:50	U	SMPL
I6157-01 - 1	RLF-SD-1	1052	0.2565	-7471A	PPB		27 Oct 2017 19:31:07	U	SMPL
I6157-02 - 1	RLF-SD-3	1411	0.3414	-7471A	PPB		27 Oct 2017 19:33:23	U	SMPL
I6157-03 - 1	RLF-SD-4	1653	0.3986	-7471A	PPB		27 Oct 2017 19:35:39	U	SMPL
I6157-04 - 1	RLF-SD-5	1978	0.4755	-7471A	PPB		27 Oct 2017 19:37:55	U	SMPL
I6157-05 - 1	RLF-SD-6	1012	0.2470	-7471A	PPB		27 Oct 2017 19:40:12	U	SMPL
I6157-06 - 1	RLF-FD-02	963	0.2354	-7471A	PPB		27 Oct 2017 19:42:29	U	SMPL
I6166-01 - 1	RO-72-11966	16863	3.9954	-7471A	PPB		27 Oct 2017 19:44:47	U	SMPL
I6169-01 - 1	SU-04-102617	1671	0.4029	-7471A	PPB		27 Oct 2017 19:47:04	U	SMPL
PB103714BL - 1	PBS	-11	0.0051	-7471A	PPB		27 Oct 2017 19:49:22	U	SMPL
PB103714BS - 1	LCSS	16523	3.9150	-7471A	PPB		27 Oct 2017 19:54:14	U	SMPL
CCV38 - 1	CCV38	20998	4.9732	-7471A	PPB		27 Oct 2017 19:56:30	U	SMPL
CCB38 - 1	CCB38	-122	-0.0211	-7471A	PPB		27 Oct 2017 19:58:47	U	SMPL
I6171-01 - 1	CL-01-102617-A	1188	0.2886	-7471A	PPB		27 Oct 2017 20:01:04	U	SMPL
I6171-03 - 1	CL-01-102617-B	6930	1.6465	-7471A	PPB		27 Oct 2017 20:03:20	U	SMPL
I6171-05 - 1	CL-01-102617-C	7380	1.7529	-7471A	PPB		27 Oct 2017 20:05:37	U	SMPL
I6171-05DUP - 1	CL-01-102617-CDUP	6190	1.4715	-7471A	PPB		27 Oct 2017 20:07:53	U	SMPL
I6171-05MS - 1	CL-01-102617-CMS	23438	5.5502	-7471A	PPB		27 Oct 2017 20:12:58	U	SMPL
I6171-05MSD - 1	CL-01-102617-CMSD	22119	5.2383	-7471A	PPB		27 Oct 2017 20:15:15	U	SMPL
I6171-07 - 1	CL-01-102617-D	4881	1.1619	-7471A	PPB		27 Oct 2017 20:17:33	U	SMPL
I6171-09 - 1	CL-01-102617-E	5226	1.2435	-7471A	PPB		27 Oct 2017 20:19:51	U	SMPL
I6178-01 - 1	EPE-120-3(10-12)	679	0.1683	-7471A	PPB		27 Oct 2017 20:22:09	U	SMPL
I6178-02 - 1	EPE-120-3(20-22)	735	0.1815	-7471A	PPB		27 Oct 2017 20:24:26	U	SMPL
CCV39 - 1	CCV39	20955	4.9630	-7471A	PPB		27 Oct 2017 20:26:42	U	SMPL
CCB39 - 1	CCB39	-158	-0.0297	-7471A	PPB		27 Oct 2017 20:28:58	U	SMPL
I6178-03 - 1	EPE-120-3(25-27)	93	0.0297	-7471A	PPB		27 Oct 2017 20:31:16	U	SMPL
I6178-04 - 1	EPE-120-3(35-37)	95	0.0302	-7471A	PPB		27 Oct 2017 20:33:32	U	SMPL
I6178-05 - 1	EPE-120-3(40-42)	82	0.0271	-7471A	PPB		27 Oct 2017 20:35:49	U	SMPL
I6178-06 - 1	EPE-M-6(0-5)	4463	1.0631	-7471A	PPB		27 Oct 2017 20:38:05	U	SMPL
I6178-07 - 1	EPE-M-6(5-10)	248	0.0664	-7471A	PPB		27 Oct 2017 20:40:22	U	SMPL
I6119-03LX5 - 1		850	0.2087	-7471A	PPB		27 Oct 2017 20:42:39	U	SMPL
I6119-03A - 1		21292	5.0427	-7471A	PPB		27 Oct 2017 20:44:57	U	SMPL
I6171-05LX5 - 1		1212	0.2943	-7471A	PPB		27 Oct 2017 20:47:14	U	SMPL
I6171-05A - 1		23724	5.6178	-7471A	PPB		27 Oct 2017 20:49:34	U	SMPL
CCV40 - 1	CCV40	20857	4.9398	-7471A	PPB		27 Oct 2017 20:51:50	U	SMPL
CCB40 - 1	CCB40	-169	-0.0323	-7471A	PPB		27 Oct 2017 20:54:07	U	SMPL