



Analytical Summary Report

Analysis Method:	SM4500-O G	Reviewed By:	Indiana
Parameter:	Dissolved Oxygen	Supervisor Review By:	apatel
Run Number:	LB91178	MANGANOUS SULFATE SOLUTION:	W2191
BOD Water:	WP59174	Alkaline Iodide Azide:	W2241
Starch:	W2121	Sodium Thiosulfate, 0.025N:	W2252
Conc Sulfuric Acid:	W2151	pH Strips:	W2106
		Zero DO Std:	WP59175

Seq	LabID	ClientID	Sample Vol (ml)	Initial Reading (mL)	Final Reading (mL)	Diff (mg/L)	Result (mg/L)
1	WINKLER1	WINKLER1	300	0.00	8.40	8.40	8.45
2	WINKLER2	WINKLER2	300	8.40	16.90	8.50	

Calibrate DO meter with average reading : 8.45

Zero DO Reading : 0.10 Calibrated to : 0.00

Seq	LabID	ClientID	DL	Sample Vol (ml)	Temperature (°C)	Result (mg/L)	Anal Date	Anal Time
1	LB91178BL	LB91178BL	1	300	20.0	0.00	10/27/2017	13:00
2	I6179-12	SWL-17	1	300	18.9	8.77	10/27/2017	13:00
3	I6179-12DUP	SWL-17DUP	1	300	18.9	8.75	10/27/2017	13:01
4	I6179-13	SW16A	1	300	18.8	7.82	10/27/2017	13:01
5	I6179-14	SW13	1	300	18.7	7.32	10/27/2017	13:02
6	I6179-15	SW12A	1	300	18.8	8.05	10/27/2017	13:02
7	I6179-16	SW3A	1	300	19.0	7.00	10/27/2017	13:03
8	I6179-17	SW11	1	300	18.9	8.96	10/27/2017	13:03
9	I6179-18	SW6	1	300	18.7	9.14	10/27/2017	13:04
10	I6179-19	SW7	1	300	19.0	7.83	10/27/2017	13:04
11	I6179-20	SW5	1	300	18.9	8.59	10/27/2017	13:05
12	I6179-21	SW9	1	300	19.0	9.43	10/27/2017	13:05
13	I6179-22	SW14	1	300	18.8	8.25	10/27/2017	13:06