


7470A

LB 91181
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34263



A= 0.0000e+000

B= 2.9391e-004

C= -3.9865e-002

Rho= 0.9998199

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	14	0.000	14				
0.2	0.200	0.198	-0.002	809	0.0 %	809				
2.5	2.500	2.525	0.025	8725	0.0 %	8725				
5.0	5.000	5.107	0.107	17510	0.0 %	17510				
7.5	7.500	7.376	-0.124	25233	0.0 %	25233				
10.0	10.000	10.030	0.030	34263	0.0 %	34263				

LB91181 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Oct 2017 21:02:57

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	14	-	0.00007470A	PPB	27 Oct 2017 21:03:07	S	Std
0.2 - 1	50.2	809	-	0.20007470A	PPB	27 Oct 2017 21:05:16	S	Std
2.5 - 1	52.5	8725	-	2.50007470A	PPB	27 Oct 2017 21:07:24	S	Std
5.0 - 1	55	17510	-	5.00007470A	PPB	27 Oct 2017 21:09:33	S	Std
7.5 - 1	57.5	25233	-	7.50007470A	PPB	27 Oct 2017 21:11:40	S	Std
10.0 - 1	510	34263	-	10.00007470A	PPB	27 Oct 2017 21:13:48	S	Std
ICV35 - 1	ICV35	13419	3.9041	-7470A	PPB	27 Oct 2017 21:16:27	U	SMPL
ICB35 - 1	ICB35	-121	-0.0754	-7470A	PPB	27 Oct 2017 21:18:34	U	SMPL
CCV41 - 1	CCV41	17575	5.1256	-7470A	PPB	27 Oct 2017 21:20:40	U	SMPL
CCB41 - 1	CCB41	-147	-0.0831	-7470A	PPB	27 Oct 2017 21:22:47	U	SMPL
CRA - 1	CRA	789	0.1920	-7470A	PPB	27 Oct 2017 21:24:54	U	SMPL
HighStd - 1	HighStd	34026	9.9608	-7470A	PPB	27 Oct 2017 21:27:01	U	SMPL
ChkStd - 1	ChkStd	23833	6.9649	-7470A	PPB	27 Oct 2017 21:29:08	U	SMPL
PB103737BL - 1	PBW	-11	-0.0431	-7470A	PPB	27 Oct 2017 21:34:18	U	SMPL
PB103737BS - 1	LCSW	13460	3.9162	-7470A	PPB	27 Oct 2017 21:36:25	U	SMPL
PB103737TB - 1	PB103699TB	-139	-0.0807	-7470A	PPB	27 Oct 2017 21:38:32	U	SMPL
I6092-01 - 1	WYANDANCH-STA(1-8)-COMP	-21	-0.0460	-7470A	PPB	27 Oct 2017 21:40:41	U	SMPL
I6092-02 - 1	WYANDANCH-STA(9-12)-COMP	-12	-0.0434	-7470A	PPB	27 Oct 2017 21:42:49	U	SMPL
I6092-06 - 1	WELLWOOD-CROSSING-EAST(0-3)	-11	-0.0431	-7470A	PPB	27 Oct 2017 21:44:58	U	SMPL
I6092-07 - 1	WELLWOOD-CROSSING-WEST-COMP(0-3)	-15	-0.0443	-7470A	PPB	27 Oct 2017 21:47:04	U	SMPL
CCV42 - 1	CCV42	17495	5.1021	-7470A	PPB	27 Oct 2017 21:49:11	U	SMPL
CCB42 - 1	CCB42	-131	-0.0784	-7470A	PPB	27 Oct 2017 21:51:18	U	SMPL
I6094-01 - 1	NH-WEST-COMP(0-3)	1	-0.0396	-7470A	PPB	27 Oct 2017 21:53:25	U	SMPL
I6094-02 - 1	COMMACK-WEST-COMP(0-3)	-6	-0.0416	-7470A	PPB	27 Oct 2017 21:55:32	U	SMPL
I6094-03 - 1	COMMACK-EAST-COMP(0-3)	-15	-0.0443	-7470A	PPB	27 Oct 2017 21:57:39	U	SMPL
I6094-04 - 1	S-18THST-EAST-COMP(0-3)	3	-0.0390	-7470A	PPB	27 Oct 2017 21:59:45	U	SMPL
I6094-05 - 1	S-18THST-WEST-COMP(0-3)	-6	-0.0416	-7470A	PPB	27 Oct 2017 22:01:52	U	SMPL
I6094-06 - 1	ISLIP-EAST-COMP(0-3)	-25	-0.0472	-7470A	PPB	27 Oct 2017 22:04:00	U	SMPL
I6094-07 - 1	ISLIP-WEST-COMP(0-3)	11	-0.0366	-7470A	PPB	27 Oct 2017 22:06:08	U	SMPL
I6153-01 - 1	WC-102517	51	-0.0249	-7470A	PPB	27 Oct 2017 22:08:16	U	SMPL
I6153-03 - 1	WC-102517	39	-0.0284	-7470A	PPB	27 Oct 2017 22:10:24	U	SMPL
I6166-03 - 1	RO-72-11966	10	-0.0369	-7470A	PPB	27 Oct 2017 22:12:31	U	SMPL
CCV43 - 1	CCV43	17205	5.0169	-7470A	PPB	27 Oct 2017 22:14:37	U	SMPL
CCB43 - 1	CCB43	-182	-0.0934	-7470A	PPB	27 Oct 2017 22:16:44	U	SMPL
I6166-05 - 1	EXCAVATION-WATER	-2	-0.0405	-7470A	PPB	27 Oct 2017 22:18:52	U	SMPL
I6169-02 - 1	SU-04-102617	-2	-0.0405	-7470A	PPB	27 Oct 2017 22:20:59	U	SMPL
I6171-02 - 1	CL-01-102617-A	28	-0.0316	-7470A	PPB	27 Oct 2017 22:23:06	U	SMPL
I6171-04 - 1	CL-01-102617-B	0	-0.0399	-7470A	PPB	27 Oct 2017 22:25:13	U	SMPL
I6171-06 - 1	CL-01-102617-C	-8	-0.0422	-7470A	PPB	27 Oct 2017 22:27:22	U	SMPL
I6171-08 - 1	CL-01-102617-D	-4	-0.0410	-7470A	PPB	27 Oct 2017 22:29:30	U	SMPL
I6171-10 - 1	CL-01-102617-E	-10	-0.0428	-7470A	PPB	27 Oct 2017 22:31:38	U	SMPL
I6171-10DUP - 1	CL-01-102617-EDUP	-7	-0.0419	-7470A	PPB	27 Oct 2017 22:33:46	U	SMPL
I6171-10MS - 1	CL-01-102617-EMS	13148	3.8245	-7470A	PPB	27 Oct 2017 22:35:55	U	SMPL
I6171-10MSD - 1	CL-01-102617-EMSD	13120	3.8162	-7470A	PPB	27 Oct 2017 22:38:02	U	SMPL
CCV44 - 1	CCV44	17489	5.1003	-7470A	PPB	27 Oct 2017 22:40:09	U	SMPL
CCB44 - 1	CCB44	-157	-0.0860	-7470A	PPB	27 Oct 2017 22:42:16	U	SMPL
I6171-10LX5 - 1		-16	-0.0446	-7470A	PPB	27 Oct 2017 22:44:23	U	SMPL
I6171-10A - 1		13462	3.9168	-7470A	PPB	27 Oct 2017 22:46:31	U	SMPL
CCV45 - 1	CCV45	17619	5.1386	-7470A	PPB	27 Oct 2017 22:48:37	U	SMPL
CCB45 - 1	CCB45	-151	-0.0842	-7470A	PPB	27 Oct 2017 22:50:45	U	SMPL