


7470A

LB 91207
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35364



A= 0.0000e+000

B= 2.8361e-004

C= 1.9926e-002

Rho= 0.9999292

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	1	0.000	1				
0.2	0.200	0.226	0.026	725	0.0 %	725				
2.5	2.500	2.447	-0.053	8557	0.0 %	8557				
5.0	5.000	5.025	0.025	17649	0.0 %	17649				
7.5	7.500	7.432	-0.068	26136	0.0 %	26136				
10.0	10.000	10.050	0.050	35364	0.0 %	35364				

LB91207 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Oct 2017 15:06:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.00007470A	PPB		30 Oct 2017 15:10:28	S	Std
0.2 - 1	50.2	725	-	0.20007470A	PPB		30 Oct 2017 15:12:36	S	Std
2.5 - 1	52.5	8557	-	2.50007470A	PPB		30 Oct 2017 15:14:44	S	Std
5.0 - 1	55	17649	-	5.00007470A	PPB		30 Oct 2017 15:16:53	S	Std
7.5 - 1	57.5	26136	-	7.50007470A	PPB		30 Oct 2017 15:19:01	S	Std
10.0 - 1	510	35364	-	10.00007470A	PPB		30 Oct 2017 15:21:08	S	Std
ICV36 - 1	ICV36	13092	3.7330	-7470A	PPB		30 Oct 2017 15:23:43	U	SMPL
ICB36 - 1	ICB36	-137	-0.0189	-7470A	PPB		30 Oct 2017 15:25:50	U	SMPL
CCV46 - 1	CCV46	17283	4.9216	-7470A	PPB		30 Oct 2017 15:27:57	U	SMPL
CCB46 - 1	CCB46	-152	-0.0232	-7470A	PPB		30 Oct 2017 15:30:03	U	SMPL
CRA - 1	CRA	724	0.2253	-7470A	PPB		30 Oct 2017 15:32:10	U	SMPL
HighStd - 1	HighStd	35686	10.1409	-7470A	PPB		30 Oct 2017 15:34:17	U	SMPL
ChkStd - 1	ChkStd	24022	6.8329	-7470A	PPB		30 Oct 2017 15:36:24	U	SMPL
PB103784BL - 1	PBW	-217	-0.0416	-7470A	PPB		30 Oct 2017 15:38:33	U	SMPL
PB103784BS - 1	LCSW	13326	3.7993	-7470A	PPB		30 Oct 2017 15:43:09	U	SMPL
I6143-01 - 1	RLF-GW-7S	74	0.0409	-7470A	PPB		30 Oct 2017 15:45:16	U	SMPL
I6143-02 - 1	RLF-GW-9S	-68	0.0006	-7470A	PPB		30 Oct 2017 15:47:25	U	SMPL
I6143-06 - 1	RLF-FD-01	6	0.0216	-7470A	PPB		30 Oct 2017 15:49:32	U	SMPL
I6160-01 - 1	RLF-SW-1	-26	0.0126	-7470A	PPB		30 Oct 2017 15:51:41	U	SMPL
I6160-01DUP - 1	RLF-SW-1DUP	-15	0.0157	-7470A	PPB		30 Oct 2017 15:53:48	U	SMPL
CCV47 - 1	CCV47	17171	4.8898	-7470A	PPB		30 Oct 2017 15:55:54	U	SMPL
CCB47 - 1	CCB47	-139	-0.0195	-7470A	PPB		30 Oct 2017 15:58:01	U	SMPL
I6160-01MS - 1	RLF-SW-1MS	13743	3.9176	-7470A	PPB		30 Oct 2017 16:02:40	U	SMPL
I6160-01MSD - 1	RLF-SW-1MSD	13820	3.9395	-7470A	PPB		30 Oct 2017 16:04:46	U	SMPL
I6160-02 - 1	RLF-SW-3	-151	-0.0229	-7470A	PPB		30 Oct 2017 16:06:53	U	SMPL
I6160-03 - 1	RLF-SW-4	-73	-0.0008	-7470A	PPB		30 Oct 2017 16:09:00	U	SMPL
I6160-04 - 1	RLF-SW-5	-47	0.0066	-7470A	PPB		30 Oct 2017 16:11:08	U	SMPL
I6160-05 - 1	RLF-SW-6	-53	0.0049	-7470A	PPB		30 Oct 2017 16:13:16	U	SMPL
I6160-06 - 1	RLF-FD-02	-37	0.0094	-7470A	PPB		30 Oct 2017 16:15:24	U	SMPL
I6166-04 - 1	EXCAVATION-WATER	863	0.2647	-7470A	PPB		30 Oct 2017 16:17:32	U	SMPL
PB103790BL - 1	PBW	-9	0.0174	-7470A	PPB		30 Oct 2017 16:19:41	U	SMPL
PB103790BS - 1	LCSW	13347	3.8053	-7470A	PPB		30 Oct 2017 16:21:48	U	SMPL
CCV48 - 1	CCV48	16857	4.8008	-7470A	PPB		30 Oct 2017 16:23:54	U	SMPL
CCB48 - 1	CCB48	-139	-0.0195	-7470A	PPB		30 Oct 2017 16:26:02	U	SMPL
I6179-01 - 1	SWL-17	-3	0.0191	-7470A	PPB		30 Oct 2017 16:28:09	U	SMPL
I6179-01DUP - 1	SWL-17DUP	-17	0.0151	-7470A	PPB		30 Oct 2017 16:30:16	U	SMPL
I6179-01MS - 1	SWL-17MS	14147	4.0322	-7470A	PPB		30 Oct 2017 16:32:23	U	SMPL
I6179-01MSD - 1	SWL-17MSD	14272	4.0676	-7470A	PPB		30 Oct 2017 16:34:30	U	SMPL
I6179-02 - 1	SW16A	-118	-0.0135	-7470A	PPB		30 Oct 2017 16:36:37	U	SMPL
I6179-03 - 1	SW13	-39	0.0089	-7470A	PPB		30 Oct 2017 16:38:45	U	SMPL
I6179-04 - 1	SW12A	-44	0.0074	-7470A	PPB		30 Oct 2017 16:40:53	U	SMPL
I6179-05 - 1	SW3A	125	0.0554	-7470A	PPB		30 Oct 2017 16:43:01	U	SMPL
I6179-06 - 1	SW11	-38	0.0091	-7470A	PPB		30 Oct 2017 16:45:10	U	SMPL
I6179-07 - 1	SW6	-5	0.0185	-7470A	PPB		30 Oct 2017 16:47:17	U	SMPL
CCV49 - 1	CCV49	16930	4.8215	-7470A	PPB		30 Oct 2017 16:49:24	U	SMPL
CCB49 - 1	CCB49	-146	-0.0215	-7470A	PPB		30 Oct 2017 16:51:31	U	SMPL
I6179-08 - 1	SW7	-27	0.0123	-7470A	PPB		30 Oct 2017 16:53:39	U	SMPL
I6179-09 - 1	SW5	-50	0.0057	-7470A	PPB		30 Oct 2017 16:55:46	U	SMPL
I6179-10 - 1	SW9	-25	0.0128	-7470A	PPB		30 Oct 2017 16:57:53	U	SMPL
I6179-11 - 1	SW14	-3	0.0191	-7470A	PPB		30 Oct 2017 17:00:00	U	SMPL
I6179-12 - 1	SWL-17	-6	0.0182	-7470A	PPB		30 Oct 2017 17:02:08	U	SMPL
I6179-13 - 1	SW16A	-19	0.0145	-7470A	PPB		30 Oct 2017 17:04:16	U	SMPL
I6179-14 - 1	SW13	3	0.0208	-7470A	PPB		30 Oct 2017 17:06:24	U	SMPL
I6179-15 - 1	SW12A	-24	0.0131	-7470A	PPB		30 Oct 2017 17:08:32	U	SMPL
I6179-16 - 1	SW3A	-28	0.0120	-7470A	PPB		30 Oct 2017 17:10:40	U	SMPL
I6179-17 - 1	SW11	-23	0.0134	-7470A	PPB		30 Oct 2017 17:12:48	U	SMPL
CCV50 - 1	CCV50	17599	5.0112	-7470A	PPB		30 Oct 2017 17:14:55	U	SMPL
CCB50 - 1	CCB50	-138	-0.0192	-7470A	PPB		30 Oct 2017 17:17:02	U	SMPL
I6179-18 - 1	SW6	-41	0.0083	-7470A	PPB		30 Oct 2017 17:19:10	U	SMPL
I6179-19 - 1	SW7	-28	0.0120	-7470A	PPB		30 Oct 2017 17:21:17	U	SMPL
I6179-20 - 1	SW5	-28	0.0120	-7470A	PPB		30 Oct 2017 17:23:24	U	SMPL
I6179-21 - 1	SW9	-37	0.0094	-7470A	PPB		30 Oct 2017 17:25:31	U	SMPL
I6179-22 - 1	SW14	-2	0.0194	-7470A	PPB		30 Oct 2017 17:27:38	U	SMPL
PB103807BL - 1	PBW	13	0.0236	-7470A	PPB		30 Oct 2017 17:29:46	U	SMPL
PB103807BS - 1	LCSW	14174	4.0399	-7470A	PPB		30 Oct 2017 17:31:55	U	SMPL
PB103807TB - 1	PB103735TB	-109	-0.0110	-7470A	PPB		30 Oct 2017 17:34:02	U	SMPL
I6092-03 - 1	STRAIGHT-PATH-EAST-COMP(0-3)	0	0.0199	-7470A	PPB		30 Oct 2017 17:36:11	U	SMPL
I6092-04 - 1	STRAIGHT-PATH-WEST-COMP(0-3)	7	0.0219	-7470A	PPB		30 Oct 2017 17:38:18	U	SMPL
CCV51 - 1	CCV51	17618	5.0166	-7470A	PPB		30 Oct 2017 17:40:26	U	SMPL
CCB51 - 1	CCB51	-130	-0.0169	-7470A	PPB		30 Oct 2017 17:42:33	U	SMPL
I6092-05 - 1	PINELAWN-STA-COMP(1-4)	1	0.0202	-7470A	PPB		30 Oct 2017 17:44:41	U	SMPL
I6092-08 - 1	CARLLS-PATH-EAST-COMP(0-3)	13	0.0236	-7470A	PPB		30 Oct 2017 17:46:49	U	SMPL
I6092-09 - 1	CARLLS-PATH-WEST-COMP(0-3)	11	0.0230	-7470A	PPB		30 Oct 2017 17:48:56	U	SMPL
I6092-10 - 1	NH-CROSSING-EAST-COMP(0-3)	4	0.0211	-7470A	PPB		30 Oct 2017 17:51:03	U	SMPL
I6167-03 - 1	S13-SP2C	11	0.0230	-7470A	PPB		30 Oct 2017 17:53:11	U	SMPL
I6174-03 - 1	B7(0-10)COMP	10	0.0228	-7470A	PPB		30 Oct 2017 17:55:18	U	SMPL
I6181-02 - 1	WASTE	115	0.0525	-7470A	PPB		30 Oct 2017 17:57:27	U	SMPL
I6182-02 - 1	RT-3026	3	0.0208	-7470A	PPB		30 Oct 2017 17:59:35	U	SMPL
I6194-02 - 1	BU-01-102717	-10	0.0171	-7470A	PPB		30 Oct 2017 18:01:44	U	SMPL
I6196-02 - 1	OK-03-102717	10	0.0228	-7470A	PPB		30 Oct 2017 18:03:51	U	SMPL
CCV52 - 1	CCV52	16697	4.7554	-7470A	PPB		30 Oct 2017 18:05:59	U	SMPL
CCB52 - 1	CCB52	-134	-0.0181	-7470A	PPB		30 Oct 2017 18:08:07	U	SMPL
I6200-02 - 1	OR-01-102717	0	0.0199	-7470A	PPB		30 Oct 2017 18:10:14	U	SMPL
I6202-02 - 1	ROLL-OFF-2	40	0.0313	-7470A	PPB		30 Oct 2017 18:12:22	U	SMPL
I6206-02 - 1	EO-01-102717-A	-6	0.0182	-7470A	PPB		30 Oct 2017 18:14:30	U	SMPL
I6206-04 - 1	EO-01-102717-B	20	0.0256	-7470A	PPB		30 Oct 2017 18:16:38	U	SMPL
I6206-06 - 1	EO-01-102717-C	75	0.0412	-7470A	PPB		30 Oct 2017 18:18:45	U	SMPL
I6206-06DUP - 1	EO-01-102717-CDUP	12	0.0233	-7470A	PPB		30 Oct 2017 18:20:53	U	SMPL
I6206-06MS - 1	EO-01-102717-CMS	12933	3.6879	-7470A	PPB		30 Oct 2017 18:23:01	U	SMPL
I6206-06MSD - 1	EO-01-102717-CMSD	13129	3.7435	-7470A	PPB		30 Oct 2017 18:25:08	U	SMPL
CCV53 - 1	CCV53	17457	4.9710	-7470A	PPB		30 Oct 2017 18:27:17	U	SMPL
CCB53 - 1	CCB53	-143	-0.0206	-7470A	PPB		30 Oct 2017 18:29:24	U	SMPL
I6160-01LX5 - 1		-9	0.0174	-7470A	PPB		30 Oct 2017 18:31:32	U	SMPL

LB91207 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Oct 2017 15:06:18

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
I6160-01A - 1		14564	4.1505		-7470A	PPB	30 Oct 2017 18:33:40	U	SMPL
I6179-01LX5 - 1		-114	-0.0124		-7470A	PPB	30 Oct 2017 18:35:48	U	SMPL
I6179-01A - 1		15139	4.3135		-7470A	PPB	30 Oct 2017 18:37:56	U	SMPL
I6206-06LX5 - 1		-107	-0.0104		-7470A	PPB	30 Oct 2017 18:40:04	U	SMPL
I6206-06A - 1		13500	3.8487		-7470A	PPB	30 Oct 2017 18:42:12	U	SMPL
CCV54 - 1	CCV54	17033	4.8507		-7470A	PPB	30 Oct 2017 18:44:20	U	SMPL
CCB54 - 1	CCB54	-127	-0.0161		-7470A	PPB	30 Oct 2017 18:46:28	U	SMPL