



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91208

Temp.: 18.7 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/30/2017	14:00
2	CAL2	1	Water			7.01	10/30/2017	14:01
3	CAL3	1	Water			10.04	10/30/2017	14:02
4	ICV	1	Water			7.00	10/30/2017	14:05
5	CCV1	1	Water			2.02	10/30/2017	14:06
6	I6228-01	1	Water			4.55	10/30/2017	14:07
7	I6228-01DUP	1	Water			4.56	10/30/2017	14:09
8	I6229-01	1	Water			4.70	10/30/2017	14:10
9	I6230-01	1	Water			5.76	10/30/2017	14:13
10	CCV2	1	Water			12.01	10/30/2017	14:16

LB 91208

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i6228

WorkList ID : 104894

Date : 10/30/2017 1:49:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2017	Water	I6228-01-1A	pH	Cool 4 deg C	STER16	D51	EXCELL	10/24/2017	SM 4500-PH B
10/23/2017	Water	I6229-01-1A	pH	Cool 4 deg C	STER16	D51	MAINCELL	10/23/2017	SM 4500-PH B
10/25/2017	Water	I6230-01-1A	pH	Cool 4 deg C	STER06	D51	POOLWATER	10/25/2017	SM 4500-PH B

Date/Time 10/30/17 1:50 PM
Received by: Jo
Relinquished by: CR

Date/Time 10/30/17 8:25 AM
Received by: CR
Relinquished by: Jo