



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JigneshDate: 10/31/2017DATE & TIME IN: 10/30/2017 18:30OVEN TEMP: 98 °CRun Number: LB91215DATE & TIME OUT: 10/31/2017 12:55OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB91215BL	LB91215BL	96.3078	96.3078	100	96.3078	96.3078	96.3078	0	0
2	I6234-04	TW-5-DEP	95.5223	95.5223	100	95.6825	95.6825	95.6825	0.1602	1602
3	I6234-04DUP	TW-5-DEPDUP	83.5584	83.5584	100	83.726	83.726	83.726	0.1676	1676

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2B91215

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I6234

WorkList ID : 104908

Date : 10/30/2017 6:07:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2017	Water	I6234-04	TS	Cool 4 deg C	AKRF02	D51	TW-5-DEP	10/28/2017	SM2540 B

Date/Time 10/30/17 6:07 PM
Received by: Jn
Relinquished by:

Date/Time 10/30/17 5:00 PM
Received by: JP
Relinquished by: JP