



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91219

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 97.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/30/2017	17:10
2	CAL2	1	Water			7.01	10/30/2017	17:11
3	CAL3	1	Water			10.02	10/30/2017	17:12
4	ICV	1	Water			6.99	10/30/2017	17:15
5	CCV1	1	Water			2.01	10/30/2017	17:16
6	I6225-01	1	Solid	20.02	20	7.91	10/30/2017	17:19
7	I6225-01DUP	1	Solid	20.03	20	7.90	10/30/2017	17:20
8	I6225-02	1	Solid	20.05	20	7.90	10/30/2017	17:21
9	I6226-08	1	Solid	20.03	20	5.87	10/30/2017	17:22
10	I6226-16	1	Solid	20.01	20	5.89	10/30/2017	17:23
11	I6226-18	1	Solid	20.03	20	10.75	10/30/2017	17:25
12	I6231-02	1	Solid	20.03	20	8.99	10/30/2017	17:26
13	I6236-02	1	Solid	20.01	20	10.96	10/30/2017	17:28
14	I6239-01	1	Solid	20.03	20	6.72	10/30/2017	17:30
15	I6239-02	1	Solid	20.02	20	6.74	10/30/2017	17:31
16	CCV2	1	Water			12.01	10/30/2017	17:32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6225

WorkList ID : 104903

Date : 10/30/2017 4:45:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2017	Solid	I6225-01	Corrosivity	Cool 4 deg C	USAR03	F61	CB-1017-S-TCLP-47S2	10/27/2017	9045C
11/04/2017	Solid	I6225-02	Corrosivity	Cool 4 deg C	USAR03	F61	CB-1017-S-TCLP-47S1	10/27/2017	9045C
11/02/2017	Solid	I6226-08	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-1	10/28/2017	9045C
11/02/2017	Solid	I6226-16	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-2	10/28/2017	9045C
11/02/2017	Solid	I6226-18	Corrosivity	Cool 4 deg C	PSEG01	D52	T-1-CONCRETE	10/28/2017	9045C
11/02/2017	Solid	I6231-02	Corrosivity	Cool 4 deg C	PSEG05	D52	TR-04-103017	10/30/2017	9045C
11/02/2017	Solid	I6236-02	Corrosivity	Cool 4 deg C	PSEG01	D52	133	10/30/2017	9045C
11/02/2017	Solid	I6239-01	Corrosivity	Cool 4 deg C	PSEG01	D52	DRUM-1-4	10/30/2017	9045C
11/02/2017	Solid	I6239-02	Corrosivity	Cool 4 deg C	PSEG01	D52	DRUM-5	10/30/2017	9045C

Date/Time 10/30/17 4:50 PM
 Received by: JL
 Relinquished by: CP

Date/Time 10/30/17 6:55 PM
 Received by: CS
 Relinquished by: JP