



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : _____

Run Number: LB91227

Temp.: 20.2 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/31/2017	09:30
2	CAL2	1	Water			6.99	10/31/2017	09:31
3	CAL3	1	Water			10.03	10/31/2017	09:32
4	ICV	1	Water			7.00	10/31/2017	09:34
5	CCV1	1	Water			2.01	10/31/2017	09:35
6	I6234-04	1	Water			6.89	10/31/2017	09:36
7	I6234-04DUP	1	Water			6.90	10/31/2017	09:37
8	I6236-04	1	Water			6.90	10/31/2017	09:39
9	CCV2	1	Water			12.01	10/31/2017	09:42

WORKLIST(Hardcopy Internal Chain)

2391227

WorkList Name : PH I6234

WorkList ID : 104935

Date : 10/31/2017 9:13:29 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/28/2017	Water	I6234-04-I	pH	Cool 4 deg C	AKRF02	D52	TW-5-DEP	10/28/2017	SM 4500-PH B
10/30/2017	Water	I6236-04-D	pH	Cool 4 deg C	PSEG01	D52	FM782	10/30/2017	SM 4500-PH B

Date/Time 10/31/17 10:00 AM
 Received by: JP
 Relinquished by: JP

Date/Time 10-31-17 10:00 AM
 Received by: JP
 Relinquished by: JP