

245.1


LB 91229

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31920



A= 0.0000e+000

B= 3.1658e-004

C= 1.0434e-002

Rho= 0.9996813

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	9	0.000	9				
0.2	0.200	0.214	0.014	644	0.0 %	644				
2.5	2.500	2.507	0.007	7885	0.0 %	7885				
5.0	5.000	5.041	0.041	15892	0.0 %	15892				
7.5	7.500	7.309	-0.191	23054	0.0 %	23054				
10.0	10.000	10.116	0.116	31920	0.0 %	31920				

LB91229 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 31 Oct 2017 10:19:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	245.1	PPB	31 Oct 2017 10:40:21	S	Std
0.2 - 1	50.2	644	-	0.2000	245.1	PPB	31 Oct 2017 10:42:29	S	Std
2.5 - 1	52.5	7885	-	2.5000	245.1	PPB	31 Oct 2017 10:44:37	S	Std
5.0 - 1	55	15892	-	5.0000	245.1	PPB	31 Oct 2017 10:46:46	S	Std
7.5 - 1	57.5	23054	-	7.5000	245.1	PPB	31 Oct 2017 10:48:53	S	Std
10.0 - 1	510	31920	-	10.0000	245.1	PPB	31 Oct 2017 10:51:02	S	Std
ICV37 - 1	ICV37	12459	3.9547	-245.1	PPB	PPB	31 Oct 2017 10:53:48	U	SMPL
ICB37 - 1	ICB37	-137	-0.0329	-245.1	PPB	PPB	31 Oct 2017 10:55:54	U	SMPL
CCV55 - 1	CCV55	15883	5.0386	-245.1	PPB	PPB	31 Oct 2017 10:58:01	U	SMPL
CCB55 - 1	CCB55	-136	-0.0326	-245.1	PPB	PPB	31 Oct 2017 11:00:08	U	SMPL
CRA - 1	CRA	639	0.2127	-245.1	PPB	PPB	31 Oct 2017 11:02:15	U	SMPL
HighStd - 1	HighStd	31603	10.0152	-245.1	PPB	PPB	31 Oct 2017 11:04:22	U	SMPL
ChkStd - 1	ChkStd	22680	7.1904	-245.1	PPB	PPB	31 Oct 2017 11:06:29	U	SMPL
PB103791BL - 1	PBW	-179	-0.0462	-245.1	PPB	PPB	31 Oct 2017 11:08:38	U	SMPL
PB103791BS - 1	LCSW	14154	4.4912	-245.1	PPB	PPB	31 Oct 2017 11:10:47	U	SMPL
I6199-01 - 1	MR-4	-153	-0.0380	-245.1	PPB	PPB	31 Oct 2017 11:12:55	U	SMPL
I6199-01DUP - 1	MR-4DUP	-43	-0.0032	-245.1	PPB	PPB	31 Oct 2017 11:15:03	U	SMPL
I6199-01MS - 1	MR-4MS	13055	4.1433	-245.1	PPB	PPB	31 Oct 2017 11:17:11	U	SMPL
I6199-01MSD - 1	MR-4MSD	12698	4.0303	-245.1	PPB	PPB	31 Oct 2017 11:19:20	U	SMPL
I6199-02 - 1	B-5W	-11	0.0070	-245.1	PPB	PPB	31 Oct 2017 11:21:26	U	SMPL
CCV56 - 1	CCV56	16101	5.1076	-245.1	PPB	PPB	31 Oct 2017 11:23:33	U	SMPL
CCB56 - 1	CCB56	-138	-0.0333	-245.1	PPB	PPB	31 Oct 2017 11:25:40	U	SMPL
I6199-03 - 1	B-6W	-5	0.0089	-245.1	PPB	PPB	31 Oct 2017 11:27:48	U	SMPL
I6203-01 - 1	102717-TIDO-E-D-B	4	0.0117	-245.1	PPB	PPB	31 Oct 2017 11:29:55	U	SMPL
I6203-02 - 1	102717-TIDO-E-D-S	-11	0.0070	-245.1	PPB	PPB	31 Oct 2017 11:32:01	U	SMPL
I6203-03 - 1	102717-TIDO-E-U-B	7	0.0127	-245.1	PPB	PPB	31 Oct 2017 11:34:08	U	SMPL
I6203-04 - 1	102717-TIDO-E-U-S	0	0.0104	-245.1	PPB	PPB	31 Oct 2017 11:36:16	U	SMPL
I6204-01 - 1	102617-TIDO-E-D-B	-1	0.0101	-245.1	PPB	PPB	31 Oct 2017 11:38:23	U	SMPL
I6204-02 - 1	102617-TIDO-E-D-S	-8	0.0079	-245.1	PPB	PPB	31 Oct 2017 11:40:31	U	SMPL
I6204-03 - 1	102617-TIDO-E-U-B	12	0.0142	-245.1	PPB	PPB	31 Oct 2017 11:42:40	U	SMPL
I6204-04 - 1	102617-TIDO-E-U-S	-3	0.0095	-245.1	PPB	PPB	31 Oct 2017 11:44:48	U	SMPL
I6205-01 - 1	102517-TIDO-F-D-B	-1	0.0101	-245.1	PPB	PPB	31 Oct 2017 11:46:55	U	SMPL
CCV57 - 1	CCV57	15850	5.0282	-245.1	PPB	PPB	31 Oct 2017 11:49:02	U	SMPL
CCB57 - 1	CCB57	-123	-0.0285	-245.1	PPB	PPB	31 Oct 2017 11:51:09	U	SMPL
I6205-02 - 1	102517-TIDO-F-D-S	-17	0.0051	-245.1	PPB	PPB	31 Oct 2017 11:53:17	U	SMPL
I6205-03 - 1	102517-TIDO-F-U-B	-15	0.0057	-245.1	PPB	PPB	31 Oct 2017 11:55:24	U	SMPL
I6205-04 - 1	102517-TIDO-F-U-S	-7	0.0082	-245.1	PPB	PPB	31 Oct 2017 11:57:30	U	SMPL
I6234-04 - 1	TW-5-DEP	153	0.0589	-245.1	PPB	PPB	31 Oct 2017 11:59:37	U	SMPL
I6199-01LX5 - 1		-12	0.0066	-245.1	PPB	PPB	31 Oct 2017 12:01:44	U	SMPL
I6199-01A - 1		13529	4.2934	-245.1	PPB	PPB	31 Oct 2017 12:03:52	U	SMPL
CCV58 - 1	CCV58	15903	5.0449	-245.1	PPB	PPB	31 Oct 2017 12:06:00	U	SMPL
CCB58 - 1	CCB58	-130	-0.0307	-245.1	PPB	PPB	31 Oct 2017 12:08:08	U	SMPL