



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB91232
Analysis Date: 10/31/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 10/31/2017
Extraction IN Time: 11:30
Extraction OUT Time: 12:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB91232BL	LB91232BL	WATER	2	1000	100	3.0163	3.0163	3.02	3.0163	3.0163	0	0
2	LB91232BS	LB91232BS	WATER	2	1000	100	3.1479	3.1479	3.01	3.1664	3.1664	0.0185	18.5
3	LB91232BSD	LB91232BSD	WATER	2	1000	100	3.0624	3.0624	3.04	3.081	3.081	0.0186	18.6
4	I6234-04	TW-5-DEP	WATER	2	970	100	3.0906	3.0906	3.05	3.0908	3.0908	0.0002	0.21



QC Batch# LB91232 Test: Non-Polar Ma Analysis Date: 10/31/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2288
pH Paper 0-14	W2248
Sodium Sulfate	EP1786
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 71 Dessicator Time In : 13:20
1.0000 gram Balance: 1.0004 In Time: 12:40
Bal Check Time: 11:45

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 70 Dessicator Time Out: 14:50
1.0000 gram Balance: 1.0005 Out Time: 14:20
Bal Check Time: 15:30

LB 9/232

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR I6234 WorkList ID : 104945 Date : 10/31/2017 11:26:41 AM

Due Date	Matrix	Sample	Test	Non-Polar Material	Conc H2SO4 to	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2017	Water	I6234-04 -D				AKRF02	D51	TW-5-DEP	10/28/2017	1664A

Date/Time 10-31-17 11:26 AM
Received by: JP
Relinquished by: CS

Date/Time 10-31-17 11:25 AM
Received by: CS
Relinquished by: JP