

7471A


LB 91233

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40740



A= 0.0000e+000

B= 2.4381e-004

C= -1.2693e-002

Rho= 0.9997837

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-11	0.000	-11				
0.2	0.200	0.205	0.005	892	0.0 %	892				
2.5	2.500	2.413	-0.087	9949	0.0 %	9949				
5.0	5.000	5.140	0.140	21136	0.0 %	21136				
7.5	7.500	7.537	0.037	30967	0.0 %	30967	0			
10.0	10.000	9.920	-0.080	40740	0.0 %	40740				

LB91233 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 31 Oct 2017 12:31:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.0000	7471A	PPB	31 Oct 2017 12:33:21	S	Std
0.2 - 1	50.2	892	-	0.2000	7471A	PPB	31 Oct 2017 12:35:38	S	Std
2.5 - 1	52.5	9949	-	2.5000	7471A	PPB	31 Oct 2017 12:37:56	S	Std
5.0 - 1	55	21136	-	5.0000	7471A	PPB	31 Oct 2017 12:40:13	S	Std
7.5 - 1	57.5	30967	-	7.5000	7471A	PPB	31 Oct 2017 12:42:31	S	Std
10.0 - 1	510	40740	-	10.0000	7471A	PPB	31 Oct 2017 12:47:29	S	Std
ICV38 - 1	ICV38	17004	4.1330	-	7471A	PPB	31 Oct 2017 12:50:25	U	SMPL
ICB38 - 1	ICB38	-152	-0.0498	-	7471A	PPB	31 Oct 2017 12:52:40	U	SMPL
CCV59 - 1	CCV59	20086	4.8844	-	7471A	PPB	31 Oct 2017 12:54:56	U	SMPL
CCB59 - 1	CCB59	-151	-0.0495	-	7471A	PPB	31 Oct 2017 12:57:12	U	SMPL
CRA - 1	CRA	871	0.1997	-	7471A	PPB	31 Oct 2017 12:59:29	U	SMPL
HighStd - 1	HighStd	41054	9.9965	-	7471A	PPB	31 Oct 2017 13:01:44	U	SMPL
ChkStd - 1	ChkStd	29540	7.1894	-	7471A	PPB	31 Oct 2017 13:04:00	U	SMPL
PB103821BL - 1	PBS	-234	-0.0697	-	7471A	PPB	31 Oct 2017 13:06:19	U	SMPL
PB103821BS - 1	LCSS	15826	3.8458	-	7471A	PPB	31 Oct 2017 13:08:37	U	SMPL
I6167-01 - 1	S13-SP2C	103614	25.2491	-	7471A	PPB	31 Oct 2017 13:10:55	U	SMPL
I6174-02 - 1	B7(0-10)COMP	19316	4.6967	-	7471A	PPB	31 Oct 2017 13:13:12	U	SMPL
I6174-02DUP - 1	B7(0-10)COMP	20374	4.9546	-	7471A	PPB	31 Oct 2017 13:15:32	U	SMPL
I6174-02MS - 1	B7(0-10)COMPMS	36257	8.8270	-	7471A	PPB	31 Oct 2017 13:17:51	U	SMPL
I6174-02MSD - 1	B7(0-10)COMPMSD	38256	9.3144	-	7471A	PPB	31 Oct 2017 13:20:08	U	SMPL
CCV60 - 1	CCV60	19787	4.8115	-	7471A	PPB	31 Oct 2017 13:22:26	U	SMPL
CCB60 - 1	CCB60	-182	-0.0571	-	7471A	PPB	31 Oct 2017 13:24:44	U	SMPL
I6181-01 - 1	WASTE	620	0.1385	-	7471A	PPB	31 Oct 2017 13:27:00	U	SMPL
I6182-01 - 1	RT-3026	2643	0.6317	-	7471A	PPB	31 Oct 2017 13:29:16	U	SMPL
I6194-01 - 1	BU-01-102717	468	0.1014	-	7471A	PPB	31 Oct 2017 13:31:32	U	SMPL
I6196-01 - 1	OK-03-102717	526	0.1155	-	7471A	PPB	31 Oct 2017 13:33:49	U	SMPL
I6200-01 - 1	OR-01-102717	3973	0.9560	-	7471A	PPB	31 Oct 2017 13:36:05	U	SMPL
I6206-01 - 1	EO-01-102717-A	2419	0.5771	-	7471A	PPB	31 Oct 2017 13:38:22	U	SMPL
I6206-03 - 1	EO-01-102717-B	2227	0.5303	-	7471A	PPB	31 Oct 2017 13:40:40	U	SMPL
I6206-05 - 1	EO-01-102717-C	7212	1.7456	-	7471A	PPB	31 Oct 2017 13:42:57	U	SMPL
I6209-01 - 1	FK-SF-02-002-A	1472	0.3462	-	7471A	PPB	31 Oct 2017 13:45:15	U	SMPL
I6209-03 - 1	FK-SF-02-002-B	1405	0.3299	-	7471A	PPB	31 Oct 2017 13:47:31	U	SMPL
CCV61 - 1	CCV61	20083	4.8837	-	7471A	PPB	31 Oct 2017 13:49:48	U	SMPL
CCB61 - 1	CCB61	-171	-0.0544	-	7471A	PPB	31 Oct 2017 13:52:04	U	SMPL
I6210-01 - 1	FK-PS-01-020	500	0.1092	-	7471A	PPB	31 Oct 2017 13:54:21	U	SMPL
I6213-01 - 1	FK-PS-01-019	497	0.1085	-	7471A	PPB	31 Oct 2017 13:56:38	U	SMPL
I6216-01 - 1	EPE-109-3(0-5)	8802	2.1333	-	7471A	PPB	31 Oct 2017 13:58:54	U	SMPL
I6216-02 - 1	EPE-109-3(5-8)	6540	1.5818	-	7471A	PPB	31 Oct 2017 14:01:11	U	SMPL
PB103828BL - 1	PBS	-54	-0.0259	-	7471A	PPB	31 Oct 2017 14:03:27	U	SMPL
PB103828BS - 1	LCSS	15511	3.7690	-	7471A	PPB	31 Oct 2017 14:05:44	U	SMPL
I6202-01 - 1	ROLL-OFF-2	87669	21.3616	-	7471A	PPB	31 Oct 2017 14:08:01	U	SMPL
I6226-01 - 1	TP-1	853	0.1953	-	7471A	PPB	31 Oct 2017 14:17:31	U	SMPL
I6226-01DUP - 1	TP-1DUP	893	0.2050	-	7471A	PPB	31 Oct 2017 14:19:49	U	SMPL
I6226-01MS - 1	TP-1MS	17209	4.1830	-	7471A	PPB	31 Oct 2017 14:22:05	U	SMPL
CCV62 - 1	CCV62	19997	4.8627	-	7471A	PPB	31 Oct 2017 14:24:21	U	SMPL
CCB62 - 1	CCB62	-174	-0.0551	-	7471A	PPB	31 Oct 2017 14:26:38	U	SMPL
I6226-01MSD - 1	TP-1MSD	17055	4.1454	-	7471A	PPB	31 Oct 2017 14:28:55	U	SMPL
I6226-09 - 1	TP-2	1315	0.3079	-	7471A	PPB	31 Oct 2017 14:31:11	U	SMPL
I6226-17 - 1	T-1-CONCRETE	272	0.0536	-	7471A	PPB	31 Oct 2017 14:33:28	U	SMPL
I6231-01 - 1	TR-04-103017	3092	0.7412	-	7471A	PPB	31 Oct 2017 14:35:44	U	SMPL
I6233-01 - 1	SB-1(3-5)	602	0.1341	-	7471A	PPB	31 Oct 2017 14:38:01	U	SMPL
I6233-02 - 1	SB-2(6-8)	916	0.2106	-	7471A	PPB	31 Oct 2017 14:40:18	U	SMPL
I6233-03 - 1	SB-3(6-8)	6030	1.4575	-	7471A	PPB	31 Oct 2017 14:42:35	U	SMPL
I6233-04 - 1	SB-4(7-9)	24729	6.0164	-	7471A	PPB	31 Oct 2017 14:44:53	U	SMPL
I6233-05 - 1	SB-5(2-4)	4318	1.0401	-	7471A	PPB	31 Oct 2017 14:47:11	U	SMPL
I6233-06 - 1	SB-6(8-10)	3339	0.8014	-	7471A	PPB	31 Oct 2017 14:49:29	U	SMPL
CCV63 - 1	CCV63	20409	4.9632	-	7471A	PPB	31 Oct 2017 14:51:45	U	SMPL
CCB63 - 1	CCB63	-152	-0.0498	-	7471A	PPB	31 Oct 2017 14:54:01	U	SMPL
I6236-01 - 1	133	28461	6.9263	-	7471A	PPB	31 Oct 2017 14:56:19	U	SMPL
I6174-02LX5 - 1		3451	0.8287	-	7471A	PPB	31 Oct 2017 14:58:36	U	SMPL
I6174-02A - 1		36935	8.9923	-	7471A	PPB	31 Oct 2017 15:00:54	U	SMPL
I6226-01LX5 - 1		-128	-0.0439	-	7471A	PPB	31 Oct 2017 15:03:10	U	SMPL
I6226-01A - 1		17617	4.2824	-	7471A	PPB	31 Oct 2017 15:05:29	U	SMPL
I6167-01DLX5 - 1		19927	4.8456	-	7471A	PPB	31 Oct 2017 15:07:46	U	SMPL
I620201DLX5 - 1		16997	4.1313	-	7471A	PPB	31 Oct 2017 15:10:04	U	SMPL
CCV64 - 1	CCV64	19981	4.8588	-	7471A	PPB	31 Oct 2017 15:12:22	U	SMPL
CCB64 - 1	CCB64	-153	-0.0500	-	7471A	PPB	31 Oct 2017 15:14:40	U	SMPL