

LB 91253


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33441



A= 0.0000e+000

B= 2.9816e-004

C= -1.5233e-002

Rho= 0.9999705

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	3	0.000	3				
0.2	0.200	0.190	-0.010	690	0.0 %	690				
2.5	2.500	2.508	0.008	8464	0.0 %	8464				
5.0	5.000	5.011	0.011	16859	0.0 %	16859				
7.5	7.500	7.548	0.048	25368	0.0 %	25368				
10.0	10.000	9.956	-0.044	33441	0.0 %	33441				

LB91253 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 01 Nov 2017 10:00:02					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7470A	PPB	01 Nov 2017 10:10:29	S	Std
0.2 - 1	50.2	690	-	0.0000	7470A	PPB	01 Nov 2017 10:12:37	S	Std
2.5 - 1	52.5	8464	-	2.5000	7470A	PPB	01 Nov 2017 10:14:46	S	Std
5.0 - 1	55	16859	-	5.0000	7470A	PPB	01 Nov 2017 10:16:54	S	Std
7.5 - 1	57.5	25368	-	7.5000	7470A	PPB	01 Nov 2017 10:19:02	S	Std
10.0 - 1	510	33441	-	10.0000	7470A	PPB	01 Nov 2017 10:21:10	S	Std
ICV39 - 1	ICV39	12886	3.8269	-	7470A	PPB	01 Nov 2017 10:23:49	U	SMPL
ICB39 - 1	ICB39	-125	-0.0525	-	7470A	PPB	01 Nov 2017 10:25:55	U	SMPL
CCV65 - 1	CCV65	17838	5.3033	-	7470A	PPB	01 Nov 2017 10:28:03	U	SMPL
CCB65 - 1	CCB65	-143	-0.0579	-	7470A	PPB	01 Nov 2017 10:30:09	U	SMPL
CRA - 1	CRA	683	0.1884	-	7470A	PPB	01 Nov 2017 10:32:16	U	SMPL
HighStd - 1	HighStd	34353	10.2274	-	7470A	PPB	01 Nov 2017 10:34:23	U	SMPL
ChkStd - 1	ChkStd	23854	7.0971	-	7470A	PPB	01 Nov 2017 10:36:30	U	SMPL
PB103844BL - 1	PBW	-214	-0.0790	-	7470A	PPB	01 Nov 2017 10:38:39	U	SMPL
PB103844BS - 1	LCSW	12661	3.7598	-	7470A	PPB	01 Nov 2017 10:40:48	U	SMPL
I6234-01 - 1	TW-1	251	0.0596	-	7470A	PPB	01 Nov 2017 10:42:56	U	SMPL
I6234-01DUP - 1	TW-1DUP	249	0.0590	-	7470A	PPB	01 Nov 2017 10:45:04	U	SMPL
I6234-01MS - 1	TW-1MS	14202	4.2192	-	7470A	PPB	01 Nov 2017 10:47:12	U	SMPL
I6234-01MSD - 1	TW-1MSD	13860	4.1173	-	7470A	PPB	01 Nov 2017 10:49:21	U	SMPL
I6234-02 - 1	TW-5	-109	-0.0477	-	7470A	PPB	01 Nov 2017 10:51:27	U	SMPL
CCV66 - 1	CCV66	17138	5.0946	-	7470A	PPB	01 Nov 2017 10:53:34	U	SMPL
CCB66 - 1	CCB66	-130	-0.0540	-	7470A	PPB	01 Nov 2017 10:55:41	U	SMPL
I6234-05 - 1	TW-1	265	0.0638	-	7470A	PPB	01 Nov 2017 10:57:48	U	SMPL
I6234-06 - 1	TW-5	326	0.0820	-	7470A	PPB	01 Nov 2017 10:59:55	U	SMPL
I6236-04 - 1	FM782	37	-0.0042	-	7470A	PPB	01 Nov 2017 11:02:02	U	SMPL
PB103866BL - 1	PBW	32	-0.0057	-	7470A	PPB	01 Nov 2017 11:04:09	U	SMPL
PB103866BS - 1	LCSW	14609	4.3406	-	7470A	PPB	01 Nov 2017 11:06:16	U	SMPL
PB103866TB - 1	PB103825TB	-103	-0.0459	-	7470A	PPB	01 Nov 2017 11:08:23	U	SMPL
I6225-01 - 1	CB-1017-S-TCLP-47S2	1	-0.0149	-	7470A	PPB	01 Nov 2017 11:10:32	U	SMPL
I6225-02 - 1	CB-1017-S-TCLP-47S1	18	-0.0099	-	7470A	PPB	01 Nov 2017 11:12:40	U	SMPL
I6226-08 - 1	TP-1	17	-0.0102	-	7470A	PPB	01 Nov 2017 11:14:49	U	SMPL
I6226-16 - 1	TP-2	18	-0.0099	-	7470A	PPB	01 Nov 2017 11:16:56	U	SMPL
CCV67 - 1	CCV67	17182	5.1077	-	7470A	PPB	01 Nov 2017 11:19:03	U	SMPL
CCB67 - 1	CCB67	-130	-0.0540	-	7470A	PPB	01 Nov 2017 11:21:10	U	SMPL
I6226-18 - 1	T-1-CONCRETE	1	-0.0149	-	7470A	PPB	01 Nov 2017 11:23:17	U	SMPL
I6231-02 - 1	TR-04-103017	1	-0.0149	-	7470A	PPB	01 Nov 2017 11:25:24	U	SMPL
I6236-02 - 1	133	15	-0.0108	-	7470A	PPB	01 Nov 2017 11:27:31	U	SMPL
I6236-05 - 1	FM782	16	-0.0105	-	7470A	PPB	01 Nov 2017 11:29:38	U	SMPL
I6236-05DUP - 1	FM782DUP	26	-0.0075	-	7470A	PPB	01 Nov 2017 11:31:46	U	SMPL
I6236-05MS - 1	FM782MS	12967	3.8510	-	7470A	PPB	01 Nov 2017 11:33:54	U	SMPL
I6236-05MSD - 1	FM782MSD	12740	3.7833	-	7470A	PPB	01 Nov 2017 11:36:02	U	SMPL
CCV68 - 1	CCV68	17860	5.3099	-	7470A	PPB	01 Nov 2017 11:38:10	U	SMPL
CCB68 - 1	CCB68	-123	-0.0519	-	7470A	PPB	01 Nov 2017 11:40:19	U	SMPL
I6234-01LX5 - 1		58	0.0021	-	7470A	PPB	01 Nov 2017 11:42:26	U	SMPL
I6234-01A - 1		14321	4.2547	-	7470A	PPB	01 Nov 2017 11:44:33	U	SMPL
I6236-05LX5 - 1		-80	-0.0391	-	7470A	PPB	01 Nov 2017 11:46:40	U	SMPL
I6236-05A - 1		13629	4.0484	-	7470A	PPB	01 Nov 2017 11:48:48	U	SMPL
CCV69 - 1	CCV69	17184	5.1083	-	7470A	PPB	01 Nov 2017 11:50:55	U	SMPL
CCB69 - 1	CCB69	-130	-0.0540	-	7470A	PPB	01 Nov 2017 11:53:03	U	SMPL