

LB91256

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/31/2017 19:30

Test: Phenolics-mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.013	0.0	0.115	
ICB1	-0.001	0.0	0.004	
CCV1	1.037	0.0	0.117	
CCB1	-0.003	0.0	0.004	
PB103862BL	-0.002	0.0	0.004	
PB103862BS	1.010	0.0	0.114	
I6234-04	-0.002	0.0	0.004	
I6234-04DUP	0.000	0.0	0.004	
I6234-04MS	0.926	0.0	0.105	
I6234-04MSD	0.927	0.0	0.105	
CCV2	1.005	0.0	0.114	
CCB2	-0.004	0.0	0.004	

N 12
Mean 0.492
SD 0.5171
CV% 105.03

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Oct 31 19:06:47 2017

Tue Oct 31 19:30:26 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0473	mg/l	10/31/2017 19:06:47	
0.1PPM	A	Phenolics-	P	0.0981	mg/l	10/31/2017 19:06:48	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	10/31/2017 19:06:49	
0.50PPM	A	Phenolics-	P	0.5	mg/l	10/31/2017 19:06:50	
1.0PPM	A	Phenolics-	P	1.0074	mg/l	10/31/2017 19:06:51	
2.0PPM	A	Phenolics-	P	1.9964	mg/l	10/31/2017 19:06:52	
ICV1	S	Phenolics-	P	1.0133	mg/l	10/31/2017 19:30:15	
ICB1	S	Phenolics-	P	-0.0006	mg/l	10/31/2017 19:30:16	
CCV1	S	Phenolics-	P	1.0372	mg/l	10/31/2017 19:30:17	
CCB1	S	Phenolics-	P	-0.0026	mg/l	10/31/2017 19:30:18	
PB103862BL	S	Phenolics-	P	-0.0021	mg/l	10/31/2017 19:30:19	
PB103862BS	S	Phenolics-	P	1.0098	mg/l	10/31/2017 19:30:20	
I6234-04	S	Phenolics-	P	-0.0018	mg/l	10/31/2017 19:30:21	
I6234-04DUP	S	Phenolics-	P	0.0003	mg/l	10/31/2017 19:30:22	
I6234-04MS	S	Phenolics-	P	0.9261	mg/l	10/31/2017 19:30:23	
I6234-04MSD	S	Phenolics-	P	0.9274	mg/l	10/31/2017 19:30:24	
CCV2	S	Phenolics-	P	1.0047	mg/l	10/31/2017 19:30:25	
CCB2	S	Phenolics-	P	-0.0039	mg/l	10/31/2017 19:30:26	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/31/2017 19:07


Test Phenolics-

Accepted 10/31/2017 19:07

Factor 9.175
Bias 0.004

Coeff. of det. 0.999971

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0473	0.0500	
2	Phenol-2	0.015	0.0981	0.1000	
3	Phenol-2	0.032	0.2508	0.2500	
4	Phenol-2	0.059	0.5000	0.5000	
5	Phenol-2	0.114	1.0074	1.0000	
6	Phenol-2	0.222	1.9964	2.0000	