


7471A

LB 91265
INSTRUMENT ID: LV2

Linear

 μ Abs.:

41983



A= 0.0000e+000

B= 2.4241e-004

C= 6.1554e-004

Rho= 0.9995026

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.068	0.068	279	0.000	279				
0.2	0.200	0.240	0.040	986	0.0 %	986				
2.5	2.500	2.480	-0.020	10229	0.0 %	10229				
5.0	5.000	4.929	-0.071	20332	0.0 %	20332				
7.5	7.500	7.305	-0.195	30132	0.0 %	30132	0			
10.0	10.000	10.178	0.178	41983	0.0 %	41983				

LB91265 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 01 Nov 2017 15:06:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	279	-	0.0000	7471A	PPB	01 Nov 2017 15:08:47	S	Std
0.2 - 1	52.5	986	-	0.2000	7471A	PPB	01 Nov 2017 15:11:05	S	Std
2.5 - 1	55	10229	-	2.5000	7471A	PPB	01 Nov 2017 15:13:22	S	Std
5.0 - 1	57.5	20332	-	5.0000	7471A	PPB	01 Nov 2017 15:15:40	S	Std
7.5 - 1	310	30132	-	7.5000	7471A	PPB	01 Nov 2017 15:17:57	S	Std
10.0 - 1		41983	-	10.0000	7471A	PPB	01 Nov 2017 15:20:14	S	Std
ICV40 - 1	ICV40	16239	3.9371	-	7471A	PPB	01 Nov 2017 15:23:17	U	SMPL
ICB40 - 1	ICB40	63	0.0159	-	7471A	PPB	01 Nov 2017 15:25:32	U	SMPL
CCV70 - 1	CCV70	20540	4.9797	-	7471A	PPB	01 Nov 2017 15:27:48	U	SMPL
CCB70 - 1	CCB70	16	0.0045	-	7471A	PPB	01 Nov 2017 15:30:04	U	SMPL
CRA - 1	CRA	966	0.2348	-	7471A	PPB	01 Nov 2017 15:32:20	U	SMPL
HighStd - 1	HighStd	41744	10.1198	-	7471A	PPB	01 Nov 2017 15:34:36	U	SMPL
ChkStd - 1	ChkStd	27879	6.7588	-	7471A	PPB	01 Nov 2017 15:36:52	U	SMPL
PB103869BL - 1	PBS	-95	-0.0224	-	7471A	PPB	01 Nov 2017 15:39:11	U	SMPL
PB103869BS - 1	LCSS	15218	3.6896	-	7471A	PPB	01 Nov 2017 15:41:29	U	SMPL
I6236-03 - 1	DRUM-1-2	984	0.2391	-	7471A	PPB	01 Nov 2017 15:43:46	U	SMPL
I6247-01 - 1	S13-SP3-C	143604	34.8116	-	7471A	PPB	01 Nov 2017 15:46:03	U	SMPL
I6249-01 - 1	COMP-1	6773	1.6425	-	7471A	PPB	01 Nov 2017 15:48:21	U	SMPL
I6249-03 - 1	COMP-2	1547	0.3756	-	7471A	PPB	01 Nov 2017 15:50:45	U	SMPL
I6252-01 - 1	HR-05-103117	1808	0.4389	-	7471A	PPB	01 Nov 2017 15:53:01	U	SMPL
CCV71 - 1	CCV71	20386	4.9424	-	7471A	PPB	01 Nov 2017 15:55:17	U	SMPL
CCB71 - 1	CCB71	24	0.0064	-	7471A	PPB	01 Nov 2017 15:57:33	U	SMPL
I6253-01 - 1	SA-01-103117	1404	0.3410	-	7471A	PPB	01 Nov 2017 16:04:59	U	SMPL
I6253-01DUP - 1	SA-01-103117DUP	1340	0.3254	-	7471A	PPB	01 Nov 2017 16:08:49	U	SMPL
I6253-01MS - 1	SA-01-103117MS	16996	4.1206	-	7471A	PPB	01 Nov 2017 16:11:05	U	SMPL
I6253-01MSD - 1	SA-01-103117MSD	16789	4.0704	-	7471A	PPB	01 Nov 2017 16:13:21	U	SMPL
I6254-01 - 1	DRUM-1-2	136	0.0336	-	7471A	PPB	01 Nov 2017 16:15:38	U	SMPL
I6257-01 - 1	EPE-110-1(0-5)	2085	0.5060	-	7471A	PPB	01 Nov 2017 16:17:55	U	SMPL
I6257-02 - 1	EPE-110-1(5-10)	1480	0.3594	-	7471A	PPB	01 Nov 2017 16:20:13	U	SMPL
I6253-01LX5 - 1		464	0.1131	-	7471A	PPB	01 Nov 2017 16:22:30	U	SMPL
I6253-01A - 1		20269	4.9140	-	7471A	PPB	01 Nov 2017 16:24:48	U	SMPL
I6247-01DLX5 - 1		28942	7.0164	-	7471A	PPB	01 Nov 2017 16:27:04	U	SMPL
CCV72 - 1	CCV72	20378	4.9404	-	7471A	PPB	01 Nov 2017 16:29:21	U	SMPL
CCB72 - 1	CCB72	9	0.0028	-	7471A	PPB	01 Nov 2017 16:31:38	U	SMPL