



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91273

Temp.: 21.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/01/2017	17:10
2	CAL2	1	Water			7.01	11/01/2017	17:11
3	CAL3	1	Water			10.03	11/01/2017	17:13
4	ICV	1	Water			6.99	11/01/2017	17:15
5	CCV1	1	Water			2.01	11/01/2017	17:17
6	I6243-02	1	Water			7.76	11/01/2017	17:18
7	I6243-02DUP	1	Water			7.77	11/01/2017	17:20
8	I6275-10	1	Water			9.30	11/01/2017	17:21
9	I6277-02	1	Water			7.60	11/01/2017	17:23
10	CCV2	1	Water			12.02	11/01/2017	17:25

1391273

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6275 WorkList ID : 105008 Date : 11/1/2017 4:50:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2017	Water	I6243-02-B	Corrosivity	Cool 4 deg C	PSEG01	D52	N35799	11/01/2017	9040C
11/01/2017	Water	I6275-10-17	Corrosivity	Cool 4 deg C	PSEG01	D52	2727	11/01/2017	9040C
11/01/2017	Water	I6277-02-F	Corrosivity	Cool 4 deg C	PSEG01	D52	FADW4807L	11/01/2017	9040C

Date/Time 11-01-17 5:10:00 AM
Received by: JR
Relinquished by: CR

Date/Time 11-01-17 6:30 PM
Received by: CR
Relinquished by: JP