



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91274

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/01/2017	17:30
2	CAL2	1	Water			7.00	11/01/2017	17:31
3	CAL3	1	Water			10.03	11/01/2017	17:33
4	ICV	1	Water			7.00	11/01/2017	17:35
5	CCV1	1	Water			2.01	11/01/2017	17:36
6	I6189-02	1	Solid	20.02	20	6.87	11/01/2017	17:37
7	I6189-02DUP	1	Solid	20.04	20	6.88	11/01/2017	17:39
8	I6258-03	1	Solid	20.02	20	7.15	11/01/2017	17:40
9	I6258-06	1	Solid	20.03	20	6.04	11/01/2017	17:41
10	I6265-03	1	Solid	20.05	20	8.08	11/01/2017	17:42
11	I6265-04	1	Solid	20.03	20	7.89	11/01/2017	17:44
12	I6275-02	1	Solid	20.02	20	6.14	11/01/2017	17:45
13	I6275-04	1	Solid	20.04	20	6.05	11/01/2017	17:46
14	I6275-07	1	Solid	20.05	20	10.10	11/01/2017	17:48
15	CCV2	1	Water			12.02	11/01/2017	17:50

2391274

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6277			WorkList ID : 105006		Date : 11/1/2017 4:49:08 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2017	Solid	I6189-02	Corrosivity	Cool 4 deg C	PSEG01	D52	LEONIA-GEN	11/01/2017	9045C
11/03/2017	Solid	I6258-03	Corrosivity	Cool 4 deg C	PSEG01	D52	WOODLAKE-DR-2-14-18	10/31/2017	9045C
11/03/2017	Solid	I6258-06	Corrosivity	Cool 4 deg C	PSEG01	D52	BOXWOOD-DR-2-4-12-14	10/31/2017	9045C
11/14/2017	Solid	I6265-03	Corrosivity	Cool 4 deg C	TULL02	D41	COMP-1	10/31/2017	9045C
11/14/2017	Solid	I6265-04	Corrosivity	Cool 4 deg C	TULL02	D41	COMP-2	10/31/2017	9045C
11/08/2017	Solid	I6275-02	Corrosivity	Cool 4 deg C	PSEG01	D52	DRUM(1-2)	11/01/2017	9045C
11/08/2017	Solid	I6275-04	Corrosivity	Cool 4 deg C	PSEG01	D52	DRUM-2	11/01/2017	9045C
11/08/2017	Solid	I6275-07	Corrosivity	Cool 4 deg C	PSEG01	D52	CONCRETE-PILE	11/01/2017	9045C

Date/Time 11-01-17 5:00 PM
Received by: JP
Relinquished by: CP

Date/Time 11-01-17 4:30 PM
Received by: CP
Relinquished by: JP