


7470A

LB 91287
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33515



A= 0.0000e+000

B= 2.9710e-004

C= -2.8178e-002

Rho= 0.9998618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	2	0.000	2				
0.2	0.200	0.168	-0.032	660	0.0 %	660				
2.5	2.500	2.495	-0.005	8492	0.0 %	8492				
5.0	5.000	5.125	0.125	17345	0.0 %	17345				
7.5	7.500	7.511	0.011	25374	0.0 %	25374				
10.0	10.000	9.929	-0.071	33515	0.0 %	33515				

LB91287 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Nov 2017 10:47:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.0000	7470A	PPB	02 Nov 2017 11:04:00	S	Std
0.2 - 1	50.2	660	-	0.2000	7470A	PPB	02 Nov 2017 11:06:09	S	Std
2.5 - 1	52.5	8492	-	2.5000	7470A	PPB	02 Nov 2017 11:08:17	S	Std
5.0 - 1	55	17345	-	5.0000	7470A	PPB	02 Nov 2017 11:10:26	S	Std
7.5 - 1	57.5	25374	-	7.5000	7470A	PPB	02 Nov 2017 11:12:34	S	Std
10.0 - 1	57	33515	-	10.0000	7470A	PPB	02 Nov 2017 11:14:41	S	Std
ICV41 - 1	ICV41	13389	3.9497	-7470A	PPB	02 Nov 2017 11:17:20	U	SMPL	
ICB41 - 1	ICB41	-148	-0.0721	-7470A	PPB	02 Nov 2017 11:19:26	U	SMPL	
CCV73 - 1	CCV73	17364	5.1307	-7470A	PPB	02 Nov 2017 11:21:33	U	SMPL	
CCB73 - 1	CCB73	-153	-0.0736	-7470A	PPB	02 Nov 2017 11:23:40	U	SMPL	
CRA - 1	CRA	623	0.1569	-7470A	PPB	02 Nov 2017 11:25:48	U	SMPL	
HighStd - 1	HighStd	33712	9.9878	-7470A	PPB	02 Nov 2017 11:27:54	U	SMPL	
ChkStd - 1	ChkStd	23800	7.0429	-7470A	PPB	02 Nov 2017 11:30:01	U	SMPL	
PB103894BL - 1	PBW	-183	-0.0825	-7470A	PPB	02 Nov 2017 11:32:10	U	SMPL	
PB103894BS - 1	LCSW	12723	3.7519	-7470A	PPB	02 Nov 2017 11:34:19	U	SMPL	
PB103894TB - 1	PB103867TB	-113	-0.0618	-7470A	PPB	02 Nov 2017 11:36:27	U	SMPL	
I6247-03 - 1	S13-SP3-C	-6	-0.0300	-7470A	PPB	02 Nov 2017 11:38:35	U	SMPL	
I6249-02 - 1	COMP-1	55	-0.0118	-7470A	PPB	02 Nov 2017 11:40:43	U	SMPL	
I6249-04 - 1	COMP-2	30	-0.0193	-7470A	PPB	02 Nov 2017 11:42:52	U	SMPL	
I6252-02 - 1	HR-05-103117	33	-0.0184	-7470A	PPB	02 Nov 2017 11:44:58	U	SMPL	
CCV74 - 1	CCV74	17298	5.1111	-7470A	PPB	02 Nov 2017 11:47:05	U	SMPL	
CCB74 - 1	CCB74	-121	-0.0641	-7470A	PPB	02 Nov 2017 11:49:12	U	SMPL	
I6253-02 - 1	SA-01-103117	40	-0.0163	-7470A	PPB	02 Nov 2017 11:51:19	U	SMPL	
I6254-03 - 1	DRUM-1-2	39	-0.0166	-7470A	PPB	02 Nov 2017 11:53:26	U	SMPL	
I6254-03DUP - 1	DRUM-1-2DUP	47	-0.0142	-7470A	PPB	02 Nov 2017 11:55:32	U	SMPL	
I6254-03MS - 1	DRUM-1-2MS	13062	3.8526	-7470A	PPB	02 Nov 2017 11:57:39	U	SMPL	
I6254-03MSD - 1	DRUM-1-2MSD	12668	3.7355	-7470A	PPB	02 Nov 2017 11:59:46	U	SMPL	
PB103896BL - 1	PBW	-94	-0.0561	-7470A	PPB	02 Nov 2017 12:01:54	U	SMPL	
PB103896BS - 1	LCSW	13458	3.9702	-7470A	PPB	02 Nov 2017 12:04:02	U	SMPL	
I6243-01 - 1	N35799	-125	-0.0653	-7470A	PPB	02 Nov 2017 12:06:10	U	SMPL	
I6255-01 - 1	432F	678	0.1733	-7470A	PPB	02 Nov 2017 12:13:01	U	SMPL	
I6255-01DUP - 1	432FDUP	640	0.1620	-7470A	PPB	02 Nov 2017 12:15:07	U	SMPL	
CCV75 - 1	CCV75	17357	5.1286	-7470A	PPB	02 Nov 2017 12:17:14	U	SMPL	
CCB75 - 1	CCB75	-104	-0.0591	-7470A	PPB	02 Nov 2017 12:19:20	U	SMPL	
I6255-01MS - 1	432FMS	13199	3.8933	-7470A	PPB	02 Nov 2017 12:24:04	U	SMPL	
I6255-01MSD - 1	432FMSD	13140	3.8758	-7470A	PPB	02 Nov 2017 12:26:11	U	SMPL	
I6275-09 - 1	2727	2584	0.7395	-7470A	PPB	02 Nov 2017 12:28:18	U	SMPL	
I6277-01 - 1	FADW4807L	294	0.0592	-7470A	PPB	02 Nov 2017 12:30:25	U	SMPL	
I6254-03LX5 - 1		26	-0.0205	-7470A	PPB	02 Nov 2017 12:32:33	U	SMPL	
I6254-03A - 1		14138	4.1723	-7470A	PPB	02 Nov 2017 12:34:40	U	SMPL	
I6255-01LX5 - 1		66	-0.0086	-7470A	PPB	02 Nov 2017 12:36:48	U	SMPL	
I6255-01A - 1		16304	4.8158	-7470A	PPB	02 Nov 2017 12:38:58	U	SMPL	
CCV76 - 1	CCV76	17217	5.0871	-7470A	PPB	02 Nov 2017 12:41:06	U	SMPL	
CCB76 - 1	CCB76	-108	-0.0603	-7470A	PPB	02 Nov 2017 12:43:14	U	SMPL	