

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91288

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59329
calibration std. phosphate 0.5 ppm	WP59330
calibration std. phosphate 0.3 ppm	WP59331
calibration std. 0 ppm	WP59334
phosphate ICV std.	WP59336
phosphate CCV std.	WP59335
calibration std. phosphate 0.1 ppm	WP59332
calibration std. phosphate 0.05 ppm	WP59333
Combined reagent	WP59338
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0009Slope: 0.6516Regression: 0.999813

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/02/2017	11:01
2	CAL2	0.05	1	50	50	0.033	11/02/2017	11:01
3	CAL3	0.10	1	50	50	0.065	11/02/2017	11:02
4	CAL4	0.30	1	50	50	0.186	11/02/2017	11:02
5	CAL5	0.50	1	50	50	0.331	11/02/2017	11:03
6	CAL6	1.00	1	50	50	0.650	11/02/2017	11:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91288

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.492	11/02/2017	11:04
2	ICB		1	50	50	0.000	0.001	11/02/2017	11:04
3	CCV1	0.50	1	50	50	0.328	0.505	11/02/2017	11:05
4	CCB1		1	50	50	0.000	0.001	11/02/2017	11:05
5	PB103926BL		1	50	50	0.000	0.001	11/02/2017	11:06
6	PB103926BS	0.50	1	50	50	0.319	0.491	11/02/2017	11:06
7	I6179-01		1	50	50	0.117	0.181	11/02/2017	11:07
8	I6179-01DUP		1	50	50	0.097	0.150	11/02/2017	11:07
9	I6179-01MS	0.50	1	50	50	0.380	0.585	11/02/2017	11:08
10	I6179-01MSD	0.50	1	50	50	0.385	0.592	11/02/2017	11:08
11	I6179-02		1	50	50	0.077	0.120	11/02/2017	11:09
12	I6179-03		1	50	50	0.035	0.055	11/02/2017	11:09
13	I6179-04		1	50	50	0.042	0.066	11/02/2017	11:10
14	I6179-05		1	50	50	0.252	0.388	11/02/2017	11:10
15	CCV2	0.50	1	50	50	0.329	0.506	11/02/2017	11:11
16	CCB2		1	50	50	0.000	0.001	11/02/2017	11:11
17	I6179-06		1	50	50	0.008	0.014	11/02/2017	11:12
18	I6179-07		1	50	50	0.065	0.101	11/02/2017	11:12
19	I6179-08		1	50	50	0.040	0.063	11/02/2017	11:13
20	I6179-09		1	50	50	0.018	0.029	11/02/2017	11:13
21	I6179-10		1	50	50	0.021	0.034	11/02/2017	11:14
22	I6179-11		1	50	50	0.080	0.124	11/02/2017	11:14
23	CCV3	0.50	1	50	50	0.327	0.503	11/02/2017	11:15
24	CCB3		1	50	50	0.000	0.001	11/02/2017	11:15