


ISM02.4

LB 91301
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45102



A= 0.0000e+000

B= 2.2086e-004

C= 1.8873e-003

Rho= 0.9998607

Accept=Accepted

slope
y=mx+c

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	52	0.000	52					-
0.2	0.200	0.196	-0.004	878	0.0 %	878					-2
2.5	2.500	2.521	0.021	11408	0.0 %	11408					1
5.0	5.000	4.901	-0.099	22184	0.0 %	22184					-2
7.5	7.500	7.605	0.105	34425	0.0 %	34425					1
10.0	10.000	9.963	-0.037	45102	0.0 %	45102					0

NB
19/02/17

LB91301
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	52	-		0 ISM02.4	PPB	11/2/2017 14:52	Std
	0.2 S0.2	878	-		0.2 ISM02.4	PPB	11/2/2017 14:54	Std
	2.5 S2.5	11408	-		2.5 ISM02.4	PPB	11/2/2017 14:56	Std
	5 S5	22184	-		5 ISM02.4	PPB	11/2/2017 14:59	Std
	7.5 S7.5	34425	-		7.5 ISM02.4	PPB	11/2/2017 15:01	Std
	10 S10	45102	-		10 ISM02.4	PPB	11/2/2017 15:03	Std
ICV	ICV	17360	3.836	-	ISM02.4	PPB	11/2/2017 15:06	SMPL
ICB	ICB	-158	-0.033	-	ISM02.4	PPB	11/2/2017 15:08	SMPL
CCV079	CCV079	22341	4.9361	-	ISM02.4	PPB	11/2/2017 15:10	SMPL
CCB079	CCB079	-168	-0.0352	-	ISM02.4	PPB	11/2/2017 15:13	SMPL
PB103903BL	PBS004	7	0.0034	-	ISM02.4	PPB	11/2/2017 15:15	SMPL
I6104-01	MBDXN0	478	0.1075	-	ISM02.4	PPB	11/2/2017 15:17	SMPL
I6104-02	MBDXN0D	426	0.096	-	ISM02.4	PPB	11/2/2017 15:19	SMPL
I6104-03	MBDXN0S	12107	2.6758	-	ISM02.4	PPB	11/2/2017 15:22	SMPL
I6104-13	MBDXM4	1172	0.2607	-	ISM02.4	PPB	11/2/2017 15:24	SMPL
I6104-14	MBDXR3	2236	0.4957	-	ISM02.4	PPB	11/2/2017 15:26	SMPL
I6104-15	MBDXS1	3491	0.7729	-	ISM02.4	PPB	11/2/2017 15:29	SMPL
PB103904BL	PBS004	-8	0.0001	-	ISM02.4	PPB	11/2/2017 15:31	SMPL
I6150-05	B-3(1-2)-102517	8341	1.8441	-	ISM02.4	PPB	11/2/2017 15:33	SMPL
I6150-06	B-3(2-3)-102517	965	0.215	-	ISM02.4	PPB	11/2/2017 15:35	SMPL
I6150-07	B-78(0.5-1.5)-102517	8500	1.8792	-	ISM02.4	PPB	11/2/2017 15:38	SMPL
I6150-08	B-78(5-7)-102517	14802	3.271	-	ISM02.4	PPB	11/2/2017 15:40	SMPL
I6175-02	B-79(1-2)-102617	6206	1.3725	-	ISM02.4	PPB	11/2/2017 15:42	SMPL
I6175-03	B-79(5-6)-102617	1351	0.3003	-	ISM02.4	PPB	11/2/2017 15:44	SMPL
CCV080	CCV080	22280	4.9226	-	ISM02.4	PPB	11/2/2017 15:47	SMPL
CCB080	CCB080	-120	-0.0246	-	ISM02.4	PPB	11/2/2017 15:49	SMPL