


7471A

LB 91304
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42734



A= 0.0000e+000

B= 2.3602e-004

C= -3.1054e-003

Rho= 0.9998544

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	12	0.000	12				
0.2	0.200	0.204	0.004	876	0.0 %	876				
2.5	2.500	2.550	0.050	10817	0.0 %	10817				
5.0	5.000	4.983	-0.017	21125	0.0 %	21125				
7.5	7.500	7.381	-0.119	31287	0.0 %	31287	0			
10.0	10.000	10.083	0.083	42734	0.0 %	42734				

LB91304 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 02 Nov 2017 16:10:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7471A	PPB	02 Nov 2017 16:13:02	S	Std
0.2 - 1	50.2	876	-	0.2000	7471A	PPB	02 Nov 2017 16:15:20	S	Std
2.5 - 1	50.5	10817	-	2.5000	7471A	PPB	02 Nov 2017 16:17:37	S	Std
5.0 - 1	55	21125	-	5.0000	7471A	PPB	02 Nov 2017 16:19:55	S	Std
7.5 - 1	57.5	31287	-	7.5000	7471A	PPB	02 Nov 2017 16:22:12	S	Std
10.0 - 1	57.5	42734	-	10.0000	7471A	PPB	02 Nov 2017 16:24:29	S	Std
ICV42 - 1	ICV42	16592	3.9129	-	7471A	PPB	02 Nov 2017 16:27:08	U	SMPL
ICB42 - 1	ICB42	-183	-0.0463	-	7471A	PPB	02 Nov 2017 16:29:25	U	SMPL
CCV81 - 1	CCV81	21234	5.0085	-	7471A	PPB	02 Nov 2017 16:31:41	U	SMPL
CCB81 - 1	CCB81	-136	-0.0352	-	7471A	PPB	02 Nov 2017 16:33:57	U	SMPL
CRA - 1	CRA	830	0.1928	-	7471A	PPB	02 Nov 2017 16:36:13	U	SMPL
HighStd - 1	HighStd	43932	10.3656	-	7471A	PPB	02 Nov 2017 16:38:30	U	SMPL
ChkStd - 1	ChkStd	29613	6.9861	-	7471A	PPB	02 Nov 2017 16:40:45	U	SMPL
PB103902BL - 1	PBS	-261	-0.0647	-	7471A	PPB	02 Nov 2017 16:43:04	U	SMPL
PB103902BS - 1	LCSS	15520	3.6599	-	7471A	PPB	02 Nov 2017 16:45:23	U	SMPL
I6189-01 - 1	LEONIA-GEN	3637	0.8553	-	7471A	PPB	02 Nov 2017 16:47:40	U	SMPL
I6258-01 - 1	WOODLAKE-DR-2-14-18	1354	0.3165	-	7471A	PPB	02 Nov 2017 16:49:57	U	SMPL
I6258-04 - 1	BOXWOOD-DR-2-4-12-14	1634	0.3825	-	7471A	PPB	02 Nov 2017 16:52:15	U	SMPL
I6261-01 - 1	HOMESTEAD-AISLE-SWITCHGEAR	396	0.0904	-	7471A	PPB	02 Nov 2017 16:54:33	U	SMPL
I6265-01 - 1	COMP-1	70326	16.5950	-	7471A	PPB	02 Nov 2017 16:56:49	U	SMPL
CCV82 - 1	CCV82	20712	4.8853	-	7471A	PPB	02 Nov 2017 16:59:05	U	SMPL
CCB82 - 1	CCB82	-184	-0.0465	-	7471A	PPB	02 Nov 2017 17:01:23	U	SMPL
I6265-02 - 1	COMP-2	56155	13.2504	-	7471A	PPB	02 Nov 2017 17:03:40	U	SMPL
I6267-01 - 1	P001-S001-0001-01	1195	0.2789	-	7471A	PPB	02 Nov 2017 17:05:55	U	SMPL
I6267-02 - 1	P001-S001-0001-02	2028	0.4755	-	7471A	PPB	02 Nov 2017 17:08:14	U	SMPL
I6267-03 - 1	P001-S002-0001-01	2220	0.5209	-	7471A	PPB	02 Nov 2017 17:10:30	U	SMPL
I6267-03DUP - 1	P001-S002-0001-01DUP	2197	0.5154	-	7471A	PPB	02 Nov 2017 17:12:47	U	SMPL
I6267-04 - 1	I6267-03MS	17638	4.1598	-	7471A	PPB	02 Nov 2017 17:15:04	U	SMPL
I6267-05 - 1	I6267-03MSD	17729	4.1812	-	7471A	PPB	02 Nov 2017 17:17:21	U	SMPL
I6267-06 - 1	P001-S003-0001-01	1496	0.3500	-	7471A	PPB	02 Nov 2017 17:19:38	U	SMPL
I6267-07 - 1	P001-S004-0001-01	1993	0.4673	-	7471A	PPB	02 Nov 2017 17:21:56	U	SMPL
I6275-01 - 1	DRUM(1-2)	130	0.0276	-	7471A	PPB	02 Nov 2017 17:24:12	U	SMPL
CCV83 - 1	CCV83	21008	4.9551	-	7471A	PPB	02 Nov 2017 17:26:28	U	SMPL
CCB83 - 1	CCB83	-151	-0.0387	-	7471A	PPB	02 Nov 2017 17:28:45	U	SMPL
I6275-06 - 1	CONCRETE-PILE	1700	0.3981	-	7471A	PPB	02 Nov 2017 17:31:01	U	SMPL
I6279-01 - 1	SAMPLE-D	183	0.0401	-	7471A	PPB	02 Nov 2017 17:33:18	U	SMPL
I6279-02 - 1	SAMPLE-C	59	0.0108	-	7471A	PPB	02 Nov 2017 17:35:34	U	SMPL
I6280-01 - 1	EPE-110-1(35-37)	679	0.1572	-	7471A	PPB	02 Nov 2017 17:37:50	U	SMPL
I6280-02 - 1	EPE-110-1(40-42)	638	0.1475	-	7471A	PPB	02 Nov 2017 17:40:07	U	SMPL
I6280-03 - 1	EPE-110-1(45-47)	520	0.1196	-	7471A	PPB	02 Nov 2017 17:42:24	U	SMPL
I6280-04 - 1	EPE-110-1(50-52)	651	0.1505	-	7471A	PPB	02 Nov 2017 17:44:41	U	SMPL
I6267-03LX5 - 1		387	0.0882	-	7471A	PPB	02 Nov 2017 17:46:59	U	SMPL
I6267-03A - 1		19503	4.5999	-	7471A	PPB	02 Nov 2017 17:49:16	U	SMPL
CCV84 - 1	CCV84	20902	4.9301	-	7471A	PPB	02 Nov 2017 17:51:33	U	SMPL
CCB84 - 1	CCB84	-187	-0.0472	-	7471A	PPB	02 Nov 2017 17:53:50	U	SMPL
I6265-01DLX5 - 1		14220	3.3531	-	7471A	PPB	02 Nov 2017 17:56:07	U	SMPL
I6265-02DLX5 - 1		11160	2.6308	-	7471A	PPB	02 Nov 2017 17:58:24	U	SMPL
CCV85 - 1	CCV85	21420	5.0524	-	7471A	PPB	02 Nov 2017 18:00:40	U	SMPL
CCB85 - 1	CCB85	-121	-0.0317	-	7471A	PPB	02 Nov 2017 18:02:56	U	SMPL