

Analysis Method: 9034

ANALYST: ketankumar

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB91306

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2206
IODINE SOLUTION .025N 1L	W2261
Starch Solution, 4L	W2121

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB103933BL		1	50	50	2.00	0.00	1.56	1.56	0.44	0.00	0.00	11/02/2017	14:17
2	I6243-02		1	50	50	2.00	0.00	1.16	1.16	0.84	0.40	3.20	11/02/2017	14:20
3	I6243-02DUP		1	50	50	2.00	1.16	2.30	1.14	0.86	0.42	3.36	11/02/2017	14:22
4	I6275-10		1	50	50	2.00	2.30	3.56	1.26	0.74	0.30	2.40	11/02/2017	14:25
5	I6277-02		1	50	50	2.00	3.56	4.68	1.12	0.88	0.44	3.52	11/02/2017	14:28

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume