



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91309

Temp.: 21.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/02/2017	17:00
2	CAL2	1	Water			6.99	11/02/2017	17:01
3	CAL3	1	Water			10.04	11/02/2017	17:03
4	ICV	1	Water			7.00	11/02/2017	17:05
5	CCV1	1	Water			2.01	11/02/2017	17:06
6	I6294-02	1	Water			7.75	11/02/2017	17:09
7	I6294-02DUP	1	Water			7.76	11/02/2017	17:10
8	I6296-02	1	Water			7.71	11/02/2017	17:11
9	CCV2	1	Water			12.02	11/02/2017	17:13

1391309

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6294 WorkList ID : 105060 Date : 11/2/2017 4:47:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2017	Water	I6294-02 - 17	Corrosivity	Cool 4 deg C	PSEG01	H52	N48971	11/02/2017	9040C
11/02/2017	Water	I6296-02 - D	Corrosivity	Cool 4 deg C	PSEG01	H52	SV2-1854L	11/02/2017	9040C

Date/Time 11-02-17 4:47 PM
 Received by: JP
 Relinquished by: CB

Date/Time 11-02-17 6:10 PM
 Received by: CB
 Relinquished by: JP