



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91310

Temp.: 21.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/02/2017	17:00
2	CAL2	1	Water			6.99	11/02/2017	17:01
3	CAL3	1	Water			10.04	11/02/2017	17:03
4	ICV	1	Water			7.00	11/02/2017	17:05
5	CCV1	1	Water			2.01	11/02/2017	17:06
6	I6294-01	1	Water			7.75	11/02/2017	17:09
7	I6294-01DUP	1	Water			7.76	11/02/2017	17:10
8	I6296-01	1	Water			7.71	11/02/2017	17:11
9	CCV2	1	Water			12.02	11/02/2017	17:13

2391310

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i6294

WorkList ID : 105061

Date : 11/2/2017 4:48:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2017	Water	I6294-01-17	pH	Cool 4 deg C	PSEG01	H52	N48971	11/02/2017	SM 4500-PH B
11/02/2017	Water	I6296-01-17	pH	Cool 4 deg C	PSEG01	H52	SV2-1854L	11/02/2017	SM 4500-PH B

Date/Time 11-02-17 4:47 PM
Received by: JP CP
Relinquished by:

Date/Time 11-02-17 6:10 PM
Received by: CP
Relinquished by: JP