



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91311

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	11/02/2017	17:35
2	CAL2	1	Water			7.01	11/02/2017	17:36
3	CAL3	1	Water			10.02	11/02/2017	17:37
4	ICV	1	Water			7.00	11/02/2017	17:39
5	CCV1	1	Water			2.01	11/02/2017	17:40
6	I6289-02	1	Solid	20.02	20	7.91	11/02/2017	17:41
7	I6289-02DUP	1	Solid	20.04	20	7.92	11/02/2017	17:42
8	I6299-02	1	Solid	20.03	20	8.92	11/02/2017	17:43
9	CCV2	1	Water			12.01	11/02/2017	17:45

2B91311

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6299 WorkList ID : 105062 Date : 11/2/2017 4:49:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2017	Solid	I6289-02	Corrosivity	Cool 4 deg C	PSEG05	H52	SU-03-110217	11/02/2017	9045C
11/07/2017	Solid	I6299-02	Corrosivity	Cool 4 deg C	PSEG05	H52	OK-02-110217	11/02/2017	9045C

Date/Time 11-02-17 4:50 PM
Received by: JP
Relinquished by: CP

Date/Time 11-02-17 6:30 PM
Received by: CP
Relinquished by: JP