

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91323BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59329
calibration std. phosphate 0.5 ppm	WP59330
calibration std. phosphate 0.3 ppm	WP59331
calibration std. 0 ppm	WP59334
phosphate ICV std.	WP59336
phosphate CCV std.	WP59335
calibration std. phosphate 0.1 ppm	WP59332
calibration std. phosphate 0.05 ppm	WP59333
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0009Slope: 0.6516Regression: 0.999813

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/02/2017	11:01
2	CAL2	0.05	1	50	50	0.033	11/02/2017	11:01
3	CAL3	0.10	1	50	50	0.065	11/02/2017	11:02
4	CAL4	0.30	1	50	50	0.186	11/02/2017	11:02
5	CAL5	0.50	1	50	50	0.331	11/02/2017	11:03
6	CAL6	1.00	1	50	50	0.650	11/02/2017	11:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91323

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.492	11/02/2017	11:04
2	ICB		1	50	50	0.000	0.001	11/02/2017	11:04
3	CCV1	0.50	1	50	50	0.328	0.505	11/02/2017	12:31
4	CCB1		1	50	50	0.000	0.001	11/02/2017	12:31
5	PB103927BL		1	1.00	50	0.000	0.001	11/02/2017	12:32
6	PB103927BS	0.50	1	1.01	50	0.321	0.494	11/02/2017	12:32
7	I6157-01		1	1.02	50	2.445	3.754	11/02/2017	12:33
8	I6157-01DUP		1	1.01	50	2.439	3.744	11/02/2017	12:33
9	I6157-01MS	0.50	1	1.03	50	2.682	4.117	11/02/2017	12:34
10	I6157-01MSD	0.50	1	1.03	50	2.675	4.107	11/02/2017	12:34
11	I6157-02		1	1.06	50	2.703	4.150	11/02/2017	12:35
12	I6157-03		1	1.01	50	2.666	4.093	11/02/2017	12:35
13	I6157-04		1	1.04	50	2.594	3.982	11/02/2017	12:36
14	I6157-05		1	1.05	50	1.618	2.484	11/02/2017	12:36
15	CCV2	0.50	1	50	50	0.329	0.506	11/02/2017	12:37
16	CCB2		1	50	50	0.000	0.001	11/02/2017	12:37
17	I6157-06		1	1.05	50	1.363	2.093	11/02/2017	12:38
18	I6157-01		20	1.02	50	0.524	0.806	11/02/2017	13:08
19	I6157-01DUP		20	1.01	50	0.520	0.799	11/02/2017	13:08
20	I6157-01MS	0.50	20	1.03	50	0.647	0.994	11/02/2017	13:09
21	I6157-01MSD	0.50	20	1.03	50	0.643	0.988	11/02/2017	13:09
22	I6157-02		20	1.06	50	0.502	0.772	11/02/2017	13:10
23	I6157-03		20	1.01	50	0.450	0.692	11/02/2017	13:11
24	I6157-04		20	1.04	50	0.490	0.753	11/02/2017	13:11
25	I6157-05		10	1.05	50	0.524	0.806	11/02/2017	13:12
26	I6157-06		10	1.05	50	0.397	0.611	11/02/2017	13:12
27	CCV3	0.50	1	50	50	0.330	0.508	11/02/2017	13:13
28	CCB3		1	50	50	0.000	0.001	11/02/2017	13:13