



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB91329

Reagent/Standard	Lot/Log #
COD calibration std. 100 ppm	WP59082
COD calibration std. 10 ppm	WP59081
COD calibration std. 150 ppm	WP59083
COD calibration std. 50 ppm	WP59084
COD calibration std. 0 ppm	WP59087
COD ICV-LCS std, 50ppm	WP59370
COD Digestion Vials Low Level 0-150Mg/L	W2268
COD ICV-LCS std, 50ppm	WP59370
COD CCV std, 50ppm	WP59371

Temp In (C): <u>150</u>	Date In: <u>11/03/2017</u>	Time In: <u>09:20</u>
Temp Out (C): <u>150</u>	Date Out: <u>11/03/2017</u>	Time Out: <u>11:25</u>

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/26/2017	11:10
2	CAL2	10	1	Water	11.000	10/26/2017	11:10
3	CAL3	50	1	Water	50.000	10/26/2017	11:11
4	CAL4	100	1	Water	102.000	10/26/2017	11:11
5	CAL5	150	1	Water	151.000	10/26/2017	11:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB91329

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	10/26/2017	11:12
2	ICB				1	Water	0.000	-0.360	11/03/2017	11:13
3	CCV1	50			1	Water	51.000	50.280	11/03/2017	12:40
4	CCB1				1	Water	0.000	-0.360	11/03/2017	12:40
5	LB91329BL				1	Water	0.000	-0.360	11/03/2017	12:41
6	LB91329BS	50			1	Water	49.000	48.294	11/03/2017	12:41
7	I6243-01				1	Water	11.000	10.562	11/03/2017	12:50
8	I6269-01				1	Water	0.000	-0.360	11/03/2017	12:50
9	I6269-01DUP				1	Water	0.000	-0.360	11/03/2017	12:51
10	I6269-01MS	50			1	Water	52.000	51.273	11/03/2017	12:51
11	I6269-01MSD	50			1	Water	51.000	50.280	11/03/2017	12:52
12	I6269-02				1	Water	0.000	-0.360	11/03/2017	12:52
13	I6275-09				1	Water	62.000	61.203	11/03/2017	12:53
14	I6277-01				1	Water	145.000	143.618	11/03/2017	12:53
15	CCV2	50			1	Water	51.000	50.280	11/03/2017	12:54
16	CCB2				1	Water	0.000	-0.360	11/03/2017	12:54
17	I6294-01				1	Water	48.000	47.301	11/03/2017	12:55
18	I6296-01				10	Water	16.000	15.527	11/03/2017	12:55
19	CCV3	50			1	Water	52.000	51.273	11/03/2017	12:56
20	CCB3				1	Water	0.000	-0.360	11/03/2017	12:56

2891329

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD110217-2

WorkList ID : 105049

Date : 11/2/2017 10:18:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2017	Water	I6243-01	COD	Conc H2SO4 tc	PSEG01	D52	N35799	11/01/2017	SM5220 D
11/08/2017	Water	I6275-09	COD	Conc H2SO4 tc	PSEG01	D52	2727	11/01/2017	SM5220 D
11/05/2017	Water	I6277-01	COD	Conc H2SO4 tc	PSEG01	D52	FADW4807L	11/01/2017	SM5220 D
11/07/2017	Water	I6294-01	COD	Conc H2SO4 tc	PSEG01	H52	N48971	11/02/2017	SM5220 D
11/07/2017	Water	I6296-01	COD	Conc H2SO4 tc	PSEG01	H52	SV2-1854L	11/02/2017	SM5220 D
11/08/2017	Water	I6269-01	COD	Conc H2SO4 tc	SHAW22	D61	OUTFALL001-171024	10/24/2017	SM5220 D
11/08/2017	Water	I6269-02	COD	Conc H2SO4 tc	SHAW22	D61	OUTFALL001-171031	10/31/2017	SM5220 D

Date/Time

11/3/17 9:00

Received by:

IB

Relinquished by:

Date/Time

11/3/17 10:00

Received by:

Relinquished by: