

7471A


LB91330

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42067



A= 0.0000e+000

B= 2.4000e-004

C= 1.3200e-003

Rho= 0.9998024

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	16	0.000	16				
0.2	0.200	0.201	0.001	830	0.0 %	830				
2.5	2.500	2.563	0.063	10674	0.0 %	10674				
5.0	5.000	4.967	-0.033	20691	0.0 %	20691				
7.5	7.500	7.367	-0.133	30690	0.0 %	30690	0			
10.0	10.000	10.097	0.097	42067	0.0 %	42067				

LB91330 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 03 Nov 2017 12:06:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7471A	PPB	03 Nov 2017 12:09:15	S	Std
0.2 - 1	51.2	830	-	0.2000	7471A	PPB	03 Nov 2017 12:11:32	S	Std
2.5 - 1	52.5	10674	-	2.5000	7471A	PPB	03 Nov 2017 12:13:50	S	Std
5.0 - 1	55	20691	-	5.0000	7471A	PPB	03 Nov 2017 12:16:07	S	Std
7.5 - 1	57.5	30690	-	7.5000	7471A	PPB	03 Nov 2017 12:18:25	S	Std
10.0 - 1	510	42067	-	10.0000	7471A	PPB	03 Nov 2017 12:20:42	S	Std
ICV44 - 1	ICV44	15649	3.7570	-7471A	PPB	PPB	03 Nov 2017 12:23:23	U	SMPL
ICB44 - 1	ICB44	-125	-0.0287	-7471A	PPB	PPB	03 Nov 2017 12:25:39	U	SMPL
CCV90 - 1	CCV90	20734	4.9774	-7471A	PPB	PPB	03 Nov 2017 12:27:56	U	SMPL
CCB90 - 1	CCB90	-158	-0.0366	-7471A	PPB	PPB	03 Nov 2017 12:30:11	U	SMPL
CRA - 1	CRA	819	0.1979	-7471A	PPB	PPB	03 Nov 2017 12:32:29	U	SMPL
HighStd - 1	HighStd	42179	10.1242	-7471A	PPB	PPB	03 Nov 2017 12:34:45	U	SMPL
ChkStd - 1	ChkStd	29471	7.0743	-7471A	PPB	PPB	03 Nov 2017 12:37:00	U	SMPL
PB103940BL - 1	PBS	-191	-0.0445	-7471A	PPB	PPB	03 Nov 2017 12:39:19	U	SMPL
PB103940BS - 1	LCSS	15440	3.7069	-7471A	PPB	PPB	03 Nov 2017 12:41:38	U	SMPL
I6285-01 - 1	P002-S001-0001-01	2570	0.6181	-7471A	PPB	PPB	03 Nov 2017 12:43:54	U	SMPL
I6285-02 - 1	P002-S001-0001-02	2627	0.6318	-7471A	PPB	PPB	03 Nov 2017 12:46:12	U	SMPL
I6285-03 - 1	P002-S002-0001-01	2437	0.5862	-7471A	PPB	PPB	03 Nov 2017 12:48:29	U	SMPL
I6285-03DUP - 1	P002-S002-0001-01DUP	2436	0.5860	-7471A	PPB	PPB	03 Nov 2017 12:50:47	U	SMPL
I6285-04 - 1	I6285-03MS	18538	4.4504	-7471A	PPB	PPB	03 Nov 2017 12:53:03	U	SMPL
CCV91 - 1	CCV91	20469	4.9138	-7471A	PPB	PPB	03 Nov 2017 12:55:19	U	SMPL
CCB91 - 1	CCB91	-113	-0.0258	-7471A	PPB	PPB	03 Nov 2017 12:57:35	U	SMPL
I6285-05 - 1	I6285-03MSD	18864	4.5286	-7471A	PPB	PPB	03 Nov 2017 12:59:52	U	SMPL
I6285-06 - 1	P002-S003-0001-01	2876	0.6916	-7471A	PPB	PPB	03 Nov 2017 13:02:07	U	SMPL
I6285-07 - 1	P002-S004-0001-01	2376	0.5716	-7471A	PPB	PPB	03 Nov 2017 13:04:25	U	SMPL
I6285-08 - 1	P002-S005-0001-01	2612	0.6282	-7471A	PPB	PPB	03 Nov 2017 13:06:41	U	SMPL
I6289-01 - 1	SU-03-110217	794	0.1919	-7471A	PPB	PPB	03 Nov 2017 13:08:58	U	SMPL
I6297-01 - 1	CHRT-24646	126	0.0316	-7471A	PPB	PPB	03 Nov 2017 13:11:15	U	SMPL
I6299-01 - 1	OK-02-110217	289	0.0707	-7471A	PPB	PPB	03 Nov 2017 13:13:32	U	SMPL
I6285-03LX5 - 1		482	0.1170	-7471A	PPB	PPB	03 Nov 2017 13:15:50	U	SMPL
I6285-03A - 1		18922	4.5426	-7471A	PPB	PPB	03 Nov 2017 13:18:07	U	SMPL
CCV92 - 1	CCV92	20508	4.9232	-7471A	PPB	PPB	03 Nov 2017 13:20:24	U	SMPL
CCB92 - 1	CCB92	-156	-0.0361	-7471A	PPB	PPB	03 Nov 2017 13:22:41	U	SMPL