



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91333

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/03/2017	12:10
2	CAL2	1	Water			7.00	11/03/2017	12:11
3	CAL3	1	Water			10.03	11/03/2017	12:12
4	ICV	1	Water			7.00	11/03/2017	12:13
5	CCV1	1	Water			2.01	11/03/2017	12:15
6	I6313-01	1	Solid	20.02	20	4.39	11/03/2017	12:16
7	I6313-01DUP	1	Solid	20.04	20	4.39	11/03/2017	12:18
8	I6313-04	1	Solid	20.02	20	4.28	11/03/2017	12:20
9	I6313-05	1	Solid	20.04	20	4.05	11/03/2017	12:21
10	I6313-06	1	Solid	20.05	20	4.21	11/03/2017	12:23
11	I6313-07	1	Solid	20.06	20	4.19	11/03/2017	12:25
12	I6313-08	1	Solid	20.03	20	4.16	11/03/2017	12:26
13	I6313-09	1	Solid	20.04	20	4.15	11/03/2017	12:29
14	I6313-10	1	Solid	20.01	20	4.19	11/03/2017	12:30
15	I6313-11	1	Solid	20.03	20	4.13	11/03/2017	12:32
16	CCV2	1	Water			12.02	11/03/2017	12:33
17	I6313-12	1	Solid	20.04	20	4.10	11/03/2017	12:35
18	I6313-13	1	Solid	20.05	20	4.19	11/03/2017	12:36
19	CCV3	1	Water			12.01	11/03/2017	12:38

2391333

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6313

WorkList ID : 105112

Date : 11/3/2017 12:00:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2017	Solid	I6313-01-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF010-01	11/02/2017	9045C
11/02/2017	Solid	I6313-04-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF010-02	11/02/2017	9045C
11/02/2017	Solid	I6313-05-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF011-01	11/02/2017	9045C
11/02/2017	Solid	I6313-06-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF012-01	11/02/2017	9045C
11/02/2017	Solid	I6313-07-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF013-01	11/02/2017	9045C
11/02/2017	Solid	I6313-08-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF014-01	11/02/2017	9045C
11/02/2017	Solid	I6313-09-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF015-01	11/02/2017	9045C
11/02/2017	Solid	I6313-10-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF016-01	11/02/2017	9045C
11/02/2017	Solid	I6313-11-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF017-01	11/02/2017	9045C
11/02/2017	Solid	I6313-12-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF018-01	11/02/2017	9045C
11/02/2017	Solid	I6313-13-1A	pH	Cool 4 deg C	ROYF02	C23	PET-BF019-01	11/02/2017	9045C

Date/Time 11-03-17 12:00 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 11-03-17 2:30 PM
Received by: [Signature]
Relinquished by: [Signature]