


245.1

LB 91346
INSTRUMENT ID: CV1

Linear

 μ Abs.:

31493



A= 0.0000e+000

B= 3.1967e-004

C= -1.6148e-002

Rho= 0.9999028

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	24	0.000	24				
0.2	0.200	0.192	-0.008	652	0.0 %	652				
2.5	2.500	2.537	0.037	7986	0.0 %	7986				
5.0	5.000	5.032	0.032	15792	0.0 %	15792				
7.5	7.500	7.396	-0.104	23188	0.0 %	23188				
10.0	10.000	10.051	0.051	31493	0.0 %	31493				

LB91346 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Nov 2017 15:27:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	245.1	PPB	03 Nov 2017 15:31:06	S	Std
0.2 - 1	50.2	652	-	0.2000	245.1	PPB	03 Nov 2017 15:33:15	S	Std
2.5 - 1	52.5	7986	-	2.5000	245.1	PPB	03 Nov 2017 15:35:23	S	Std
5.0 - 1	55	15792	-	5.0000	245.1	PPB	03 Nov 2017 15:37:31	S	Std
7.5 - 1	57.5	23188	-	7.5000	245.1	PPB	03 Nov 2017 15:39:39	S	Std
10.0 - 1	50	31493	-	10.0000	245.1	PPB	03 Nov 2017 15:44:27	S	Std
ICV45 - 1	ICV45	12392	3.9452	-245.1		PPB	03 Nov 2017 15:46:57	U	SMPL
ICB45 - 1	ICB45	-132	-0.0583	-245.1		PPB	03 Nov 2017 15:49:05	U	SMPL
CCV93 - 1	CCV93	15574	4.9624	-245.1		PPB	03 Nov 2017 15:51:12	U	SMPL
CCB93 - 1	CCB93	-126	-0.0564	-245.1		PPB	03 Nov 2017 15:53:19	U	SMPL
CRA - 1	CRA	626	0.1840	-245.1		PPB	03 Nov 2017 15:55:26	U	SMPL
HighStd - 1	HighStd	31313	9.9936	-245.1		PPB	03 Nov 2017 15:57:33	U	SMPL
ChkStd - 1	ChkStd	21609	6.8915	-245.1		PPB	03 Nov 2017 15:59:39	U	SMPL
PB103942BL - 1	PBW	-164	-0.0686	-245.1		PPB	03 Nov 2017 16:01:49	U	SMPL
PB103942BS - 1	LCSW	13308	4.2380	-245.1		PPB	03 Nov 2017 16:03:58	U	SMPL
I6273-01 - 1	103117-TIDO-E-D-B	-122	-0.0551	-245.1		PPB	03 Nov 2017 16:06:06	U	SMPL
I6273-02 - 1	103117-TIDO-E-D-S	-46	-0.0309	-245.1		PPB	03 Nov 2017 16:08:14	U	SMPL
I6273-03 - 1	103117-TIDO-E-U-B	-16	-0.0213	-245.1		PPB	03 Nov 2017 16:10:22	U	SMPL
I6273-04 - 1	103117-TIDO-E-U-S	-9	-0.0190	-245.1		PPB	03 Nov 2017 16:12:31	U	SMPL
I6273-05 - 1	103117-SIDI-E-D-B	-6	-0.0181	-245.1		PPB	03 Nov 2017 16:14:37	U	SMPL
CCV94 - 1	CCV94	15684	4.9975	-245.1		PPB	03 Nov 2017 16:16:44	U	SMPL
CCB94 - 1	CCB94	-128	-0.0571	-245.1		PPB	03 Nov 2017 16:18:50	U	SMPL
I6273-06 - 1	103117-SIDI-E-D-S	0	-0.0161	-245.1		PPB	03 Nov 2017 16:20:58	U	SMPL
I6273-07 - 1	103117-SIDI-E-U-B	-4	-0.0174	-245.1		PPB	03 Nov 2017 16:23:05	U	SMPL
I6273-08 - 1	103117-SIDI-E-U-S	22	-0.0091	-245.1		PPB	03 Nov 2017 16:25:11	U	SMPL
I6273-09 - 1	103117-SIDI-E-D-M	-9	-0.0190	-245.1		PPB	03 Nov 2017 16:27:18	U	SMPL
I6273-10 - 1	103117-SIDI-E-U-M	0	-0.0161	-245.1		PPB	03 Nov 2017 16:29:26	U	SMPL
I6274-01 - 1	110117-TIDO-F-D-B	-5	-0.0177	-245.1		PPB	03 Nov 2017 16:31:33	U	SMPL
I6274-02 - 1	110117-TIDO-F-D-S	-2	-0.0168	-245.1		PPB	03 Nov 2017 16:33:41	U	SMPL
I6274-03 - 1	110117-TIDO-F-U-B	-13	-0.0203	-245.1		PPB	03 Nov 2017 16:35:50	U	SMPL
I6274-04 - 1	110117-TIDO-F-U-S	-4	-0.0174	-245.1		PPB	03 Nov 2017 16:37:58	U	SMPL
I6278-01 - 1	COMP	70	0.0062	-245.1		PPB	03 Nov 2017 16:40:05	U	SMPL
CCV95 - 1	CCV95	15282	4.8690	-245.1		PPB	03 Nov 2017 16:42:12	U	SMPL
CCB95 - 1	CCB95	-115	-0.0529	-245.1		PPB	03 Nov 2017 16:44:19	U	SMPL
I6278-01DUP - 1	COMPDUP	86	0.0113	-245.1		PPB	03 Nov 2017 16:46:27	U	SMPL
I6278-01MS - 1	COMPMS	12145	3.8662	-245.1		PPB	03 Nov 2017 16:48:34	U	SMPL
I6278-01MSD - 1	COMPMSD	12319	3.9218	-245.1		PPB	03 Nov 2017 16:50:40	U	SMPL
I6329-01 - 1	B-EMH-F	391	0.1088	-245.1		PPB	03 Nov 2017 16:52:47	U	SMPL
I6278-01LX5 - 1		-1	-0.0165	-245.1		PPB	03 Nov 2017 16:54:55	U	SMPL
I6278-01A - 1		13272	4.2265	-245.1		PPB	03 Nov 2017 16:57:03	U	SMPL
CCV96 - 1	CCV96	15435	4.9179	-245.1		PPB	03 Nov 2017 17:01:46	U	SMPL
CCB96 - 1	CCB96	-122	-0.0551	-245.1		PPB	03 Nov 2017 17:03:54	U	SMPL