



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91354

Temp.: 21.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/03/2017	17:30
2	CAL2	1	Water			7.00	11/03/2017	17:31
3	CAL3	1	Water			10.03	11/03/2017	17:33
4	ICV	1	Water			7.00	11/03/2017	17:35
5	CCV1	1	Water			2.02	11/03/2017	17:36
6	I6338-03	1	Water			7.67	11/03/2017	17:37
7	I6338-03DUP	1	Water			7.68	11/03/2017	17:39
8	I6339-02	1	Water			9.64	11/03/2017	17:41
9	I6341-02	1	Water			8.02	11/03/2017	17:43
10	CCV2	1	Water			12.01	11/03/2017	17:45

18A1354

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6340 WorkList ID : 105132 Date : 11/3/2017 5:23:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/03/2017	Water	I6338-03-17	Corrosivity	Cool 4 deg C	PSEG01	H52	TK2021L	11/03/2017	9040C
11/03/2017	Water	I6339-02-17	Corrosivity	Cool 4 deg C	SEAC01	H43	A017	11/03/2017	9040C
11/03/2017	Water	I6341-02-17	Corrosivity	Cool 4 deg C	PSEG01	H52	A6835DW-A6865DW	11/03/2017	9040C

Date/Time 11-03-17 5:23 PM
 Received by: JN
 Relinquished by: CA

Date/Time 11-03-17 6:00 PM
 Received by: CA
 Relinquished by: JD