



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91356

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/03/2017	17:00
2	CAL2	1	Water			7.01	11/03/2017	17:01
3	CAL3	1	Water			10.04	11/03/2017	17:03
4	ICV	1	Water			7.00	11/03/2017	17:05
5	CCV1	1	Water			2.01	11/03/2017	17:06
6	I6305-02	1	Solid	20.02	20	9.02	11/03/2017	17:08
7	I6305-02DUP	1	Solid	20.04	20	9.03	11/03/2017	17:10
8	I6334-02	1	Solid	20.01	20	7.62	11/03/2017	17:11
9	I6337-02	1	Solid	20.04	20	7.50	11/03/2017	17:14
10	I6338-01	1	Solid	20.04	20	8.40	11/03/2017	17:15
11	I6339-01	1	Solid	20.02	20	11.42	11/03/2017	17:17
12	CCV2	1	Water			12.01	11/03/2017	17:18

2891356

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6341

WorkList ID : 105130

Date : 11/3/2017 4:51:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2017	Solid	I6305-02	Corrosivity	Cool 4 deg C	PSEG05	H52	PL-01-110317	11/03/2017	9045C
11/08/2017	Solid	I6334-02	Corrosivity	Cool 4 deg C	PSEG01	H52	278	11/03/2017	9045C
11/08/2017	Solid	I6337-02	Corrosivity	Cool 4 deg C	PSEG01	H52	WOODLAKE-COMP-5.1	11/03/2017	9045C
11/08/2017	Solid	I6338-01	Corrosivity	Cool 4 deg C	PSEG01	H52	2015	11/03/2017	9045C
11/17/2017	Solid	I6339-01	Corrosivity	Cool 4 deg C	SEAC01	H43	DRUMS	11/03/2017	9045C

Date/Time

11-03-17 5:00 PM

Received by:

JD

Relinquished by:

WD

Date/Time

11-03-17

Received by:

WJ

Relinquished by:

JD