



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91358

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	11/03/2017	09:15
2	I6243-02		1	>100.0	>212.0	11/03/2017	09:46
3	I6243-02DUP		1	>100.0	>212.0	11/03/2017	10:10
4	I6275-10		1	>100.0	>212.0	11/03/2017	10:45
5	I6277-02		1	>100.0	>212.0	11/03/2017	11:15
6	I6294-02		1	>100.0	>212.0	11/03/2017	11:40
7	I6296-02		1	>100.0	>212.0	11/03/2017	12:15
8	I6329-01		1	>100.0	>212.0	11/03/2017	12:50
9	I6338-03		1	>100.0	>212.0	11/03/2017	13:15
10	I6339-02		1	>100.0	>212.0	11/03/2017	13:52
11	I6341-02		1	>100.0	>212.0	11/03/2017	14:16

Result = (Celsius * 1.8) + 32

LB91358

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP11317

WorkList ID : 105131

Date : 11/3/2017 08:30AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2017	Water	I6243-02	Flash Point	Cool 4 deg C	PSEG01	D52	N35799	11/01/2017	1010A
11/08/2017	Water	I6275-10	Flash Point	Cool 4 deg C	PSEG01	D52	2727	11/01/2017	1010A
11/05/2017	Water	I6277-02	Flash Point	Cool 4 deg C	PSEG01	D52	FADW4807L	11/01/2017	1010A
11/07/2017	Water	I6294-02	Flash Point	Cool 4 deg C	PSEG01	H52	N48971	11/02/2017	1010A
11/07/2017	Water	I6296-02	Flash Point	Cool 4 deg C	PSEG01	H52	SV2-1854L	11/02/2017	1010A
11/08/2017	Water	I6329-01	Flash Point	Cool 4 deg C	DIST01	H53	B-EMH-F	11/03/2017	1010A
11/08/2017	Water	I6338-03	Flash Point	Cool 4 deg C	PSEG01	H52	TK2021L	11/03/2017	1010A
11/17/2017	Water	I6339-02	Flash Point	Cool 4 deg C	SEAC01	H43	A017	11/03/2017	1010A
11/08/2017	Water	I6341-02	Flash Point	Cool 4 deg C	PSEG01	H52	A6835DW-A6865DW	11/03/2017	1010A

Date/Time 11/3/17 8:35AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/3/17 14:25
 Received by: CP
 Relinquished by: KS