

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/06/2017

DATE & TIME IN: 11/06/2017 09:50

OVEN TEMP: 103 °C

Run Number: LB91367

DATE & TIME OUT: 11/06/2017 13:50

OVEN TEMP: 105 °C

BalanceID: WC Sc-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB91367BL	LB91367BL	1.4199	1.4199	100	1.4199	1.4199	1.4199	0	0
2	LB91367BS	LB91367BS	1.4021	1.4021	100	1.4572	1.4572	1.4572	0.0551	551
3	I6327-01	110217-TIDO-F-D-B	1.4137	1.4137	200	1.4353	1.4353	1.4353	0.0216	108
4	I6327-01DUP	110217-TIDO-F-D-BDUP	1.4191	1.4191	200	1.4409	1.4409	1.4409	0.0218	109
5	I6327-02	110217-TIDO-F-D-S	1.4252	1.4252	400	1.4408	1.4408	1.4408	0.0156	39
6	I6327-03	110217-110217-TIDO-F-U-B	1.4211	1.4211	300	1.4353	1.4353	1.4353	0.0142	47.3
7	I6327-04	110217-TIDO-F-U-S	1.4218	1.4218	400	1.4347	1.4347	1.4347	0.0129	32.3
8	I6327-05	110217-JETT-F-D-B	1.4162	1.4162	300	1.4387	1.4387	1.4387	0.0225	75
9	I6327-06	110217-JETT-F-D-S	1.413	1.413	300	1.4317	1.4317	1.4317	0.0187	62.3
10	I6327-07	110217-JETT-F-U-B	1.4084	1.4084	300	1.4271	1.4271	1.4271	0.0187	62.3
11	I6327-08	110217-JETT-F-U-S	1.4241	1.4241	700	1.4357	1.4357	1.4357	0.0116	16.6
12	I6327-09	110217-SIDI-F-D-B	1.4145	1.4145	300	1.4223	1.4223	1.4223	0.0078	26
13	I6327-10	110217-SIDI-F-D-S	1.4125	1.4125	700	1.4225	1.4225	1.4225	0.01	14.3
14	I6327-11	110217-SIDI-F-U-B	1.413	1.413	400	1.4203	1.4203	1.4203	0.0073	18.3
15	I6327-12	110217-SIDI-F-U-S	1.4128	1.4128	600	1.4242	1.4242	1.4242	0.0114	19

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/06/2017

DATE & TIME IN: 11/06/2017 09:50

OVEN TEMP: 103 °C

Run Number: LB91367

DATE & TIME OUT: 11/06/2017 13:50

OVEN TEMP: 105 °C

BalanceID: WC Sc-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	I6329-01	B-EMH-F	1.4113	1.4113	20	1.4324	1.4324	1.4324	0.0211	1055
17	I6330-01	001M-WILLETS-PTBLVD (NOV)	1.4302	1.4302	100	1.4411	1.4411	1.4411	0.0109	109
18	I6330-02	002M-35THAVE (NOV)	1.4118	1.4118	200	1.4211	1.4211	1.4211	0.0093	46.5
19	I6338-02	TK2021L	1.4181	1.4181	25	1.4456	1.4456	1.4456	0.0275	1100
20	I6341-01	A6835DW-A6865DW	1.4234	1.4234	600	1.4293	1.4293	1.4293	0.0059	9.8

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS110617

WorkList ID : 105139

Date : 11/06/2017 8:14:36 AM

LB91367

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2017	Water	I6327-01 - C	TSS	Cool 4 deg C	HDR103	H52	110217-TIDO-F-D-B	11/02/2017	SM2540 D
11/09/2017	Water	I6327-02 - C	TSS	Cool 4 deg C	HDR103	H52	110217-TIDO-F-D-S	11/02/2017	SM2540 D
11/09/2017	Water	I6327-03 - C	TSS	Cool 4 deg C	HDR103	H52	110217-110217-TIDO-F-U-B	11/02/2017	SM2540 D
11/09/2017	Water	I6327-04 - C	TSS	Cool 4 deg C	HDR103	H52	110217-TIDO-F-U-S	11/02/2017	SM2540 D
11/09/2017	Water	I6327-05 - C	TSS	Cool 4 deg C	HDR103	H52	110217-JETT-F-D-B	11/02/2017	SM2540 D
11/09/2017	Water	I6327-06 - C	TSS	Cool 4 deg C	HDR103	H52	110217-JETT-F-D-S	11/02/2017	SM2540 D
11/09/2017	Water	I6327-07 - C	TSS	Cool 4 deg C	HDR103	H52	110217-JETT-F-U-B	11/02/2017	SM2540 D
11/09/2017	Water	I6327-08 - C	TSS	Cool 4 deg C	HDR103	H52	110217-JETT-F-U-S	11/02/2017	SM2540 D
11/09/2017	Water	I6327-09 - C	TSS	Cool 4 deg C	HDR103	H52	110217-SIDI-F-D-B	11/02/2017	SM2540 D
11/09/2017	Water	I6327-10 - C	TSS	Cool 4 deg C	HDR103	H52	110217-SIDI-F-D-S	11/02/2017	SM2540 D
11/09/2017	Water	I6327-11 - C	TSS	Cool 4 deg C	HDR103	H52	110217-SIDI-F-U-B	11/02/2017	SM2540 D
11/09/2017	Water	I6327-12 - C	TSS	Cool 4 deg C	HDR103	H52	110217-SIDI-F-U-S	11/02/2017	SM2540 D
11/08/2017	Water	I6329-01 - F	TSS	Cool 4 deg C	DIST01	H53	B-EMH-F	11/03/2017	SM2540 D
11/05/2017	Water	I6330-01 - E	TSS	Cool 4 deg C	TULL01	H21	001M-WILLETS-PTBLVD(NO)	11/02/2017	SM2540 D
11/05/2017	Water	I6330-02 - F	TSS	Cool 4 deg C	TULL01	H21	002M-35THAVE(NOV)	11/02/2017	SM2540 D
11/08/2017	Water	I6338-02 - F	TSS	Cool 4 deg C	PSEG01	H52	TK2021L	11/03/2017	SM2540 D
11/08/2017	Water	I6341-01 - F	TSS	Cool 4 deg C	PSEG01	H52	A6835DW-A6865DW	11/03/2017	SM2540 D

Date/Time 11/06/17, 9:00
 Received by: CPAS
 Relinquished by: CPAS

Date/Time 11/06/17, 10:10
 Received by: CPAS
 Relinquished by: CPAS